

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE091415\
 Data File : BE090808.D
 Acq On : 14 Sep 2015 20:55
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 15 00:58:37 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 11:40:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	120	0.00
2	1,4-Dioxane	1.000	0.848	15.2	98	0.00
3	n-Nitrosodimethylamine	1.000	0.841	15.9	99	0.00
4 S	2-Fluorophenol	1.000	0.798	20.2	93	0.00
5 S	Phenol-d6	1.000	0.849	15.1	107	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	131	0.00
7 S	Nitrobenzene-d5	1.000	0.836	16.4	110	0.00
8	Nitrobenzene	1.000	0.850	15.0	115	0.00
9	Naphthalene	1.000	0.873	12.7	115	0.00
10	Hexachlorobutadiene	1.000	0.868	13.2	110	0.00
11	2-Methylnaphthalene	1.000	0.887	11.3	120	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	142	0.00
13 S	2,4,6-Tribromophenol	1.000	0.749	25.1#	105	0.00
14 S	2-Fluorobiphenyl	1.000	0.833	16.7	120	0.00
15	Acenaphthylene	1.000	0.833	16.7	123	0.00
16	Acenaphthene	1.000	0.854	14.6	120	0.00
17	Fluorene	1.000	0.844	15.6	121	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	134	0.00
19	4-Bromophenyl-phenylether	1.000	0.853	14.7	119	-0.02
20	Hexachlorobenzene	1.000	0.844	15.6	113	0.00
21	Pentachlorophenol	1.000	0.937	6.3	100	0.00
22	Phenanthrene	1.000	0.871	12.9	117	-0.02
23	Anthracene	1.000	0.895	10.5	125	0.00
24	Fluoranthene	1.000	0.863	13.7	125	-0.01
25 I	Chrysene-d12	5.000	5.000	0.0	140	0.00
26	Pyrene	1.000	0.983	1.7	145	-0.01
27 S	Terphenyl-d14	1.000	0.903	9.7	132	-0.01
28	Benzo(a)anthracene	1.000	0.932	6.8	132	0.00
29	Chrysene	1.000	0.913	8.7	124	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.910	9.0	141	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	0.968	3.2	146	-0.01
32 I	Perylene-d12	5.000	5.000	0.0	148	0.00
33	Benzo(b)fluoranthene	1.000	0.898	10.2	137	0.00
34	Benzo(k)fluoranthene	1.000	0.910	9.0	137	0.00
35 C	Benzo(a)pyrene	1.000	0.916	8.4	145	0.00
36	Dibenzo(a,h)anthracene	1.000	0.925	7.5	145	-0.02
37	Benzo(g,h,i)perylene	1.000	0.856	14.4	134	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0