

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE091415\
 Data File : BE090808.D
 Acq On : 14 Sep 2015 20:55
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 15 00:58:37 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 11:40:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	0.00
2	1,4-Dioxane	0.814	0.648	20.4	98	0.00
3	n-Nitrosodimethylamine	1.033	0.806	22.0	99	0.00
4 S	2-Fluorophenol	1.039	0.762	26.7#	93	0.00
5 S	Phenol-d6	1.392	1.182	15.1	107	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	131	0.00
7 S	Nitrobenzene-d5	0.330	0.276	16.4	110	0.00
8	Nitrobenzene	0.365	0.308	15.6	115	0.00
9	Naphthalene	1.091	0.952	12.7	115	0.00
10	Hexachlorobutadiene	0.171	0.148	13.5	110	0.00
11	2-Methylnaphthalene	0.590	0.523	11.4	120	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	142	0.00
13 S	2,4,6-Tribromophenol	0.143	0.089	37.8#	105	0.00
14 S	2-Fluorobiphenyl	1.385	1.153	16.8	120	0.00
15	Acenaphthylene	7.992	6.653	16.8	123	0.00
16	Acenaphthene	1.335	1.140	14.6	120	0.00
17	Fluorene	1.666	1.406	15.6	121	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	134	0.00
19	4-Bromophenyl-phenylether	0.170	0.145	14.7	119	-0.02
20	Hexachlorobenzene	0.928	0.771	16.9	113	0.00
21	Pentachlorophenol	0.054	0.035	35.2#	100	0.00
22	Phenanthrene	1.237	1.077	12.9	117	-0.02
23	Anthracene	1.035	0.926	10.5	125	0.00
24	Fluoranthene	1.304	1.125	13.7	125	-0.01
25 I	Chrysene-d12	1.000	1.000	0.0	140	0.00
26	Pyrene	1.035	1.017	1.7	145	-0.01
27 S	Terphenyl-d14	0.553	0.499	9.8	132	-0.01
28	Benzo(a)anthracene	1.156	1.035	10.5	132	0.00
29	Chrysene	1.279	1.124	12.1	124	0.00
30	Bis(2-ethylhexyl)phthalate	0.345	0.256	25.8#	141	0.00
31	Indeno(1,2,3-cd)pyrene	1.181	1.143	3.2	146	-0.01
32 I	Perylene-d12	1.000	1.000	0.0	148	0.00
33	Benzo(b)fluoranthene	1.084	0.974	10.1	137	0.00
34	Benzo(k)fluoranthene	1.263	1.149	9.0	137	0.00
35 C	Benzo(a)pyrene	0.970	0.888	8.5	145	0.00
36	Dibenzo(a,h)anthracene	1.020	0.956	6.3	145	-0.02
37	Benzo(g,h,i)perylene	1.116	0.955	14.4	134	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0