

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE091415\
 Data File : BE090810.D
 Acq On : 14 Sep 2015 22:06
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 15 00:58:53 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 11:40:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	115	0.00
2	1,4-Dioxane	1.000	0.820	18.0	92	0.00
3	n-Nitrosodimethylamine	1.000	0.832	16.8	94	0.00
4 S	2-Fluorophenol	1.000	0.773	22.7	87	0.00
5 S	Phenol-d6	1.000	0.786	21.4	95	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	124	0.00
7 S	Nitrobenzene-d5	1.000	0.832	16.8	104	0.00
8	Nitrobenzene	1.000	0.858	14.2	110	0.00
9	Naphthalene	1.000	0.873	12.7	109	-0.01
10	Hexachlorobutadiene	1.000	0.889	11.1	107	0.00
11	2-Methylnaphthalene	1.000	0.886	11.4	114	-0.01
12 I	Acenaphthene-d10	5.000	5.000	0.0	137	0.00
13 S	2,4,6-Tribromophenol	1.000	0.711	28.9#	94	0.00
14 S	2-Fluorobiphenyl	1.000	0.832	16.8	115	0.00
15	Acenaphthylene	1.000	0.827	17.3	118	0.00
16	Acenaphthene	1.000	0.860	14.0	116	0.00
17	Fluorene	1.000	0.838	16.2	116	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	129	0.00
19	4-Bromophenyl-phenylether	1.000	0.850	15.0	114	-0.02
20	Hexachlorobenzene	1.000	0.864	13.6	111	0.00
21	Pentachlorophenol	1.000	0.916	8.4	92	0.00
22	Phenanthrene	1.000	0.889	11.1	115	-0.02
23	Anthracene	1.000	0.876	12.4	118	0.00
24	Fluoranthene	1.000	0.849	15.1	118	-0.01
25 I	Chrysene-d12	5.000	5.000	0.0	135	0.00
26	Pyrene	1.000	0.977	2.3	139	-0.01
27 S	Terphenyl-d14	1.000	0.898	10.2	127	-0.01
28	Benzo(a)anthracene	1.000	0.931	6.9	127	0.00
29	Chrysene	1.000	0.901	9.9	118	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.862	13.8	128	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.005	-0.5	147	-0.01
32 I	Perylene-d12	5.000	5.000	0.0	147	-0.01
33	Benzo(b)fluoranthene	1.000	0.916	8.4	139	0.00
34	Benzo(k)fluoranthene	1.000	0.882	11.8	132	0.00
35 C	Benzo(a)pyrene	1.000	0.909	9.1	143	0.00
36	Dibenzo(a,h)anthracene	1.000	0.927	7.3	144	-0.02
37	Benzo(g,h,i)perylene	1.000	0.859	14.1	134	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0