

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE091415\
 Data File : BE090810.D
 Acq On : 14 Sep 2015 22:06
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 15 00:58:53 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 11:40:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	0.00
2	1,4-Dioxane	0.814	0.631	22.5	92	0.00
3	n-Nitrosodimethylamine	1.033	0.798	22.7	94	0.00
4 S	2-Fluorophenol	1.039	0.737	29.1#	87	0.00
5 S	Phenol-d6	1.392	1.094	21.4	95	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	124	0.00
7 S	Nitrobenzene-d5	0.330	0.275	16.7	104	0.00
8	Nitrobenzene	0.365	0.311	14.8	110	0.00
9	Naphthalene	1.091	0.953	12.6	109	-0.01
10	Hexachlorobutadiene	0.171	0.152	11.1	107	0.00
11	2-Methylnaphthalene	0.590	0.522	11.5	114	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	137	0.00
13 S	2,4,6-Tribromophenol	0.143	0.083	42.0#	94	0.00
14 S	2-Fluorobiphenyl	1.385	1.152	16.8	115	0.00
15	Acenaphthylene	7.992	6.612	17.3	118	0.00
16	Acenaphthene	1.335	1.149	13.9	116	0.00
17	Fluorene	1.666	1.395	16.3	116	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	129	0.00
19	4-Bromophenyl-phenylether	0.170	0.145	14.7	114	-0.02
20	Hexachlorobenzene	0.928	0.787	15.2	111	0.00
21	Pentachlorophenol	0.054	0.034	37.0#	92	0.00
22	Phenanthrene	1.237	1.101	11.0	115	-0.02
23	Anthracene	1.035	0.907	12.4	118	0.00
24	Fluoranthene	1.304	1.108	15.0	118	-0.01
25 I	Chrysene-d12	1.000	1.000	0.0	135	0.00
26	Pyrene	1.035	1.011	2.3	139	-0.01
27 S	Terphenyl-d14	0.553	0.497	10.1	127	-0.01
28	Benzo(a)anthracene	1.156	1.035	10.5	127	0.00
29	Chrysene	1.279	1.110	13.2	118	0.00
30	Bis(2-ethylhexyl)phthalate	0.345	0.241	30.1#	128	0.00
31	Indeno(1,2,3-cd)pyrene	1.181	1.187	-0.5	147	-0.01
32 I	Perylene-d12	1.000	1.000	0.0	147	-0.01
33	Benzo(b)fluoranthene	1.084	0.993	8.4	139	0.00
34	Benzo(k)fluoranthene	1.263	1.114	11.8	132	0.00
35 C	Benzo(a)pyrene	0.970	0.882	9.1	143	0.00
36	Dibenzo(a,h)anthracene	1.020	0.959	6.0	144	-0.02
37	Benzo(g,h,i)perylene	1.116	0.958	14.2	134	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0