

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE091417\
 Data File : BE094012.D
 Acq On : 14 Sep 2017 19:07
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 15 01:53:26 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE090517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 05 13:34:56 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2	1,4-Dioxane	0.551	0.582	-5.6	113	0.00
3	n-Nitrosodimethylamine	0.646	0.675	-4.5	110	0.00
4 S	2-Fluorophenol	1.294	1.313	-1.5	108	-0.02
5 S	Phenol-d6	1.941	1.926	0.8	108	0.00
6	bis(2-Chloroethyl)ether	1.587	1.563	1.5	104	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
8 S	Nitrobenzene-d5	0.338	0.337	0.3	106	0.00
9	Naphthalene	4.585	4.562	0.5	104	0.00
10	Hexachlorobutadiene	0.167	0.158	5.4	100	-0.02
11 SURR	2-Methylnaphthalene-d10	0.640	0.651	-1.7	105	0.00
12	2-Methylnaphthalene	0.762	0.755	0.9	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
14 S	2,4,6-Tribromophenol	0.124	0.123	0.8	104	0.00
15 S	2-Fluorobiphenyl	1.582	1.550	2.0	104	0.00
16	Acenaphthylene	1.997	1.978	1.0	110	0.00
17	Acenaphthene	1.349	1.345	0.3	106	0.00
18	Fluorene	1.741	1.729	0.7	109	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
20	4-Bromophenyl-phenylether	0.219	0.221	-0.9	90	0.00
21	Hexachlorobenzene	0.236	0.192	18.6	80	0.00
22	Pentachlorophenol	0.063	0.071	-12.7	118	0.00
23	Phenanthrene	1.293	1.254	3.0	90	0.00
24	Anthracene	1.218	1.248	-2.5	105	0.00
25 SURR	Fluoranthene-d10	5.960	5.939	0.4	107	0.00
26	Fluoranthene	1.783	1.757	1.5	107	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	104	-0.01
28	Pvrene	1.376	1.394	-1.3	106	0.00
29 S	Terphenyl-d14	0.736	0.745	-1.2	105	0.00
30	Benzo(a)anthracene	1.229	1.152	6.3	97	0.00
31	Chrysene	1.300	1.309	-0.7	106	-0.01
32	Bis(2-ethylhexyl)phthalate	2.983	2.750	7.8	101	-0.01
33	Indeno(1,2,3-cd)pyrene	0.867	0.637	26.5#	78	0.00
34 I	Perylene-d12	1.000	1.000	0.0	94	-0.01
35	Benzo(b)fluoranthene	1.274	1.270	0.3	94	-0.01
36	Benzo(k)fluoranthene	1.561	1.647	-5.5	100	0.00
37 C	Benzo(a)pyrene	1.297	1.284	1.0	94	0.00
38	Dibenzo(a,h)anthracene	0.860	0.696	19.1	76	0.00
39	Benzo(a,h,i)perylene	0.907	0.770	15.1	80	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE091417\
Data File : BE094012.D
Acq On : 14 Sep 2017 19:07
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 15 01:53:26 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE090517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 05 13:34:56 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			