

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE091517\
 Data File : BE094014.D
 Acq On : 15 Sep 2017 10:22
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 18 14:29:58 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 05 13:34:56 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	113	0.00
2	1,4-Dioxane	0.551	0.538	2.4	112	0.00
3	n-Nitrosodimethylamine	0.646	0.684	-5.9	120	-0.01
4 S	2-Fluorophenol	1.294	1.300	-0.5	115	-0.01
5 S	Phenol-d6	1.941	1.858	4.3	112	-0.01
6	bis(2-Chloroethyl)ether	1.587	1.564	1.4	112	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
8 S	Nitrobenzene-d5	0.338	0.334	1.2	113	0.00
9	Naphthalene	4.585	4.595	-0.2	113	0.00
10	Hexachlorobutadiene	0.167	0.163	2.4	111	-0.02
11 SURR	2-Methylnaphthalene-d10	0.640	0.648	-1.3	113	0.00
12	2-Methylnaphthalene	0.762	0.759	0.4	112	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	113	0.00
14 S	2,4,6-Tribromophenol	0.124	0.109	12.1	96	0.00
15 S	2-Fluorobiphenyl	1.582	1.604	-1.4	112	0.00
16	Acenaphthylene	1.997	1.899	4.9	111	0.00
17	Acenaphthene	1.349	1.341	0.6	111	0.00
18	Fluorene	1.741	1.716	1.4	113	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	108	0.00
20	4-Bromophenyl-phenylether	0.219	0.230	-5.0	96	0.00
21	Hexachlorobenzene	0.236	0.189	19.9	81	0.00
22	Pentachlorophenol	0.063	0.066	-4.8	112	0.00
23	Phenanthrene	1.293	1.319	-2.0	97	0.00
24	Anthracene	1.218	1.320	-8.4	115	0.00
25 SURR	Fluoranthene-d10	5.960	5.746	3.6	107	0.00
26	Fluoranthene	1.783	1.689	5.3	106	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	104	-0.01
28	Pyrene	1.376	1.403	-2.0	106	0.00
29 S	Terphenyl-d14	0.736	0.744	-1.1	105	0.00
30	Benzo(a)anthracene	1.229	1.107	9.9	93	0.00
31	Chrysene	1.300	1.352	-4.0	109	0.00
32	Bis(2-ethylhexyl)phthalate	2.983	2.570	13.8	94	0.00
33	Indeno(1,2,3-cd)pyrene	0.867	0.724	16.5	88	0.00
34 I	Perylene-d12	1.000	1.000	0.0	90	0.00
35	Benzo(b)fluoranthene	1.274	1.265	0.7	89	0.00
36	Benzo(k)fluoranthene	1.561	1.751	-12.2	101	0.00
37 C	Benzo(a)pyrene	1.297	1.290	0.5	90	0.00
38	Dibenzo(a,h)anthracene	0.860	0.704	18.1	73	0.00
39	Benzo(g,h,i)perylene	0.907	0.839	7.5	84	0.00

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE091517\
Data File : BE094014.D
Acq On : 15 Sep 2017 10:22
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 18 14:29:58 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 05 13:34:56 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		