

Data Path : Z:\HPCHEM1\BNA_E\Data\BE091517\
 Data File : BE094018.D
 Acq On : 15 Sep 2017 16:59
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 18 14:42:54 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 05 13:34:56 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	120	0.00
2	1,4-Dioxane	0.400	0.456	-14.0	139	0.00
3	n-Nitrosodimethylamine	0.400	0.398	0.5	120	0.00
4 S	2-Fluorophenol	0.400	0.386	3.5	118	-0.01
5 S	Phenol-d6	0.400	0.378	5.5	117	0.00
6	bis(2-Chloroethyl)ether	0.400	0.399	0.3	121	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	119	0.00
8 S	Nitrobenzene-d5	0.400	0.390	2.5	117	0.00
9	Naphthalene	0.400	0.402	-0.5	118	0.00
10	Hexachlorobutadiene	0.400	0.383	4.3	114	-0.02
11 SURR	2-Methylnaphthalene-d10	0.400	0.404	-1.0	118	0.00
12	2-Methylnaphthalene	0.400	0.401	-0.3	117	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	116	0.00
14 S	2,4,6-Tribromophenol	0.400	0.346	13.5	97	0.00
15 S	2-Fluorobiphenyl	0.400	0.410	-2.5	116	0.00
16	Acenaphthylene	0.400	0.389	2.8	116	0.00
17	Acenaphthene	0.400	0.403	-0.8	115	0.00
18	Fluorene	0.400	0.395	1.3	116	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	114	0.00
20	4-Bromophenyl-phenylether	0.400	0.393	1.8	101	0.00
21	Hexachlorobenzene	0.400	0.321	19.8	85	0.00
22	Pentachlorophenol	0.400	0.357	10.8	101	0.00
23	Phenanthrene	0.400	0.358	10.5	97	0.00
24	Anthracene	0.400	0.399	0.3	111	0.00
25 SURR	Fluoranthene-d10	0.400	0.368	8.0	107	0.00
26	Fluoranthene	0.400	0.367	8.3	108	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	104	-0.01
28	Pyrene	0.400	0.410	-2.5	107	0.00
29 S	Terphenyl-d14	0.400	0.402	-0.5	104	0.00
30	Benzo(a)anthracene	0.400	0.359	10.3	92	0.00
31	Chrysene	0.400	0.416	-4.0	109	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.317	20.8	86	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.299	25.3#	79	0.00
34 I	Perylene-d12	0.400	0.400	0.0	88	0.00
35	Benzo(b)fluoranthene	0.400	0.384	4.0	84	0.00
36	Benzo(k)fluoranthene	0.400	0.453	-13.2	100	0.00
37 C	Benzo(a)pyrene	0.400	0.401	-0.3	89	0.00
38	Dibenzo(a,h)anthracene	0.400	0.334	16.5	73	0.00
39	Benzo(g,h,i)perylene	0.400	0.358	10.5	79	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE091517\
Data File : BE094018.D
Acq On : 15 Sep 2017 16:59
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 18 14:42:54 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 05 13:34:56 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				