

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE091616\
 Data File : BE092384.D
 Acq On : 16 Sep 2016 9:21
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 16 11:17:59 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 19:34:51 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	127	-0.02
2	1,4-Dioxane	0.707	0.676	4.4	117	-0.01
3	n-Nitrosodimethylamine	0.661	0.596	9.8	136	-0.02
4 S	2-Fluorophenol	1.189	1.063	10.6	133	0.00
5 S	Phenol-d6	1.586	1.397	11.9	145	0.00
6	bis(2-Chloroethyl)ether	1.288	1.150	10.7	143	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	132	0.00
8 S	Nitrobenzene-d5	0.359	0.348	3.1	162#	0.00
9	Naphthalene	1.120	1.148	-2.5	142	0.00
10	Hexachlorobutadiene	0.175	0.186	-6.3	142	0.00
11	2-Methylnaphthalene	0.733	0.754	-2.9	146	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	136	0.00
13 S	2,4,6-Tribromophenol	0.155	0.158	-1.9	173#	0.00
14 S	2-Fluorobiphenyl	1.534	1.514	1.3	149	0.00
15	Acenaphthylene	8.606	8.674	-0.8	153#	0.00
16	Acenaphthene	1.393	1.418	-1.8	150#	0.00
17	Fluorene	1.749	1.776	-1.5	153#	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	140	0.02
19	Hexachlorobenzene	0.239	0.208	13.0	134	0.00
20	Pentachlorophenol	0.053	0.061	-15.1	238#	0.00
21	Phenanthrene	1.405	1.198	14.7	132	0.02
22	Anthracene	1.250	1.101	11.9	149	0.02
23	Fluoranthene	6.135	5.712	6.9	148	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	142	0.00
25	Pyrene	4.864	4.844	0.4	148	0.02
26 S	Terphenyl-d14	0.740	0.698	5.7	142	0.02
27	Benzo(a)anthracene	5.396	5.527	-2.4	148	0.00
28	Chrysene	5.570	5.356	3.8	151#	0.02
29	Bis(2-ethylhexyl)phthalate	0.749	0.565	24.6	129	0.00
30	Indeno(1,2,3-cd)pyrene	5.684	5.428	4.5	138	0.05
31 I	Perylene-d12	1.000	1.000	0.0	124	0.04
32	Benzo(b)fluoranthene	1.284	1.409	-9.7	154#	0.02
33	Benzo(k)fluoranthene	1.357	1.256	7.4	119	0.04
34 C	Benzo(a)pyrene	1.265	1.290	-2.0	142	0.02
35	Dibenzo(a,h)anthracene	1.170	1.178	-0.7	141	0.05
36	Benzo(g,h,i)perylene	4.933	5.045	-2.3	136	0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0