

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE091616\
 Data File : BE092384.D
 Acq On : 16 Sep 2016 9:21
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 16 11:17:59 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 02 19:34:51 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	127	-0.02
2	1,4-Dioxane	0.400	0.353	11.8	117	-0.01
3	n-Nitrosodimethylamine	0.400	0.372	7.0	136	-0.02
4 S	2-Fluorophenol	0.400	0.358	10.5	133	0.00
5 S	Phenol-d6	0.400	0.352	12.0	145	0.00
6	bis(2-Chloroethyl)ether	0.400	0.370	7.5	143	-0.02
7 I	Naphthalene-d8	5.000	5.000	0.0	132	0.00
8 S	Nitrobenzene-d5	0.400	0.388	3.0	162	0.00
9	Naphthalene	0.400	0.410	-2.5	142	0.00
10	Hexachlorobutadiene	0.400	0.426	-6.5	142	0.00
11	2-Methylnaphthalene	0.400	0.411	-2.7	146	-0.01
12 I	Acenaphthene-d10	5.000	5.000	0.0	136	0.00
13 S	2,4,6-Tribromophenol	0.400	0.409	-2.2	173	0.00
14 S	2-Fluorobiphenyl	0.400	0.395	1.3	149	0.00
15	Acenaphthylene	0.400	0.403	-0.8	153	0.00
16	Acenaphthene	0.400	0.407	-1.7	150	0.00
17	Fluorene	0.400	0.406	-1.5	153	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	140	0.02
19	Hexachlorobenzene	0.400	0.347	13.3	134	0.00
20	Pentachlorophenol	0.400	0.514	-28.5#	238	0.00
21	Phenanthrene	0.400	0.341	14.8	132	0.02
22	Anthracene	0.400	0.352	12.0	149	0.02
23	Fluoranthene	0.400	0.372	7.0	148	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	142	0.00
25	Pyrene	0.400	0.398	0.5	148	0.02
26 S	Terphenyl-d14	0.400	0.378	5.5	142	0.02
27	Benzo(a)anthracene	0.400	0.410	-2.5	148	0.00
28	Chrysene	0.400	0.323	19.3	151	0.02
29	Bis(2-ethylhexyl)phthalate	0.400	0.312	22.0	129	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.382	4.5	138	0.05
31 I	Perylene-d12	5.000	5.000	0.0	124	0.04
32	Benzo(b)fluoranthene	0.400	0.439	-9.7	154	0.02
33	Benzo(k)fluoranthene	0.400	0.370	7.5	119	0.04
34 C	Benzo(a)pyrene	0.400	0.408	-2.0	142	0.02
35	Dibenzo(a,h)anthracene	0.400	0.403	-0.8	141	0.05
36	Benzo(g,h,i)perylene	0.400	0.409	-2.2	136	0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0