

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE091715\
 Data File : BE090857.D
 Acq On : 17 Sep 2015 9:41
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 17 11:34:18 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 11:40:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	73	0.00
2	1,4-Dioxane	1.000	0.945	5.5	65	-0.01
3	n-Nitrosodimethylamine	1.000	0.954	4.6	67	0.00
4 S	2-Fluorophenol	1.000	1.072	-7.2	77	0.00
5 S	Phenol-d6	1.000	1.010	-1.0	77	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	82	0.00
7 S	Nitrobenzene-d5	1.000	0.916	8.4	75	0.00
8	Nitrobenzene	1.000	0.902	9.8	77	0.00
9	Naphthalene	1.000	0.947	5.3	78	-0.02
10	Hexachlorobutadiene	1.000	0.908	9.2	72	0.00
11	2-Methylnaphthalene	1.000	1.038	-3.8	88	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	93	0.00
13 S	2,4,6-Tribromophenol	1.000	0.977	2.3	95	0.00
14 S	2-Fluorobiphenyl	1.000	0.901	9.9	85	0.00
15	Acenaphthylene	1.000	0.953	4.7	92	0.00
16	Acenaphthene	1.000	0.937	6.3	86	0.00
17	Fluorene	1.000	0.960	4.0	90	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	86	0.00
19	4-Bromophenyl-phenylether	1.000	0.958	4.2	86	-0.02
20	Hexachlorobenzene	1.000	0.971	2.9	82	0.00
21	Pentachlorophenol	1.000	1.014	-1.4	74	0.00
22	Phenanthrene	1.000	0.973	2.7	84	-0.02
23	Anthracene	1.000	0.993	0.7	89	0.00
24	Fluoranthene	1.000	0.942	5.8	88	-0.01
25 I	Chrysene-d12	5.000	5.000	0.0	82	0.00
26	Pyrene	1.000	1.035	-3.5	89	-0.01
27 S	Terphenyl-d14	1.000	1.019	-1.9	87	0.00
28	Benzo(a)anthracene	1.000	0.980	2.0	81	0.00
29	Chrysene	1.000	1.001	-0.1	79	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.962	3.8	88	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	0.917	8.3	81	-0.02
32 I	Perylene-d12	5.000	5.000	0.0	82	-0.01
33	Benzo(b)fluoranthene	1.000	0.979	2.1	83	0.00
34	Benzo(k)fluoranthene	1.000	0.947	5.3	79	0.00
35 C	Benzo(a)pyrene	1.000	0.972	2.8	85	-0.01
36	Dibenzo(a,h)anthracene	1.000	0.921	7.9	80	-0.02
37	Benzo(g,h,i)perylene	1.000	0.946	5.4	82	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0