

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE091715\
 Data File : BE090857.D
 Acq On : 17 Sep 2015 9:41
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 17 11:34:18 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 11:40:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	73	0.00
2	1,4-Dioxane	0.814	0.709	12.9	65	-0.01
3	n-Nitrosodimethylamine	1.033	0.904	12.5	67	0.00
4 S	2-Fluorophenol	1.039	1.029	1.0	77	0.00
5 S	Phenol-d6	1.392	1.406	-1.0	77	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	82	0.00
7 S	Nitrobenzene-d5	0.330	0.303	8.2	75	0.00
8	Nitrobenzene	0.365	0.329	9.9	77	0.00
9	Naphthalene	1.091	1.033	5.3	78	-0.02
10	Hexachlorobutadiene	0.171	0.155	9.4	72	0.00
11	2-Methylnaphthalene	0.590	0.613	-3.9	88	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.00
13 S	2,4,6-Tribromophenol	0.143	0.123	14.0	95	0.00
14 S	2-Fluorobiphenyl	1.385	1.248	9.9	85	0.00
15	Acenaphthylene	7.992	7.619	4.7	92	0.00
16	Acenaphthene	1.335	1.251	6.3	86	0.00
17	Fluorene	1.666	1.598	4.1	90	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	86	0.00
19	4-Bromophenyl-phenylether	0.170	0.163	4.1	86	-0.02
20	Hexachlorobenzene	0.928	0.872	6.0	82	0.00
21	Pentachlorophenol	0.054	0.041	24.1	74	0.00
22	Phenanthrene	1.237	1.205	2.6	84	-0.02
23	Anthracene	1.035	1.027	0.8	89	0.00
24	Fluoranthene	1.304	1.229	5.8	88	-0.01
25 I	Chrysene-d12	1.000	1.000	0.0	82	0.00
26	Pyrene	1.035	1.071	-3.5	89	-0.01
27 S	Terphenyl-d14	0.553	0.564	-2.0	87	0.00
28	Benzo(a)anthracene	1.156	1.088	5.9	81	0.00
29	Chrysene	1.279	1.223	4.4	79	0.00
30	Bis(2-ethylhexyl)phthalate	0.345	0.273	20.9	88	0.00
31	Indeno(1,2,3-cd)pyrene	1.181	1.083	8.3	81	-0.02
32 I	Perylene-d12	1.000	1.000	0.0	82	-0.01
33	Benzo(b)fluoranthene	1.084	1.061	2.1	83	0.00
34	Benzo(k)fluoranthene	1.263	1.196	5.3	79	0.00
35 C	Benzo(a)pyrene	0.970	0.943	2.8	85	-0.01
36	Dibenzo(a,h)anthracene	1.020	0.951	6.8	80	-0.02
37	Benzo(g,h,i)perylene	1.116	1.056	5.4	82	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0