

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE091715\
 Data File : BE090874.D
 Acq On : 17 Sep 2015 20:56
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 18 02:29:53 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE090115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 02 11:40:50 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	68	0.00
2	1,4-Dioxane	1.000	0.949	5.1	61	0.00
3	n-Nitrosodimethylamine	1.000	0.827	17.3	55	0.00
4 S	2-Fluorophenol	1.000	0.959	4.1	64	0.00
5 S	Phenol-d6	1.000	0.938	6.2	67	0.01
6 I	Naphthalene-d8	5.000	5.000	0.0	78	0.00
7 S	Nitrobenzene-d5	1.000	0.885	11.5	69	0.00
8	Nitrobenzene	1.000	0.860	14.0	69	0.00
9	Naphthalene	1.000	0.946	5.4	74	-0.01
10	Hexachlorobutadiene	1.000	0.913	8.7	69	0.00
11	2-Methylnaphthalene	1.000	1.012	-1.2	82	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	86	0.00
13 S	2,4,6-Tribromophenol	1.000	0.841	15.9	73	0.00
14 S	2-Fluorobiphenyl	1.000	0.903	9.7	79	0.00
15	Acenaphthylene	1.000	0.947	5.3	85	0.00
16	Acenaphthene	1.000	0.963	3.7	82	0.00
17	Fluorene	1.000	0.913	8.7	79	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	82	0.00
19	4-Bromophenyl-phenylether	1.000	0.936	6.4	80	0.00
20	Hexachlorobenzene	1.000	0.983	1.7	79	0.00
21	Pentachlorophenol	1.000	0.918	8.2	59	0.02
22	Phenanthrene	1.000	0.965	3.5	80	-0.02
23	Anthracene	1.000	0.927	7.3	79	0.00
24	Fluoranthene	1.000	0.881	11.9	78	-0.01
25 I	Chrysene-d12	5.000	5.000	0.0	74	0.00
26	Pyrene	1.000	1.013	-1.3	80	0.00
27 S	Terphenyl-d14	1.000	0.989	1.1	77	0.00
28	Benzo(a)anthracene	1.000	0.962	3.8	73	0.00
29	Chrysene	1.000	1.000	0.0	72	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.830	17.0	68	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	0.847	15.3	68	0.00
32 I	Perylene-d12	5.000	5.000	0.0	73	-0.01
33	Benzo(b)fluoranthene	1.000	0.940	6.0	71	0.00
34	Benzo(k)fluoranthene	1.000	0.947	5.3	71	0.00
35 C	Benzo(a)pyrene	1.000	0.921	7.9	72	-0.01
36	Dibenzo(a,h)anthracene	1.000	0.885	11.5	68	0.00
37	Benzo(g,h,i)perylene	1.000	0.884	11.6	69	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0