

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE091724\
 Data File : BE101344.D
 Acq On : 17 Sep 2024 9:41
 Operator : RC/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC2.5

Quant Time: Sep 17 15:40:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-BE091724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 17 15:24:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.576	152	153355	20.000	ng	0.00
21) Naphthalene-d8	10.343	136	740085	20.000	ng	0.00
39) Acenaphthene-d10	14.185	164	543755	20.000	ng	0.00
64) Phenanthrene-d10	16.923	188	1285990	20.000	ng	0.00
76) Chrysene-d12	21.083	240	1511159	20.000	ng	0.00
86) Perylene-d12	23.375	264	1904881	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						
19) n-Nitroso-di-n-propyla...	8.392	70	17894	2.039	ng	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE091724\
Data File : BE101344.D
Acq On : 17 Sep 2024 9:41
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDICC2.5

Quant Time: Sep 17 15:40:23 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-BE091724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 17 15:24:04 2024
Response via : Initial Calibration



Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE091724\
Data File : BE101344.D
Acq On : 17 Sep 2024 9:41
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDICC2.5

Quant Time: Sep 17 15:40:23 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-BE091724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 17 15:24:04 2024
Response via : Initial Calibration

