

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE091724\
 Data File : BE101352.D
 Acq On : 17 Sep 2024 15:34
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE091724

Quant Time: Sep 17 16:29:59 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-BE091724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 17 16:03:42 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	97	0.00
2	1,4-Dioxane	40.000	39.410	1.5	98	0.00
3	Pyridine	40.000	40.556	-1.4	94	0.00
4	n-Nitrosodimethylamine	40.000	42.680	-6.7	103	0.00
5 S	2-Fluorophenol	80.000	80.523	-0.7	98	0.00
6	Aniline	40.000	39.656	0.9	98	0.00
7 S	Phenol-d6	80.000	83.654	-4.6	99	0.00
8	2-Chlorophenol	40.000	40.430	-1.1	98	0.00
9	Benzaldehyde	40.000	45.088	-12.7	104	0.00
10 C	Phenol	40.000	42.225	-5.6	100	0.00
11	bis(2-Chloroethyl)ether	40.000	34.679	13.3	80	0.00
12	1,3-Dichlorobenzene	40.000	39.174	2.1	97	0.00
13 C	1,4-Dichlorobenzene	40.000	39.159	2.1	97	0.00
14	1,2-Dichlorobenzene	40.000	39.472	1.3	97	0.00
15	Benzyl Alcohol	40.000	42.944	-7.4	96	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.887	0.3	98	0.00
17	2-Methylphenol	40.000	41.768	-4.4	99	0.00
18	Hexachloroethane	40.000	40.512	-1.3	98	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	43.121	-7.8	98	0.00
20	3+4-Methylphenols	40.000	42.604	-6.5	100	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	97	0.00
22	Acetophenone	40.000	41.315	-3.3	98	0.00
23 S	Nitrobenzene-d5	80.000	84.464	-5.6	99	0.00
24	Nitrobenzene	40.000	41.435	-3.6	98	0.00
25	Isophorone	40.000	41.688	-4.2	97	0.00
26 C	2-Nitrophenol	40.000	43.255	-8.1	99	0.00
27	2,4-Dimethylphenol	40.000	41.352	-3.4	97	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.058	-2.6	97	0.00
29 C	2,4-Dichlorophenol	40.000	42.008	-5.0	98	0.00
30	1,2,4-Trichlorobenzene	40.000	39.985	0.0	98	0.00
31	Naphthalene	40.000	40.065	-0.2	98	0.00
32	Benzoic acid	40.000	40.056	-0.1	101	0.00
33	4-Chloroaniline	40.000	42.839	-7.1	100	0.00
34 C	Hexachlorobutadiene	40.000	39.457	1.4	96	0.00
35	Caprolactam	40.000	44.246	-10.6	100	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.369	-5.9	99	0.00
37	2-Methylnaphthalene	40.000	40.630	-1.6	98	0.00
38	1-Methylnaphthalene	40.000	40.543	-1.4	97	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	98	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.565	1.1	97	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.137	2.2	90	0.00
42 S	2,4,6-Tribromophenol	80.000	82.047	-2.6	98	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.523	-3.8	98	0.00
44	2,4,5-Trichlorophenol	40.000	41.331	-3.3	98	0.00
45 S	2-Fluorobiphenyl	80.000	79.686	0.4	97	0.00
46	1,1'-Biphenyl	40.000	39.570	1.1	97	0.00
47	2-Chloronaphthalene	40.000	39.751	0.6	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE091724\
 Data File : BE101352.D
 Acq On : 17 Sep 2024 15:34
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE091724

Quant Time: Sep 17 16:29:59 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-BE091724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 17 16:03:42 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.776	-9.4	98	0.00
49	Acenaphthylene	40.000	40.545	-1.4	99	0.00
50	Dimethylphthalate	40.000	39.416	1.5	97	0.00
51	2,6-Dinitrotoluene	40.000	41.783	-4.5	99	0.00
52 C	Acenaphthene	40.000	39.717	0.7	97	0.00
53	3-Nitroaniline	40.000	43.192	-8.0	100	0.00
54 P	2,4-Dinitrophenol	40.000	42.766	-6.9	100	0.00
55	Dibenzofuran	40.000	39.703	0.7	98	0.00
56 P	4-Nitrophenol	40.000	45.218	-13.0	104	0.00
57	2,4-Dinitrotoluene	40.000	42.835	-7.1	100	0.00
58	Fluorene	40.000	39.910	0.2	98	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.892	-2.2	98	0.00
60	Diethylphthalate	40.000	39.778	0.6	97	0.00
61	4-Chlorophenyl-phenylether	40.000	39.683	0.8	97	0.00
62	4-Nitroaniline	40.000	44.523	-11.3	100	0.00
63	Azobenzene	40.000	40.128	-0.3	97	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	98	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.533	-6.3	100	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.518	-1.3	97	0.00
67	4-Bromophenyl-phenylether	40.000	39.976	0.1	96	0.00
68	Hexachlorobenzene	40.000	40.527	-1.3	97	0.00
69	Atrazine	40.000	39.655	0.9	97	0.00
70 C	Pentachlorophenol	40.000	43.810	-9.5	100	0.00
71	Phenanthrene	40.000	39.900	0.3	97	0.00
72	Anthracene	40.000	40.656	-1.6	97	0.00
73	Carbazole	40.000	40.822	-2.1	98	0.00
74	Di-n-butylphthalate	40.000	42.382	-6.0	99	0.00
75 C	Fluoranthene	40.000	40.499	-1.2	98	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	98	0.00
77	Benzidine	40.000	61.193	-53.0#	127	0.00
78	Pyrene	40.000	40.669	-1.7	99	0.00
79 S	Terphenyl-d14	80.000	67.895	15.1	98	0.00
80	Butylbenzylphthalate	40.000	41.927	-4.8	98	0.00
81	Benzo(a)anthracene	40.000	40.111	-0.3	97	0.00
82	3,3'-Dichlorobenzidine	40.000	42.782	-7.0	97	0.00
83	Chrysene	40.000	39.915	0.2	99	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.046	-7.6	99	0.00
85 c	Di-n-octyl phthalate	40.000	41.721	-4.3	98	0.00
86 I	Perylene-d12	20.000	20.000	0.0	97	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.682	-1.7	97	0.01
88	Benzo(b)fluoranthene	40.000	40.831	-2.1	99	0.00
89	Benzo(k)fluoranthene	40.000	40.183	-0.5	96	0.00
90 C	Benzo(a)pyrene	40.000	40.927	-2.3	98	0.00
91	Dibenzo(a,h)anthracene	40.000	40.923	-2.3	96	0.00
92	Benzo(g,h,i)perylene	40.000	40.519	-1.3	97	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE091724\
Data File : BE101352.D
Acq On : 17 Sep 2024 15:34
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE091724

Quant Time: Sep 17 16:29:59 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-BE091724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 17 16:03:42 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0