

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE091815\
 Data File : BE090890.D
 Acq On : 18 Sep 2015 17:49
 Operator : UM/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE091815

Quant Time: Sep 18 18:33:10 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 18 18:17:34 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	102	0.00
2	1,4-Dioxane	1.000	0.991	0.9	93	0.00
3	n-Nitrosodimethylamine	1.000	0.964	3.6	92	0.00
4 S	2-Fluorophenol	1.000	0.918	8.2	90	0.00
5 S	Phenol-d6	1.000	0.886	11.4	93	0.00
6	bis(2-Chloroethyl)ether	1.000	1.015	-1.5	105	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	101	0.00
8 S	Nitrobenzene-d5	1.000	0.920	8.0	94	0.00
9	Nitrobenzene	1.000	0.907	9.3	95	0.00
10	Naphthalene	1.000	0.976	2.4	94	0.00
11	Hexachlorobutadiene	1.000	0.982	1.8	95	0.00
12	2-Methylnaphthalene	1.000	0.918	8.2	95	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	103	0.00
14 S	2,4,6-Tribromophenol	1.000	0.853	14.7	90	-0.02
15 S	2-Fluorobiphenyl	1.000	0.939	6.1	96	0.00
16	Acenaphthylene	1.000	0.916	8.4	95	0.00
17	Acenaphthene	1.000	0.930	7.0	97	0.00
18	Fluorene	1.000	0.952	4.8	97	0.00
19 I	Phenanthrene-d10	5.000	5.000	0.0	102	0.00
20	4-Bromophenyl-phenylether	1.000	0.930	7.0	96	-0.02
21	Hexachlorobenzene	1.000	0.987	1.3	96	0.00
22	Pentachlorophenol	1.000	0.800	20.0	77	0.00
23	Phenanthrene	1.000	1.009	-0.9	99	0.00
24	Anthracene	1.000	0.901	9.9	95	0.00
25	Fluoranthene	1.000	0.898	10.2	93	0.00
26 I	Chrysene-d12	5.000	5.000	0.0	95	0.00
27	Pyrene	1.000	0.903	9.7	92	0.00
28 S	Terphenyl-d14	1.000	0.920	8.0	91	0.00
29	Benzo(a)anthracene	1.000	0.875	12.5	88	0.00
30	Chrysene	1.000	0.965	3.5	88	0.00
31	Bis(2-ethylhexyl)phthalate	1.000	0.768	23.2	75	0.00
32	Indeno(1,2,3-cd)pyrene	1.000	0.871	12.9	83	0.00
33 I	Perylene-d12	5.000	5.000	0.0	92	0.00
34	Benzo(b)fluoranthene	1.000	0.868	13.2	83	0.00
35	Benzo(k)fluoranthene	1.000	0.923	7.7	86	0.00
36 C	Benzo(a)pyrene	1.000	0.866	13.4	83	0.00
37	Dibenzo(a,h)anthracene	1.000	0.872	12.8	81	0.00
38	Benzo(g,h,i)perylene	1.000	0.907	9.3	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0