

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE091815\
 Data File : BE090892.D
 Acq On : 19 Sep 2015 2:25
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 19 03:47:26 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 18 18:17:34 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	113	0.00
2	1,4-Dioxane	0.795	0.724	8.9	101	0.00
3	n-Nitrosodimethylamine	1.067	0.891	16.5	97	0.00
4 S	2-Fluorophenol	1.238	1.009	18.5	93	0.00
5 S	Phenol-d6	1.559	1.248	19.9	93	0.00
6	bis(2-Chloroethyl)ether	1.546	1.385	10.4	111	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	113	0.00
8 S	Nitrobenzene-d5	0.345	0.299	13.3	99	0.00
9	Nitrobenzene	0.383	0.344	10.2	105	0.00
10	Naphthalene	1.126	1.033	8.3	105	0.00
11	Hexachlorobutadiene	0.166	0.165	0.6	111	0.00
12	2-Methylnaphthalene	0.672	0.605	10.0	104	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	118	0.00
14 S	2,4,6-Tribromophenol	0.196	0.135	31.1#	93	-0.02
15 S	2-Fluorobiphenyl	1.418	1.305	8.0	107	0.00
16	Acenaphthylene	8.705	7.758	10.9	106	0.00
17	Acenaphthene	1.341	1.257	6.3	111	0.00
18	Fluorene	1.705	1.550	9.1	110	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	123	0.00
20	4-Bromophenyl-phenylether	0.193	0.170	11.9	109	-0.02
21	Hexachlorobenzene	0.906	0.836	7.7	113	0.00
22	Pentachlorophenol	0.087	0.056	35.6#	86	0.00
23	Phenanthrene	1.292	1.148	11.1	112	0.00
24	Anthracene	1.144	0.977	14.6	108	0.00
25	Fluoranthene	1.525	1.327	13.0	108	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	113	0.00
27	Pyrene	1.141	1.018	10.8	108	0.00
28 S	Terphenyl-d14	0.638	0.540	15.4	105	0.00
29	Benzo(a)anthracene	1.235	1.123	9.1	108	0.00
30	Chrysene	1.352	1.248	7.7	105	0.00
31	Bis(2-ethylhexyl)phthalate	0.515	0.379	26.4#	95	0.00
32	Indeno(1,2,3-cd)pyrene	1.455	1.257	13.6	97	0.00
33 I	Perylene-d12	1.000	1.000	0.0	110	0.00
34	Benzo(b)fluoranthene	1.170	1.060	9.4	102	0.00
35	Benzo(k)fluoranthene	1.347	1.214	9.9	99	0.00
36 C	Benzo(a)pyrene	1.113	0.972	12.7	99	0.00
37	Dibenzo(a,h)anthracene	1.158	1.013	12.5	97	0.00
38	Benzo(g,h,i)perylene	1.240	1.104	11.0	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0