

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE091815\
 Data File : BE090882.D
 Acq On : 18 Sep 2015 12:50
 Operator : UM/IZ
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Manual Integrations
 APPROVED

MMDadoda
 9/21/2015 2:24:08 PM

Quant Time: Sep 18 17:48:45 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 18 15:49:17 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.54	152	27290	5.00	ng	0.00
7) Naphthalene-d8	10.30	136	118311	5.00	ng	0.00
13) Acenaphthene-d10	14.16	164	61849	5.00	ng	0.00
19) Phenanthrene-d10	16.91	188	147280	5.00	ng	0.00
26) Chrysene-d12	21.07	240	213284	5.00	ng	0.00
33) Perylene-d12	23.35	264	216708	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.17	112	742	0.13	ng	-0.01
5) Phenol-d6	6.75	99	860	0.11	ng	0.00
8) Nitrobenzene-d5	8.70	82	780	0.10	ng	0.00
14) 2,4,6-Tribromophenol	15.67	330	329	0.17	ng	-0.02
15) 2-Fluorobiphenyl	12.79	172	1743	0.10	ng	0.00
28) Terphenyl-d14	19.53	244	3486	0.14	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.15	88	455	0.11	ng	# 85
3) n-Nitrosodimethylamine	3.45	42	806	0.14	ng	# 81
6) bis(2-Chloroethyl)ether	6.98	93	921	0.10	ng	# 89
9) Nitrobenzene	8.74	77	836	0.09	ng	98
10) Naphthalene	10.35	128	2942	0.11	ng	# 88
11) Hexachlorobutadiene	10.63	225	405	0.10	ng	99
12) 2-Methylnaphthalene	11.98	142	1599	0.11	ng	# 95
16) Acenaphthylene	13.88	152	10676	0.11	ng	100
17) Acenaphthene	14.23	154	1675	0.10	ng	91
18) Fluorene	15.22	166	2271	0.11	ng	# 98
20) 4-Bromophenyl-phenylether	16.11	248	641	0.12	ng	97
21) Hexachlorobenzene	16.23	284	2806	0.10	ng	94
23) Phenanthrene	16.94	178	4409m	0.12	ng	
24) Anthracene	17.05	178	3480	0.11	ng	98
25) Fluoranthene	18.96	202	4954	0.13	ng	99
27) Pyrene	19.32	202	5246	0.12	ng	99
29) Benzo(a)anthracene	21.05	228	5938	0.12	ng	98
30) Chrysene	21.10	228	6501	0.12	ng	99
31) Bis(2-ethylhexyl)phthalate	21.00	149	2547	0.16	ng	# 98
32) Indeno(1,2,3-cd)pyrene	25.73	276	6399	0.12	ng	98
34) Benzo(b)fluoranthene	22.67	252	5285	0.11	ng	# 85
35) Benzo(k)fluoranthene	22.71	252	6072	0.11	ng	# 86
36) Benzo(a)pyrene	23.25	252	4927	0.11	ng	# 88
37) Dibenzo(a,h)anthracene	25.74	278	4765	0.10	ng	# 83
38) Benzo(g,h,i)perylene	26.45	276	5277	0.11	ng	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_E\DATA\BE091815\
Data File : BE090882.D
Acq On : 18 Sep 2015 12:50
Operator : UM/IZ
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICC0.1

Manual Integrations
APPROVED

MMDadoda
9/21/2015 2:24:08 PM

Quant Time: Sep 18 17:48:45 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 18 15:49:17 2015
Response via : Initial Calibration






#1
 1,4-Dichlorobenzene-d4
 Concen: 5.00 ng
 RT: 7.54 min Scan# 725
 Delta R.T. -0.01 min
 Lab File: BE090882.D
 Acq: 18 Sep 2015 12:50

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1




Tgt Ion	Ratio	Lower	Upper
152	100		
150	156.7	122.2	183.4
115	68.9	51.0	76.4

Manual Integrations
 APPROVED

MMDadoda
 9/21/2015 2:24:08 PM




Abundance Ion 152.00 (151.70 to 152.70): E
 Ion 150.00 (149.70 to 150.70): E
 Ion 115.00 (114.70 to 115.70): E






#2
 1,4-Dioxane
 Concen: 0.11 ng
 RT: 3.15 min Scan# 81
 Delta R.T. -0.01 min
 Lab File: BE090882.D
 Acq: 18 Sep 2015 12:50

Tgt Ion	Ratio	Lower	Upper
88	100		
58	91.9	68.4	102.6
43	54.7	26.9	40.3



Abundance Ion 88.00 (87.70 to 88.70): BE0
 Ion 58.00 (57.70 to 58.70): BE0
 Ion 43.00 (42.70 to 43.70): BE0





#3




n-Nitrosodimethylamine
Concen: 0.14 ng
RT: 3.45 min Scan# 125
Delta R.T. -0.01 min
Lab File: BE090882.D
Acq: 18 Sep 2015 12:50

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED

MMDadoda
9/21/2015 2:24:08 PM





Tgt Ion	Ratio	Lower	Upper
42	100		
74	86.8	83.7	125.5
44	0.0	10.0	15.0#






#4

2-Fluorophenol
Concen: 0.13 ng
RT: 5.17 min Scan# 379
Delta R.T. -0.01 min
Lab File: BE090882.D
Acq: 18 Sep 2015 12:50

Tgt Ion	Ratio	Lower	Upper
112	100		
64	70.1	57.3	85.9
63	42.3	29.2	43.8





#5

Phenol-d6

Concen: 0.11 ng

RT: 6.75 min Scan# 608

Delta R.T. -0.01 min

Lab File: BE090882.D

Acq: 18 Sep 2015 12:50

Tgt Ion:	99	Resp:	860
Ion	Ratio	Lower	Upper
99	100		
42	10.7	16.8	25.2#
71	0.0	28.3	42.5#

Instrument :





BNA_E

ClientSampleId :

SSTDIC0.1

Manual Integrations
APPROVED

MMDadoda
9/21/2015 2:24:08 PM



#6

bis(2-Chloroethyl)ether

Concen: 0.10 ng


RT: 6.98 min Scan# 643


Delta R.T. -0.00 min

Lab File: BE090882.D

Acq: 18 Sep 2015 12:50

Tgt Ion:	93	Resp:	921
Ion	Ratio	Lower	Upper
93	100		
63	76.3	68.5	102.7
95	26.0	26.9	40.3#





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.30 min Scan# 1211

Delta R.T. -0.01 min

Lab File: BE090882.D


Acq: 18 Sep 2015 12:50

Instrument :

BNA_E

ClientSampleId :

SSTDIC0.1



Tgt Ion: 136 Resp: 118311

Ion Ratio Lower Upper

136 100


137 11.2 8.7 13.1

54 12.4 9.0 13.6

68 9.0 6.3 9.5

Manual Integrations
APPROVED

MMDadoda
9/21/2015 2:24:08 PM





Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0

Time-->



#8

Nitrobenzene-d5

Concen: 0.10 ng

RT: 8.70 min Scan# 932

Delta R.T. -0.00 min

Lab File: BE090882.D

Acq: 18 Sep 2015 12:50


Tgt Ion: 82 Resp: 780

Ion Ratio Lower Upper

82 100

128 44.8 25.0 37.4#

54 76.2 46.3 69.5#






Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0

Time-->





#9

Nitrobenzene

Concen: 0.09 ng

RT: 8.74 min Scan# 941

Delta R.T. -0.00 min

Lab File: BE090882.D

Acq: 18 Sep 2015 12:50

Instrument :

BNA_E





ClientSampleId :

SSTDIC0.1

Tgt Ion:	77	Resp:	836
Ion Ratio	Lower	Upper	
77	100		
123	32.2	25.1	37.7
65	11.1	9.9	14.9

Manual Integrations
APPROVED

MMDadoda
9/21/2015 2:24:08 PM



#10

Naphthalene

Concen: 0.11 ng




RT: 10.35 min Scan# 1218


Delta R.T. -0.01 min

Lab File: BE090882.D

Acq: 18 Sep 2015 12:50

Tgt Ion:	128	Resp:	2942
Ion Ratio	Lower	Upper	
128	100		
129	14.9	8.4	12.6#
127	17.2	9.9	14.9#









#11
Hexachlorobutadiene
Concen: 0.10 ng
RT: 10.63 min Scan# 1256
Delta R.T. -0.01 min
Lab File: BE090882.D
Acq: 18 Sep 2015 12:50

Instrument :
BNA_E
ClientSampleId :
SSTDICC0.1

Tgt Ion	Ratio	Lower	Upper
225	100		
223	63.2	50.6	76.0
227	62.2	51.0	76.6




Manual Integrations
APPROVED


MMDadoda
9/21/2015 2:24:08 PM



#12
2-Methylnaphthalene
Concen: 0.11 ng
RT: 11.98 min Scan# 1489
Delta R.T. -0.01 min
Lab File: BE090882.D
Acq: 18 Sep 2015 12:50

Tgt Ion	Ratio	Lower	Upper
142	100		
141	88.6	71.4	107.2
115	46.9	31.0	46.4#









#13
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.16 min Scan# 1915
Delta R.T. -0.01 min
Lab File: BE090882.D
Acq: 18 Sep 2015 12:50

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tgt Ion	Ratio	Lower	Upper
164	100		
162	106.5	86.8	130.2
160	53.4	53.9	80.9#




Manual Integrations
APPROVED


MMDadoda
9/21/2015 2:24:08 PM



#14
2,4,6-Tribromophenol
Concen: 0.17 ng
RT: 15.67 min Scan# 2050
Delta R.T. -0.02 min
Lab File: BE090882.D
Acq: 18 Sep 2015 12:50

Tgt Ion	Ratio	Lower	Upper
330	100		
332	96.0	74.9	112.3
141	203.6	145.2	217.8









#15
2-Fluorobiphenyl
Concen: 0.10 ng
RT: 12.79 min Scan# 1668
Delta R.T. -0.01 min
Lab File: BE090882.D
Acq: 18 Sep 2015 12:50

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tgt Ion	Ratio	Lower	Upper
172	100		
171	39.3	27.8	41.8
170	27.8	19.8	29.6




Manual Integrations
APPROVED


MMDadoda
9/21/2015 2:24:08 PM



#16
Acenaphthylene
Concen: 0.11 ng
RT: 13.88 min Scan# 1872
Delta R.T. 0.00 min
Lab File: BE090882.D
Acq: 18 Sep 2015 12:50

Tgt Ion	Ratio	Lower	Upper
152	100		
151	4.7	3.9	5.9
153	12.9	10.3	15.5





#17

Acenaphthene

Concen: 0.10 ng

RT: 14.23 min Scan# 1925

Delta R.T. 0.00 min

Lab File: BE090882.D


Acq: 18 Sep 2015 12:50

Instrument :

BNA_E

ClientSampleId :

SSTDIC0.1



Tgt Ion:154 Resp: 1675

Ion Ratio Lower Upper




154 100

153 438.9 354.5 531.7

152 247.3 227.8 341.6

Manual Integrations
APPROVED

MMDadoda
9/21/2015 2:24:08 PM



#18

Fluorene

Concen: 0.11 ng

RT: 15.22 min Scan# 2024

Delta R.T. 0.00 min

Lab File: BE090882.D

Acq: 18 Sep 2015 12:50




Tgt Ion:166 Resp: 2271


Ion Ratio Lower Upper

166 100

165 101.7 82.7 124.1

167 16.6 10.7 16.1#





#19

Phenanthrene-d10

Concen: 5.00 ng

RT: 16.91 min Scan# 2121

Delta R.T. 0.00 min

Lab File: BE090882.D

Acq: 18 Sep 2015 12:50

Instrument :

BNA_E





ClientSampleId :

SSTDICC0.1

Tgt Ion:	188	Resp:	147280
Ion Ratio	Lower	Upper	
188	100		
94	9.4	10.3	15.5#
80	10.0	11.0	16.6#

Manual Integrations
APPROVED

MMDadoda
9/21/2015 2:24:08 PM



#20

4-Bromophenyl-phenylether

Concen: 0.12 ng




RT: 16.11 min Scan# 2075


Delta R.T. -0.02 min

Lab File: BE090882.D

Acq: 18 Sep 2015 12:50

Tgt Ion:	248	Resp:	641
Ion Ratio	Lower	Upper	
248	100		
250	88.0	76.7	115.1
141	369.7	292.6	439.0





#21

Hexachlorobenzene

Concen: 0.10 ng

RT: 16.23 min Scan# 2082

Delta R.T. 0.00 min

Lab File: BE090882.D

Acq: 18 Sep 2015 12:50

Instrument :

BNA_E





ClientSampleId :

SSTDICC0.1

Tgt Ion:	284	Resp:	2806
Ion Ratio	Lower	Upper	
284	100		
142	48.8	35.0	52.4
249	30.4	25.5	38.3

Manual Integrations
APPROVED

MMDadoda
9/21/2015 2:24:08 PM



#23

Phenanthrene

Concen: 0.12 ng m




RT: 16.94 min Scan# 2123


Delta R.T. 0.00 min

Lab File: BE090882.D

Acq: 18 Sep 2015 12:50

Tgt Ion:	178	Resp:	4409
Ion Ratio	Lower	Upper	
178	100		
176	11.8	15.6	23.4#
179	9.7	12.8	19.2#





#24

Anthracene

Concen: 0.11 ng

RT: 17.05 min Scan# 2129

Delta R.T. 0.00 min

Lab File: BE090882.D

Acq: 18 Sep 2015 12:50

Instrument :

BNA_E





ClientSampleId :

SSTDIC0.1

Tgt Ion:	178	Resp:	3480
Ion Ratio	Lower	Upper	
178	100		
176	17.1	15.0	22.6
179	15.6	12.3	18.5

Manual Integrations
APPROVED

MMDadoda
9/21/2015 2:24:08 PM



#25

Fluoranthene

Concen: 0.13 ng




RT: 18.96 min Scan# 2346


Delta R.T. -0.00 min

Lab File: BE090882.D

Acq: 18 Sep 2015 12:50

Tgt Ion:	202	Resp:	4954
Ion Ratio	Lower	Upper	
202	100		
101	13.5	11.0	16.6
203	17.6	13.9	20.9






#26
Chrysene-d12
Concen: 5.00 ng
RT: 21.07 min Scan# 2705
Delta R.T. -0.00 min
Lab File: BE090882.D
Acq: 18 Sep 2015 12:50

Instrument :
BNA_E
ClientSampleId :
SSTDICC0.1




Tgt Ion	Ratio	Resp	Lower	Upper
240	100	213284		
120	15.9	10.1	15.1	#
236	27.8	21.9	32.9	

Manual Integrations
APPROVED

MMDadoda
9/21/2015 2:24:08 PM




Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E






#27
Pyrene
Concen: 0.12 ng
RT: 19.32 min Scan# 2414
Delta R.T. -0.00 min
Lab File: BE090882.D
Acq: 18 Sep 2015 12:50

Tgt Ion	Ratio	Resp	Lower	Upper
202	100	5246		
200	21.3	16.7	25.1	
203	17.0	13.8	20.6	



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 200.00 (199.70 to 200.70): E
Ion 203.00 (202.70 to 203.70): E









#28
 Terphenyl-d14
 Concen: 0.14 ng
 RT: 19.53 min Scan# 2455
 Delta R.T. 0.00 min
 Lab File: BE090882.D
 Acq: 18 Sep 2015 12:50

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Tgt Ion	Ratio	Lower	Upper
244	100		
212	10.0	6.9	10.3
122	22.0	12.9	19.3




Manual Integrations
 APPROVED


MMDadoda
 9/21/2015 2:24:08 PM



#29
 Benzo(a)anthracene
 Concen: 0.12 ng
 RT: 21.05 min Scan# 2703
 Delta R.T. -0.00 min
 Lab File: BE090882.D
 Acq: 18 Sep 2015 12:50

Tgt Ion	Ratio	Lower	Upper
228	100		
226	26.2	21.6	32.4
229	21.0	15.9	23.9





#30

Chrysene

Concen: 0.12 ng

RT: 21.10 min Scan# 2710

Delta R.T. -0.00 min

Lab File: BE090882.D

Acq: 18 Sep 2015 12:50

Instrument :

BNA_E





ClientSampleId :

SSTDIC0.1

Tgt	Ion	228	Resp:	6501
	Ratio	100	Lower	Upper
	226	31.0	24.2	36.4
	229	19.6	15.3	22.9

Manual Integrations
APPROVED

MMDadoda
9/21/2015 2:24:08 PM



#31

Bis(2-ethylhexyl)phthalate

Concen: 0.16 ng




RT: 21.00 min Scan# 2695


Delta R.T. -0.00 min

Lab File: BE090882.D

Acq: 18 Sep 2015 12:50

Tgt	Ion	149	Resp:	2547
	Ratio	100	Lower	Upper
	167	26.1	20.1	30.1
	279	3.7	2.3	3.5#









#32
 Indeno(1,2,3-cd)pyrene
 Concen: 0.12 ng
 RT: 25.73 min Scan# 3662
 Delta R.T. 0.02 min
 Lab File: BE090882.D
 Acq: 18 Sep 2015 12:50

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Tgt Ion:	276	Resp:	6399
Ion Ratio	Lower	Upper	
276	100		
138	27.6	23.3	34.9
277	24.8	19.8	29.6




Manual Integrations
 APPROVED


MMDadoda
 9/21/2015 2:24:08 PM



#33
 Perylene-d12
 Concen: 5.00 ng
 RT: 23.35 min Scan# 3143
 Delta R.T. 0.00 min
 Lab File: BE090882.D
 Acq: 18 Sep 2015 12:50

Tgt Ion:	264	Resp:	216708
Ion Ratio	Lower	Upper	
264	100		
260	25.3	19.7	29.5
265	21.7	17.5	26.3








#34
Benzo(b)fluoranthene
Concen: 0.11 ng
RT: 22.67 min Scan# 2993
Delta R.T. 0.01 min
Lab File: BE090882.D
Acq: 18 Sep 2015 12:50

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tgt Ion	Ratio	Lower	Upper
252	100		
253	28.0	17.9	26.9#
125	21.2	10.6	16.0#





Manual Integrations
APPROVED


MMDadoda
9/21/2015 2:24:08 PM



#35
Benzo(k)fluoranthene
Concen: 0.11 ng
RT: 22.71 min Scan# 3002
Delta R.T. 0.01 min
Lab File: BE090882.D
Acq: 18 Sep 2015 12:50

Tgt Ion	Ratio	Lower	Upper
252	100		
253	27.9	17.9	26.9#
125	20.3	10.1	15.1#





#36

Benzo(a)pyrene

Concen: 0.11 ng

RT: 23.25 min Scan# 3121

Delta R.T. 0.00 min

Lab File: BE090882.D

Acq: 18 Sep 2015 12:50

Instrument :

BNA_E





ClientSampleId :

SSTDIC0.1

Tgt Ion:	252	Resp:	4927
Ion Ratio	Lower	Upper	
252	100		
253	30.8	17.2	25.8#
125	25.1	18.0	27.0

Manual Integrations
APPROVED

MMDadoda
9/21/2015 2:24:08 PM



#37

Dibenzo(a,h)anthracene

Concen: 0.10 ng




RT: 25.74 min Scan# 3665




Delta R.T. 0.00 min

Lab File: BE090882.D

Acq: 18 Sep 2015 12:50

Tgt Ion:	278	Resp:	4765
Ion Ratio	Lower	Upper	
278	100		
139	30.6	15.4	23.0#
279	28.2	18.8	28.2





#38

Benzo(g,h,i)perylene

Concen: 0.11 ng

RT: 26.45 min Scan# 3819

Delta R.T. 0.01 min

Lab File: BE090882.D

Acq: 18 Sep 2015 12:50

Instrument :

BNA_E

ClientSampleId :

SSTDIC0.1

Tgt Ion	Ratio	Lower	Upper
276	100		
277	27.3	18.5	27.7
138	35.2	20.1	30.1

Manual Integrations
APPROVED

MMDadoda
9/21/2015 2:24:08 PM

