

Data Path : Z:\HPCHEM1\BNA E\Data\BE091817\
 Data File : BE094022.D
 Acq On : 18 Sep 2017 15:40
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 19 15:30:14 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE090517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 05 13:34:56 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	129	0.00
2	1,4-Dioxane	0.551	0.614	-11.4	146	0.00
3	n-Nitrosodimethylamine	0.646	0.664	-2.8	133	0.00
4 S	2-Fluorophenol	1.294	1.255	3.0	127	-0.01
5 S	Phenol-d6	1.941	1.746	10.0	120	0.00
6	bis(2-Chloroethyl)ether	1.587	1.644	-3.6	135	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	129	0.00
8 S	Nitrobenzene-d5	0.338	0.324	4.1	124	-0.02
9	Naphthalene	4.585	4.625	-0.9	128	0.00
10	Hexachlorobutadiene	0.167	0.159	4.8	122	-0.02
11 SURR	2-Methylnaphthalene-d10	0.640	0.650	-1.6	127	0.00
12	2-Methylnaphthalene	0.762	0.765	-0.4	127	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	128	0.00
14 S	2,4,6-Tribromophenol	0.124	0.098	21.0	98	0.00
15 S	2-Fluorobiphenyl	1.582	1.570	0.8	124	0.00
16	Acenaphthylene	1.997	1.907	4.5	126	0.00
17	Acenaphthene	1.349	1.351	-0.1	126	0.00
18	Fluorene	1.741	1.693	2.8	126	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	117	-0.03
20	4-Bromophenyl-phenylether	0.219	0.225	-2.7	102	0.00
21	Hexachlorobenzene	0.236	0.175	25.8#	81	0.00
22	Pentachlorophenol	0.063	0.048	23.8	89	0.00
23	Phenanthrene	1.293	1.192	7.8	95	0.00
24	Anthracene	1.218	1.228	-0.8	116	0.00
25 SURR	Fluoranthene-d10	5.960	5.558	6.7	111	0.00
26	Fluoranthene	1.783	1.658	7.0	112	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	106	-0.01
28	Pvrene	1.376	1.462	-6.3	113	0.00
29 S	Terphenyl-d14	0.736	0.761	-3.4	109	0.00
30	Benzo(a)anthracene	1.229	1.109	9.8	95	0.00
31	Chrysene	1.300	1.319	-1.5	108	-0.01
32	Bis(2-ethylhexyl)phthalate	2.983	2.430	18.5	90	-0.01
33	Indeno(1,2,3-cd)pyrene	0.867	0.590	31.9#	73	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	88	0.00
35	Benzo(b)fluoranthene	1.274	1.248	2.0	86	0.00
36	Benzo(k)fluoranthene	1.561	1.808	-15.8	103	0.00
37 C	Benzo(a)pyrene	1.297	1.244	4.1	85	0.00
38	Dibenzo(a,h)anthracene	0.860	0.693	19.4	71	0.00
39	Benzo(a,h,i)perylene	0.907	0.794	12.5	78	-0.01

Data Path : Z:\HPCHEM1\BNA E\Data\BE091817\
Data File : BE094022.D
Acq On : 18 Sep 2017 15:40
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 19 15:30:14 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE090517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 05 13:34:56 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			