

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE091824\
 Data File : BE101355.D
 Acq On : 18 Sep 2024 9:19
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 18 10:54:35 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-BE091724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 17 16:03:42 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	108	0.00
2	1,4-Dioxane	40.000	39.978	0.1	110	0.00
3	Pyridine	40.000	42.594	-6.5	110	-0.01
4	n-Nitrosodimethylamine	40.000	41.573	-3.9	112	-0.01
5 S	2-Fluorophenol	80.000	79.528	0.6	107	0.00
6	Aniline	40.000	41.250	-3.1	113	0.00
7 S	Phenol-d6	80.000	80.172	-0.2	106	0.00
8	2-Chlorophenol	40.000	39.929	0.2	108	0.00
9	Benzaldehyde	40.000	36.273	9.3	93	0.00
10 C	Phenol	40.000	40.294	-0.7	105	0.00
11	bis(2-Chloroethyl)ether	40.000	40.435	-1.1	103	0.00
12	1,3-Dichlorobenzene	40.000	39.295	1.8	108	0.00
13 C	1,4-Dichlorobenzene	40.000	39.115	2.2	107	0.00
14	1,2-Dichlorobenzene	40.000	39.267	1.8	107	0.00
15	Benzyl Alcohol	40.000	43.123	-7.8	107	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.403	-1.0	110	-0.01
17	2-Methylphenol	40.000	40.650	-1.6	107	0.00
18	Hexachloroethane	40.000	41.258	-3.1	111	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.865	-7.2	108	0.00
20	3+4-Methylphenols	40.000	40.718	-1.8	106	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	106	0.00
22	Acetophenone	40.000	40.780	-2.0	106	0.00
23 S	Nitrobenzene-d5	80.000	82.762	-3.5	107	0.00
24	Nitrobenzene	40.000	40.727	-1.8	106	0.00
25	Isophorone	40.000	41.626	-4.1	106	0.00
26 C	2-Nitrophenol	40.000	42.714	-6.8	108	0.00
27	2,4-Dimethylphenol	40.000	40.965	-2.4	106	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.864	-4.7	109	0.00
29 C	2,4-Dichlorophenol	40.000	41.600	-4.0	107	0.00
30	1,2,4-Trichlorobenzene	40.000	40.202	-0.5	108	0.00
31	Naphthalene	40.000	40.008	-0.0	107	0.00
32	Benzoic acid	40.000	37.759	5.6	103	0.00
33	4-Chloroaniline	40.000	40.980	-2.4	105	0.00
34 C	Hexachlorobutadiene	40.000	41.329	-3.3	111	0.00
35	Caprolactam	40.000	40.961	-2.4	101	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.508	-1.3	104	0.00
37	2-Methylnaphthalene	40.000	40.456	-1.1	107	0.00
38	1-Methylnaphthalene	40.000	39.995	0.0	106	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.793	-2.0	107	0.00
41 P	Hexachlorocyclopentadiene	40.000	44.618	-11.5	110	0.00
42 S	2,4,6-Tribromophenol	80.000	81.991	-2.5	104	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.729	-4.3	105	0.00
44	2,4,5-Trichlorophenol	40.000	41.583	-4.0	105	0.00
45 S	2-Fluorobiphenyl	80.000	81.729	-2.2	106	0.00
46	1,1'-Biphenyl	40.000	40.414	-1.0	106	0.00
47	2-Chloronaphthalene	40.000	40.509	-1.3	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE091824\
 Data File : BE101355.D
 Acq On : 18 Sep 2024 9:19
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 18 10:54:35 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-BE091724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 17 16:03:42 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.490	-6.2	102	0.00
49	Acenaphthylene	40.000	40.051	-0.1	104	0.00
50	Dimethylphthalate	40.000	39.761	0.6	105	0.00
51	2,6-Dinitrotoluene	40.000	40.974	-2.4	104	0.00
52 C	Acenaphthene	40.000	39.856	0.4	104	0.00
53	3-Nitroaniline	40.000	42.052	-5.1	104	0.00
54 P	2,4-Dinitrophenol	40.000	41.774	-4.4	104	0.00
55	Dibenzofuran	40.000	39.406	1.5	104	0.00
56 P	4-Nitrophenol	40.000	42.870	-7.2	106	0.00
57	2,4-Dinitrotoluene	40.000	41.353	-3.4	103	0.00
58	Fluorene	40.000	39.345	1.6	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.955	0.1	102	0.00
60	Diethylphthalate	40.000	40.146	-0.4	104	0.00
61	4-Chlorophenyl-phenylether	40.000	40.409	-1.0	106	0.00
62	4-Nitroaniline	40.000	42.551	-6.4	102	0.00
63	Azobenzene	40.000	40.729	-1.8	105	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.322	-5.8	105	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.642	-1.6	103	0.00
67	4-Bromophenyl-phenylether	40.000	41.307	-3.3	105	0.00
68	Hexachlorobenzene	40.000	40.372	-0.9	102	0.00
69	Atrazine	40.000	39.775	0.6	102	0.00
70 C	Pentachlorophenol	40.000	43.517	-8.8	105	0.00
71	Phenanthrene	40.000	40.065	-0.2	103	0.00
72	Anthracene	40.000	40.606	-1.5	103	0.00
73	Carbazole	40.000	40.511	-1.3	103	0.00
74	Di-n-butylphthalate	40.000	43.114	-7.8	107	0.00
75 C	Fluoranthene	40.000	40.341	-0.9	103	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	102	0.00
77	Benzydine	40.000	58.893	-47.2#	128	0.00
78	Pyrene	40.000	40.848	-2.1	103	0.00
79 S	Terphenyl-d14	80.000	68.639	14.2	103	0.00
80	Butylbenzylphthalate	40.000	43.279	-8.2	106	0.00
81	Benzo(a)anthracene	40.000	39.936	0.2	101	0.00
82	3,3'-Dichlorobenzidine	40.000	42.841	-7.1	102	0.00
83	Chrysene	40.000	39.427	1.4	102	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.827	-12.1	108	0.00
85 c	Di-n-octyl phthalate	40.000	42.477	-6.2	105	0.00
86 I	Perylene-d12	20.000	20.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.304	-0.8	99	0.00
88	Benzo(b)fluoranthene	40.000	41.271	-3.2	103	0.00
89	Benzo(k)fluoranthene	40.000	39.506	1.2	98	0.00
90 C	Benzo(a)pyrene	40.000	40.750	-1.9	100	0.00
91	Dibenzo(a,h)anthracene	40.000	40.624	-1.6	98	0.00
92	Benzo(g,h,i)perylene	40.000	40.169	-0.4	99	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE091824\
Data File : BE101355.D
Acq On : 18 Sep 2024 9:19
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC040

Quant Time: Sep 18 10:54:35 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-BE091724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 17 16:03:42 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0