

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE091918\
 Data File : BE097568.D
 Acq On : 20 Sep 2018 8:58
 Operator : SJ/JU
 Sample : J4958-22
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 BPS1-TT-MW309D-20180915

Manual Integrations
 APPROVED

Sohil
 9/20/2018 3:35:04 PM

Quant Time: Sep 20 11:01:15 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 20 10:54:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.38	152	623	0.40	ng	0.00
7) Naphthalene-d8	10.15	136	2812	0.40	ng	0.01
13) Acenaphthene-d10	14.01	164	1868	0.40	ng	0.01
19) Phenanthrene-d10	16.77	188	6132	0.40	ng	0.01
27) Chrysene-d12	20.97	240	8056	0.40	ng	0.00
34) Perylene-d12	23.21	264	7896	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.06	112	269	0.14	ng	0.01
5) Phenol-d6	6.69	99	170m	0.07	ng	0.08
8) Nitrobenzene-d5	8.58	82	574	0.26	ng	0.05
11) 2-Methylnaphthalene-d10	11.75	152	1541	0.40	ng	0.04
14) 2,4,6-Tribromophenol	15.54	330	292m	0.41	ng	0.02
15) 2-Fluorobiphenyl	12.64	172	3312	0.50	ng	0.02
25) Fluoranthene-d10	18.80	212	30628	0.39	ng	0.00
29) Terphenyl-d14	19.42	244	7496	0.51	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.08	88	100	0.059	ng	# 47
9) Naphthalene	10.20	128	737	0.029	ng	# 71
32) Bis(2-ethylhexyl)phthalate	20.91	149	3045	0.100	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE091918\
Data File : BE097568.D
Acq On : 20 Sep 2018 8:58
Operator : SJ/JU
Sample : J4958-22
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
BPS1-TT-MW309D-20180915

Manual Integrations
APPROVED

Sohil
9/20/2018 3:35:04 PM

Quant Time: Sep 20 11:01:15 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 20 10:54:43 2018
Response via : Initial Calibration

