

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE091918\
 Data File : BE097542.D
 Acq On : 19 Sep 2018 16:19
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE091918

Quant Time: Sep 19 16:57:55 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 19 16:26:00 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	-0.01
2	1,4-Dioxane	0.707	0.733	-3.7	109	0.00
3	n-Nitrosodimethylamine	0.624	0.607	2.7	127	0.00
4 S	2-Fluorophenol	1.199	1.230	-2.6	119	-0.01
5 S	Phenol-d6	1.532	1.457	4.9	108	-0.01
6	bis(2-Chloroethyl)ether	1.421	1.457	-2.5	119	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
8 S	Nitrobenzene-d5	0.309	0.304	1.6	115	-0.01
9	Naphthalene	3.602	3.763	-4.5	117	0.00
10	Hexachlorobutadiene	0.166	0.171	-3.0	113	0.00
11 SURR	2-Methylnaphthalene-d10	0.551	0.562	-2.0	116	0.00
12	2-Methylnaphthalene	0.562	0.572	-1.8	116	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.166	0.155	6.6	123	0.00
15 S	2-Fluorobiphenyl	1.413	1.520	-7.6	120	0.00
16	Acenaphthylene	1.581	1.578	0.2	116	0.00
17	Acenaphthene	1.060	1.121	-5.8	117	0.00
18	Fluorene	1.382	1.410	-2.0	116	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	111	0.00
20	4-Bromophenyl-phenylether	0.180	0.179	0.6	121	0.00
21	Hexachlorobenzene	0.198	0.202	-2.0	121	0.00
22	Pentachlorophenol	0.069	0.062	10.1	125	0.00
23	Phenanthrene	0.950	0.958	-0.8	123	0.00
24	Anthracene	0.839	0.804	4.2	120	0.00
25 SURR	Fluoranthene-d10	5.121	5.070	1.0	120	0.00
26	Fluoranthene	1.318	1.322	-0.3	122	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	110	0.00
28	Pvrene	0.910	0.956	-5.1	126	0.00
29 S	Terphenyl-d14	0.732	0.752	-2.7	124	0.00
30	Benzo(a)anthracene	1.007	0.970	3.7	119	0.00
31	Chrysene	1.080	1.105	-2.3	118	0.00
32	Bis(2-ethylhexyl)phthalate	2.006	1.444	28.0#	108	0.00
33	Indeno(1,2,3-cd)pyrene	1.301	1.302	-0.1	114	0.00
34 I	Perylene-d12	1.000	1.000	0.0	103	0.00
35	Benzo(b)fluoranthene	1.026	1.000	2.5	113	0.00
36	Benzo(k)fluoranthene	1.171	1.206	-3.0	119	0.00
37 C	Benzo(a)pyrene	1.053	1.015	3.6	115	0.00
38	Dibenzo(a,h)anthracene	1.075	1.115	-3.7	117	0.00
39	Benzo(a,h,i)perylene	1.046	1.079	-3.2	115	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE091918\
Data File : BE097542.D
Acq On : 19 Sep 2018 16:19
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE091918

Quant Time: Sep 19 16:57:55 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 19 16:26:00 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			