

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE091918\
 Data File : BE097561.D
 Acq On : 20 Sep 2018 4:03
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 20 06:50:44 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 19 16:26:00 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.707	0.671	5.1	92	0.01
3	n-Nitrosodimethylamine	0.624	0.556	10.9	108	0.00
4 S	2-Fluorophenol	1.199	1.195	0.3	107	0.00
5 S	Phenol-d6	1.532	1.634	-6.7	112	0.00
6	bis(2-Chloroethyl)ether	1.421	1.458	-2.6	111	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
8 S	Nitrobenzene-d5	0.309	0.298	3.6	117	0.00
9	Naphthalene	3.602	3.574	0.8	115	0.00
10	Hexachlorobutadiene	0.166	0.156	6.0	106	0.00
11 SURR	2-Methylnaphthalene-d10	0.551	0.597	-8.3	127	0.00
12	2-Methylnaphthalene	0.562	0.597	-6.2	125	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	137	0.00
14 S	2,4,6-Tribromophenol	0.166	0.148	10.8	155#	0.01
15 S	2-Fluorobiphenyl	1.413	1.269	10.2	133	0.00
16	Acenaphthylene	1.581	1.612	-2.0	157#	0.00
17	Acenaphthene	1.060	1.068	-0.8	148	0.00
18	Fluorene	1.382	1.461	-5.7	159#	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	150#	0.00
20	4-Bromophenyl-phenylether	0.180	0.176	2.2	160#	0.00
21	Hexachlorobenzene	0.198	0.197	0.5	159#	0.00
22	Pentachlorophenol	0.069	0.033	52.2#	89	0.01
23	Phenanthrene	0.950	0.902	5.1	157#	0.00
24	Anthracene	0.839	0.794	5.4	160#	0.00
25 SURR	Fluoranthene-d10	5.121	4.218	17.6	135	0.00
26	Fluoranthene	1.318	1.091	17.2	135	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
28	Pvrene	0.910	1.034	-13.6	132	0.00
29 S	Terphenyl-d14	0.732	0.762	-4.1	122	0.00
30	Benzo(a)anthracene	1.007	1.006	0.1	119	0.00
31	Chrysene	1.080	1.069	1.0	111	0.00
32	Bis(2-ethylhexyl)phthalate	2.006	1.779	11.3	129	0.00
33	Indeno(1,2,3-cd)pyrene	1.301	1.283	1.4	109	0.00
34 I	Perylene-d12	1.000	1.000	0.0	102	0.00
35	Benzo(b)fluoranthene	1.026	0.981	4.4	110	0.00
36	Benzo(k)fluoranthene	1.171	1.105	5.6	108	0.00
37 C	Benzo(a)pyrene	1.053	0.991	5.9	111	0.00
38	Dibenzo(a,h)anthracene	1.075	1.064	1.0	111	0.00
39	Benzo(a,h,i)perylene	1.046	1.031	1.4	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE091918\
Data File : BE097561.D
Acq On : 20 Sep 2018 4:03
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 20 06:50:44 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 19 16:26:00 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		