

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE091918\
 Data File : BE097563.D
 Acq On : 20 Sep 2018 5:53
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 20 06:55:22 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 19 16:26:00 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.707	0.641	9.3	87	0.00
3	n-Nitrosodimethylamine	0.624	0.543	13.0	104	0.00
4 S	2-Fluorophenol	1.199	1.144	4.6	102	0.00
5 S	Phenol-d6	1.532	1.673	-9.2	115	0.00
6	bis(2-Chloroethyl)ether	1.421	1.394	1.9	105	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
8 S	Nitrobenzene-d5	0.309	0.280	9.4	108	0.00
9	Naphthalene	3.602	3.552	1.4	112	0.00
10	Hexachlorobutadiene	0.166	0.157	5.4	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.551	0.597	-8.3	125	0.00
12	2-Methylnaphthalene	0.562	0.609	-8.4	126	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	132	0.00
14 S	2,4,6-Tribromophenol	0.166	0.181	-9.0	183#	0.00
15 S	2-Fluorobiphenyl	1.413	1.315	6.9	132	0.00
16	Acenaphthylene	1.581	1.644	-4.0	154#	0.00
17	Acenaphthene	1.060	1.075	-1.4	143	0.00
18	Fluorene	1.382	1.472	-6.5	154#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	144	0.00
20	4-Bromophenyl-phenylether	0.180	0.177	1.7	155#	0.00
21	Hexachlorobenzene	0.198	0.202	-2.0	156#	0.00
22	Pentachlorophenol	0.069	0.051	26.1#	132	0.00
23	Phenanthrene	0.950	0.935	1.6	156#	0.00
24	Anthracene	0.839	0.811	3.3	157#	0.00
25 SURR	Fluoranthene-d10	5.121	4.304	16.0	132	0.00
26	Fluoranthene	1.318	1.121	14.9	133	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
28	Pvrene	0.910	1.008	-10.8	128	0.00
29 S	Terphenyl-d14	0.732	0.757	-3.4	121	0.00
30	Benzo(a)anthracene	1.007	0.974	3.3	115	0.00
31	Chrysene	1.080	1.059	1.9	109	0.00
32	Bis(2-ethylhexyl)phthalate	2.006	1.579	21.3	114	0.00
33	Indeno(1,2,3-cd)pyrene	1.301	1.272	2.2	108	0.00
34 I	Perylene-d12	1.000	1.000	0.0	101	0.00
35	Benzo(b)fluoranthene	1.026	0.970	5.5	107	0.00
36	Benzo(k)fluoranthene	1.171	1.137	2.9	110	0.00
37 C	Benzo(a)pyrene	1.053	0.974	7.5	108	0.00
38	Dibenzo(a,h)anthracene	1.075	1.060	1.4	109	0.00
39	Benzo(a,h,i)perylene	1.046	1.022	2.3	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE091918\
Data File : BE097563.D
Acq On : 20 Sep 2018 5:53
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 20 06:55:22 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 19 16:26:00 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		