

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE091924\
 Data File : BE101370.D
 Acq On : 19 Sep 2024 9:14
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 19 11:09:22 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-BE091724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 17 16:03:42 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	92	0.00
2	1,4-Dioxane	40.000	40.074	-0.2	94	0.00
3	Pyridine	40.000	41.279	-3.2	91	-0.01
4	n-Nitrosodimethylamine	40.000	42.460	-6.2	97	0.00
5 S	2-Fluorophenol	80.000	80.055	-0.1	92	0.00
6	Aniline	40.000	44.069	-10.2	103	0.00
7 S	Phenol-d6	80.000	80.761	-1.0	91	0.00
8	2-Chlorophenol	40.000	40.066	-0.2	92	0.00
9	Benzaldehyde	40.000	46.573	-16.4	101	0.00
10 C	Phenol	40.000	41.097	-2.7	92	0.00
11	bis(2-Chloroethyl)ether	40.000	36.992	7.5	80	0.00
12	1,3-Dichlorobenzene	40.000	39.015	2.5	91	0.00
13 C	1,4-Dichlorobenzene	40.000	39.195	2.0	91	0.00
14	1,2-Dichlorobenzene	40.000	38.930	2.7	91	0.00
15	Benzyl Alcohol	40.000	43.607	-9.0	92	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.557	1.1	92	0.00
17	2-Methylphenol	40.000	40.696	-1.7	91	0.00
18	Hexachloroethane	40.000	40.507	-1.3	93	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.541	-6.4	91	0.00
20	3+4-Methylphenols	40.000	41.352	-3.4	91	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	90	0.00
22	Acetophenone	40.000	40.818	-2.0	89	0.00
23 S	Nitrobenzene-d5	80.000	83.998	-5.0	91	0.00
24	Nitrobenzene	40.000	41.310	-3.3	91	0.00
25	Isophorone	40.000	41.613	-4.0	89	0.00
26 C	2-Nitrophenol	40.000	43.293	-8.2	92	0.00
27	2,4-Dimethylphenol	40.000	41.290	-3.2	90	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.538	-1.3	89	0.00
29 C	2,4-Dichlorophenol	40.000	41.468	-3.7	90	0.00
30	1,2,4-Trichlorobenzene	40.000	39.669	0.8	90	0.00
31	Naphthalene	40.000	40.257	-0.6	91	0.00
32	Benzoic acid	40.000	40.629	-1.6	95	0.00
33	4-Chloroaniline	40.000	41.907	-4.8	90	0.00
34 C	Hexachlorobutadiene	40.000	39.526	1.2	89	0.00
35	Caprolactam	40.000	43.633	-9.1	91	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.652	-4.1	90	0.00
37	2-Methylnaphthalene	40.000	40.110	-0.3	89	0.00
38	1-Methylnaphthalene	40.000	40.191	-0.5	90	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	89	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.963	0.1	90	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.279	-3.2	87	0.00
42 S	2,4,6-Tribromophenol	80.000	83.880	-4.8	91	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.895	-4.7	91	0.00
44	2,4,5-Trichlorophenol	40.000	41.966	-4.9	91	0.00
45 S	2-Fluorobiphenyl	80.000	81.226	-1.5	90	0.00
46	1,1'-Biphenyl	40.000	39.819	0.5	89	0.00
47	2-Chloronaphthalene	40.000	40.225	-0.6	90	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE091924\
 Data File : BE101370.D
 Acq On : 19 Sep 2024 9:14
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 19 11:09:22 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-BE091724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 17 16:03:42 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.533	-11.3	92	0.00
49	Acenaphthylene	40.000	40.637	-1.6	90	0.00
50	Dimethylphthalate	40.000	39.904	0.2	90	0.00
51	2,6-Dinitrotoluene	40.000	41.408	-3.5	90	0.00
52 C	Acenaphthene	40.000	40.282	-0.7	90	0.00
53	3-Nitroaniline	40.000	44.028	-10.1	93	0.00
54 P	2,4-Dinitrophenol	40.000	42.643	-6.6	91	0.00
55	Dibenzofuran	40.000	39.888	0.3	90	0.00
56 P	4-Nitrophenol	40.000	44.001	-10.0	93	0.00
57	2,4-Dinitrotoluene	40.000	43.129	-7.8	92	0.00
58	Fluorene	40.000	40.343	-0.9	90	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.497	-3.7	91	0.00
60	Diethylphthalate	40.000	40.298	-0.7	90	0.00
61	4-Chlorophenyl-phenylether	40.000	39.868	0.3	89	0.00
62	4-Nitroaniline	40.000	45.106	-12.8	93	0.00
63	Azobenzene	40.000	41.098	-2.7	91	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	91	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.070	-5.2	92	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.498	-1.2	90	0.00
67	4-Bromophenyl-phenylether	40.000	40.026	-0.1	90	0.00
68	Hexachlorobenzene	40.000	40.581	-1.5	91	0.00
69	Atrazine	40.000	35.323	11.7	80	0.00
70 C	Pentachlorophenol	40.000	43.006	-7.5	91	0.00
71	Phenanthrene	40.000	40.502	-1.3	91	0.00
72	Anthracene	40.000	41.572	-3.9	92	0.00
73	Carbazole	40.000	41.128	-2.8	92	0.00
74	Di-n-butylphthalate	40.000	42.521	-6.3	93	0.00
75 C	Fluoranthene	40.000	41.058	-2.6	92	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	92	0.00
77	Benzidine	40.000	66.089	-65.2#	129	0.00
78	Pyrene	40.000	40.906	-2.3	93	0.00
79 S	Terphenyl-d14	80.000	70.541	11.8	96	0.00
80	Butylbenzylphthalate	40.000	42.465	-6.2	94	0.00
81	Benzo(a)anthracene	40.000	40.743	-1.9	93	0.00
82	3,3'-Dichlorobenzidine	40.000	43.976	-9.9	94	0.00
83	Chrysene	40.000	40.583	-1.5	94	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.428	-8.6	94	0.00
85 c	Di-n-octyl phthalate	40.000	42.679	-6.7	95	0.00
86 I	Perylene-d12	20.000	20.000	0.0	92	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.153	-2.9	93	0.00
88	Benzo(b)fluoranthene	40.000	40.799	-2.0	93	0.00
89	Benzo(k)fluoranthene	40.000	41.037	-2.6	93	0.00
90 C	Benzo(a)pyrene	40.000	41.118	-2.8	93	0.00
91	Dibenzo(a,h)anthracene	40.000	41.359	-3.4	92	0.00
92	Benzo(g,h,i)perylene	40.000	40.579	-1.4	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE091924\
Data File : BE101370.D
Acq On : 19 Sep 2024 9:14
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC040

Quant Time: Sep 19 11:09:22 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-BE091724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 17 16:03:42 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0