

Data Path : Z:\HPCHEM1\BNA E\DATA\BE092115\
 Data File : BE090917.D
 Acq On : 21 Sep 2015 14:21
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001

Quant Time: Sep 21 15:57:29 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE091815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 18 18:17:34 2015
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.55	152	31787	5.00	ng	0.00
7) Naphthalene-d8	10.30	136	141317	5.00	ng	0.00
13) Acenaphthene-d10	14.17	164	72422	5.00	ng	0.00
19) Phenanthrene-d10	16.91	188	174574	5.00	ng	0.00
26) Chrysene-d12	21.07	240	228603	5.00	ng	0.00
33) Perylene-d12	23.35	264	205569	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.19	112	2405	0.29	ng	0.00
5) Phenol-d6	6.77	99	7612	0.77	ng	0.01
8) Nitrobenzene-d5	8.71	82	8334	0.86	ng	0.00
14) 2,4,6-Tribromophenol	15.69	330	1547	0.61	ng	0.00
15) 2-Fluorobiphenyl	12.80	172	18586	0.91	ng	0.00
28) Terphenyl-d14	19.54	244	24221	0.88	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.15	88	4573	0.96	ng	98
3) n-Nitrosodimethylamine	3.45	42	5111	0.81	ng	# 90
6) bis(2-Chloroethyl)ether	6.99	93	8408	0.92	ng	89
9) Nitrobenzene	8.75	77	8720	0.81	ng	99
10) Naphthalene	10.35	128	29101	0.97	ng	98
11) Hexachlorobutadiene	10.64	225	4699	1.05	ng	99
12) 2-Methylnaphthalene	11.99	142	16812	0.88	ng	97
16) Acenaphthylene	13.88	152	110877	0.88	ng	100
17) Acenaphthene	14.23	154	18637	0.96	ng	92
18) Fluorene	15.22	166	22328	0.93	ng	97
20) 4-Bromophenyl-phenylether	16.13	248	5915	0.88	ng	88
21) Hexachlorobenzene	16.23	284	30908	1.03	ng	97
22) Pentachlorophenol	16.61	266	1107	0.51	ng	# 85
23) Phenanthrene	16.94	178	41378	0.97	ng	99
24) Anthracene	17.05	178	34950	0.88	ng	99
25) Fluoranthene	18.96	202	43799	0.82	ng	100
27) Pyrene	19.32	202	46004	0.88	ng	100
29) Benzo(a)anthracene	21.05	228	48243	0.85	ng	100
30) Chrysene	21.10	228	57329	0.97	ng	100
31) Bis(2-ethylhexyl)phthalate	21.00	149	11083	0.58	ng	100
32) Indeno(1,2,3-cd)pyrene	25.72	276	49746	0.75	ng	100
34) Benzo(b)fluoranthene	22.66	252	41623	0.86	ng	100
35) Benzo(k)fluoranthene	22.70	252	50612	0.91	ng	99
36) Benzo(a)pyrene	23.25	252	37928	0.83	ng	# 91
37) Dibenzo(a,h)anthracene	25.75	278	38535	0.81	ng	99
38) Benzo(g,h,i)perylene	26.44	276	42059	0.83	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE092115\
Data File : BE090917.D
Acq On : 21 Sep 2015 14:21
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDCCC001

Quant Time: Sep 21 15:57:29 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE091815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 18 18:17:34 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.55 min Scan# 726

Delta R.T. -0.00 min

Lab File: BE090917.D

Acq: 21 Sep 2015 14:21

Instrument :

BNA_E

ClientSampleId :

SSTDCCC001


Tot Ion:152 Resp: 31787

Ion Ratio Lower Upper

152 100

150 159.3 122.2 183.4




115 68.9 51.0 76.4



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 0.96 ng

RT: 3.15 min Scan# 82

Delta R.T. 0.00 min

Lab File: BE090917.D

Acq: 21 Sep 2015 14:21


Tgt Ion: 88 Resp: 4573

Ion Ratio Lower Upper

88 100

58 88.4 68.4 102.6



43 34.0 26.9 40.3




Abundance Ion 88.00 (87.70 to 88.70): BE0

Ion 58.00 (57.70 to 58.70): BE0

Ion 43.00 (42.70 to 43.70): BE0









#3

n-Nitrosodimethylamine
 Concen: 0.81 ng
 RT: 3.45 min Scan# 126
 Delta R.T. 0.00 min
 Lab File: BE090917.D
 Acq: 21 Sep 2015 14:21

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001




Tot Ion: 42 Resp: 5111
 Ion Ratio Lower Upper
 42 100
 74 114.6 83.7 125.5
 44 16.5 10.0 15.0#






#4

2-Fluorophenol
 Concen: 0.29 ng
 RT: 5.19 min Scan# 382
 Delta R.T. 0.01 min
 Lab File: BE090917.D
 Acq: 21 Sep 2015 14:21

Tgt Ion: 112 Resp: 2405
 Ion Ratio Lower Upper
 112 100
 64 66.7 57.3 85.9
 63 35.2 29.2 43.8





#5

Phenol-d6

Concen: 0.77 ng

RT: 6.77 min Scan# 611

Delta R.T. 0.01 min

Lab File: BE090917.D

Acq: 21 Sep 2015 14:21

Instrument :

BNA_E

ClientSampleId :

SSTDCCC001





Tot Ion: 99 Resp: 7612

Ion Ratio Lower Upper

99 100

42 21.5 16.8 25.2

71 37.2 28.3 42.5



#6

bis(2-Chloroethyl)ether

Concen: 0.92 ng

RT: 6.99 min Scan# 644

Delta R.T. 0.01 min

Lab File: BE090917.D

Acq: 21 Sep 2015 14:21


Tgt Ion: 93 Resp: 8408


Ion Ratio Lower Upper

93 100

63 73.8 68.5 102.7

95 30.1 26.9 40.3





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.30 min Scan# 1211

Delta R.T. 0.00 min

Lab File: BE090917.D

Acq: 21 Sep 2015 14:21

Instrument :

BNA_E

ClientSampleId :

SSTDCCC001

Tot Ion:136 Resp: 141317


Ion Ratio Lower Upper

136 100

137 10.9 8.7 13.1

54 12.0 9.0 13.6

68 8.4 6.3 9.5






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0



#8

Nitrobenzene-d5

Concen: 0.86 ng

RT: 8.71 min Scan# 934

Delta R.T. 0.00 min

Lab File: BE090917.D

Acq: 21 Sep 2015 14:21


Tgt Ion: 82 Resp: 8334

Ion Ratio Lower Upper

82 100

128 30.5 25.0 37.4



54 58.5 46.3 69.5




Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0





#9

Nitrobenzene

Concen: 0.81 ng

RT: 8.75 min Scan# 942

Delta R.T. 0.00 min

Lab File: BE090917.D

Acq: 21 Sep 2015 14:21

Instrument :

BNA_E

ClientSampleId :

SSTDCCC001





Tot Ion: 77 Resp: 8720

Ion Ratio Lower Upper

77 100

123 30.8 25.1 37.7

65 13.1 9.9 14.9



#10

Naphthalene

Concen: 0.97 ng

RT: 10.35 min Scan# 1218

Delta R.T. 0.00 min

Lab File: BE090917.D

Acq: 21 Sep 2015 14:21




Tgt Ion: 128 Resp: 29101


Ion Ratio Lower Upper

128 100

129 11.4 8.4 12.6

127 13.1 9.9 14.9





#11

Hexachlorobutadiene

Concen: 1.05 ng

RT: 10.64 min Scan# 1256

Delta R.T. 0.00 min

Lab File: BE090917.D

Acq: 21 Sep 2015 14:21

Instrument :

BNA_E

ClientSampleId :

SSTDCCC001


Tot Ion:225 Resp: 4699

Ion Ratio Lower Upper

225 100

223 62.0 50.6 76.0




227 62.9 51.0 76.6



Abundance Ion 225.00 (224.70 to 225.70): E

Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E



#12

2-Methylnaphthalene

Concen: 0.88 ng

RT: 11.99 min Scan# 1490

Delta R.T. 0.00 min

Lab File: BE090917.D

Acq: 21 Sep 2015 14:21


Tgt Ion:142 Resp: 16812

Ion Ratio Lower Upper

142 100

141 86.7 71.4 107.2



115 37.5 31.0 46.4




Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E









#13
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.17 min Scan# 1915
Delta R.T. -0.00 min
Lab File: BE090917.D
Acq: 21 Sep 2015 14:21




Instrument :
BNA_E
ClientSampleId :
SSTDCCC001


Tot Ion:164 Resp: 72422
Ion Ratio Lower Upper
164 100
162 103.7 86.8 130.2
160 55.5 53.9 80.9



#14
2,4,6-Tribromophenol
Concen: 0.61 ng
RT: 15.69 min Scan# 2050
Delta R.T. 0.00 min
Lab File: BE090917.D
Acq: 21 Sep 2015 14:21

Tgt Ion:330 Resp: 1547
Ion Ratio Lower Upper
330 100
332 97.7 74.9 112.3
141 178.7 145.2 217.8









#15
2-Fluorobiphenyl
Concen: 0.91 ng
RT: 12.80 min Scan# 1668
Delta R.T. 0.00 min
Lab File: BE090917.D
Acq: 21 Sep 2015 14:21




Instrument :
BNA_E
ClientSampleId :
SSTDCCC001


Tot Ion:172 Resp: 18586
Ion Ratio Lower Upper
172 100
171 35.8 27.8 41.8
170 24.8 19.8 29.6



#16
Acenaphthylene
Concen: 0.88 ng
RT: 13.88 min Scan# 1871
Delta R.T. -0.00 min
Lab File: BE090917.D
Acq: 21 Sep 2015 14:21

Tgt Ion:152 Resp: 110877
Ion Ratio Lower Upper
152 100
151 4.9 3.9 5.9
153 13.1 10.3 15.5









#17
Acenaphthene
Concen: 0.96 ng
RT: 14.23 min Scan# 1924
Delta R.T. -0.00 min
Lab File: BE090917.D
Acq: 21 Sep 2015 14:21




Instrument :
BNA_E
ClientSampleId :
SSTDCCC001


Tot Ion:154 Resp: 18637
Ion Ratio Lower Upper
154 100
153 438.2 354.5 531.7
152 252.5 227.8 341.6



#18
Fluorene
Concen: 0.93 ng
RT: 15.22 min Scan# 2023
Delta R.T. 0.00 min
Lab File: BE090917.D
Acq: 21 Sep 2015 14:21

Tgt Ion:166 Resp: 22328
Ion Ratio Lower Upper
166 100
165 100.0 82.7 124.1
167 13.5 10.7 16.1





#19

Phenanthrene-d10

Concen: 5.00 ng

RT: 16.91 min Scan# 2120

Delta R.T. 0.00 min

Lab File: BE090917.D

Acq: 21 Sep 2015 14:21

Instrument :

BNA_E

ClientSampleId :

SSTDCCC001


Tot Ion:188 Resp: 174574

Ion Ratio Lower Upper

188 100

94 10.8 10.3 15.5




80 11.9 11.0 16.6



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BE0

Ion 80.00 (79.70 to 80.70): BE0



#20

4-Bromophenyl-phenylether

Concen: 0.88 ng

RT: 16.13 min Scan# 2075

Delta R.T. 0.00 min

Lab File: BE090917.D

Acq: 21 Sep 2015 14:21


Tgt Ion:248 Resp: 5915

Ion Ratio Lower Upper

248 100

250 96.6 76.7 115.1



141 333.7 292.6 439.0




Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E









#21
Hexachlorobenzene
Concen: 1.03 ng
RT: 16.23 min Scan# 2081
Delta R.T. 0.00 min
Lab File: BE090917.D
Acq: 21 Sep 2015 14:21




Instrument :
BNA_E
ClientSampleId :
SSTDCCC001


Tot Ion:284 Resp: 30908
Ion Ratio Lower Upper
284 100
142 41.6 35.0 52.4
249 30.9 25.5 38.3



#22
Pentachlorophenol
Concen: 0.51 ng
RT: 16.61 min Scan# 2103
Delta R.T. 0.02 min
Lab File: BE090917.D
Acq: 21 Sep 2015 14:21

Tgt Ion:266 Resp: 1107
Ion Ratio Lower Upper
266 100
268 77.6 49.6 74.4#
264 75.2 53.9 80.9





#23

Phenanthrene

Concen: 0.97 ng

RT: 16.94 min Scan# 2122

Delta R.T. 0.00 min

Lab File: BE090917.D

Acq: 21 Sep 2015 14:21

Instrument :

BNA_E

ClientSampleId :

SSTDCCC001


Tot Ion:178 Resp: 41378

Ion Ratio Lower Upper

178 100

176 19.2 15.6 23.4




179 15.2 12.8 19.2



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#24

Anthracene

Concen: 0.88 ng

RT: 17.05 min Scan# 2128

Delta R.T. 0.00 min

Lab File: BE090917.D

Acq: 21 Sep 2015 14:21


Tgt Ion:178 Resp: 34950

Ion Ratio Lower Upper

178 100

176 18.4 15.0 22.6



179 15.0 12.3 18.5




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E









#25
 Fluoranthene
 Concen: 0.82 ng
 RT: 18.96 min Scan# 2346
 Delta R.T. -0.00 min
 Lab File: BE090917.D
 Acq: 21 Sep 2015 14:21




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001


Tot Ion	Ratio	Lower	Upper
202	100		
101	13.9	11.0	16.6
203	17.1	13.9	20.9



#26
 Chrysene-d12
 Concen: 5.00 ng
 RT: 21.07 min Scan# 2705
 Delta R.T. 0.00 min
 Lab File: BE090917.D
 Acq: 21 Sep 2015 14:21

Tgt Ion	Ratio	Lower	Upper
240	100		
120	14.8	10.1	15.1
236	27.4	21.9	32.9









#27
Pvrene
Concen: 0.88 ng
RT: 19.32 min Scan# 2414
Delta R.T. -0.00 min
Lab File: BE090917.D
Acq: 21 Sep 2015 14:21




Instrument :
BNA_E
ClientSampleId :
SSTDCCC001


Tot Ion:202 Resp: 46004
Ion Ratio Lower Upper
202 100
200 21.1 16.7 25.1
203 17.3 13.8 20.6



#28
Terphenyl-d14
Concen: 0.88 ng
RT: 19.54 min Scan# 2455
Delta R.T. 0.01 min
Lab File: BE090917.D
Acq: 21 Sep 2015 14:21

Tgt Ion:244 Resp: 24221
Ion Ratio Lower Upper
244 100
212 8.4 6.9 10.3
122 15.1 12.9 19.3





#29

Benzo(a)anthracene

Concen: 0.85 ng

RT: 21.05 min Scan# 2703

Delta R.T. 0.00 min

Lab File: BE090917.D

Acq: 21 Sep 2015 14:21

Instrument :

BNA_E

ClientSampleId :

SSTDCCC001





Tot Ion:228 Resp: 48243

Ion Ratio Lower Upper

228 100

226 27.2 21.6 32.4

229 19.6 15.9 23.9



#30

Chrysene

Concen: 0.97 ng

RT: 21.10 min Scan# 2710

Delta R.T. 0.00 min

Lab File: BE090917.D

Acq: 21 Sep 2015 14:21




Tgt Ion:228 Resp: 57329


Ion Ratio Lower Upper

228 100

226 30.2 24.2 36.4

229 19.6 15.3 22.9









#31
 Bis(2-ethylhexyl)phthalate
 Concen: 0.58 ng
 RT: 21.00 min Scan# 2695
 Delta R.T. 0.00 min
 Lab File: BE090917.D
 Acq: 21 Sep 2015 14:21




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001


Tot Ion:149 Resp: 11083
 Ion Ratio Lower Upper
 149 100
 167 25.1 20.1 30.1
 279 3.0 2.3 3.5



#32
 Indeno(1,2,3-cd)pyrene
 Concen: 0.75 ng
 RT: 25.72 min Scan# 3658
 Delta R.T. 0.00 min
 Lab File: BE090917.D
 Acq: 21 Sep 2015 14:21

Tgt Ion:276 Resp: 49746
 Ion Ratio Lower Upper
 276 100
 138 29.5 23.3 34.9
 277 24.6 19.8 29.6









#33
 Pervlene-d12
 Concen: 5.00 ng
 RT: 23.35 min Scan# 3142
 Delta R.T. 0.00 min
 Lab File: BE090917.D
 Acq: 21 Sep 2015 14:21




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001


Tot Ion:264 Resp: 205569
 Ion Ratio Lower Upper
 264 100
 260 24.8 19.7 29.5
 265 21.7 17.5 26.3



#34
 Benzo(b)fluoranthene
 Concen: 0.86 ng
 RT: 22.66 min Scan# 2990
 Delta R.T. 0.00 min
 Lab File: BE090917.D
 Acq: 21 Sep 2015 14:21

Tgt Ion:252 Resp: 41623
 Ion Ratio Lower Upper
 252 100
 253 22.4 17.9 26.9
 125 13.2 10.6 16.0









#35
Benzo(k)fluoranthene
Concen: 0.91 ng
RT: 22.70 min Scan# 3000
Delta R.T. 0.00 min
Lab File: BE090917.D
Acq: 21 Sep 2015 14:21




Instrument :
BNA_E
ClientSampleId :
SSTDCCC001


Tot Ion:252 Resp: 50612
Ion Ratio Lower Upper
252 100
253 22.4 17.9 26.9
125 13.3 10.1 15.1



#36
Benzo(a)pyrene
Concen: 0.83 ng
RT: 23.25 min Scan# 3120
Delta R.T. 0.00 min
Lab File: BE090917.D
Acq: 21 Sep 2015 14:21

Tgt Ion:252 Resp: 37928
Ion Ratio Lower Upper
252 100
253 22.6 17.2 25.8
125 15.0 18.0 27.0#









#37
Dibenzo(a,h)anthracene
Concen: 0.81 ng
RT: 25.75 min Scan# 3664
Delta R.T. 0.00 min
Lab File: BE090917.D
Acq: 21 Sep 2015 14:21

Instrument :
BNA_E
ClientSampleId :
SSTDCCC001

Tot Ion:278 Resp: 38535

Ion	Ratio	Lower	Upper
278	100		
139	19.5	15.4	23.0
279	24.1	18.8	28.2



#38
Benzo(g,h,i)perylene
Concen: 0.83 ng
RT: 26.44 min Scan# 3817
Delta R.T. 0.00 min
Lab File: BE090917.D
Acq: 21 Sep 2015 14:21

Tgt Ion:276 Resp: 42059

Ion	Ratio	Lower	Upper
276	100		
277	24.4	18.5	27.7
138	25.1	20.1	30.1

