

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE092115\
 Data File : BE090910.D
 Acq On : 21 Sep 2015 5:09
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 21 07:28:11 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 18 18:17:34 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	122	0.00
2	1,4-Dioxane	0.795	0.736	7.4	111	0.00
3	n-Nitrosodimethylamine	1.067	0.819	23.2	97	0.00
4 S	2-Fluorophenol	1.238	0.930	24.9	93	0.00
5 S	Phenol-d6	1.559	1.247	20.0	101	0.01
6	bis(2-Chloroethyl)ether	1.546	1.459	5.6	126	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	125	0.00
8 S	Nitrobenzene-d5	0.345	0.300	13.0	110	0.00
9	Nitrobenzene	0.383	0.318	17.0	108	0.00
10	Naphthalene	1.126	1.028	8.7	116	0.00
11	Hexachlorobutadiene	0.166	0.163	1.8	121	0.00
12	2-Methylnaphthalene	0.672	0.616	8.3	117	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	134	0.00
14 S	2,4,6-Tribromophenol	0.196	0.136	30.6#	106	0.00
15 S	2-Fluorobiphenyl	1.418	1.288	9.2	121	0.00
16	Acenaphthylene	8.705	7.810	10.3	121	0.00
17	Acenaphthene	1.341	1.285	4.2	129	0.00
18	Fluorene	1.705	1.562	8.4	126	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	138	0.00
20	4-Bromophenyl-phenylether	0.193	0.171	11.4	124	0.00
21	Hexachlorobenzene	0.906	0.845	6.7	127	0.00
22	Pentachlorophenol	0.087	0.037	57.5#	65	0.02
23	Phenanthrene	1.292	1.175	9.1	128	0.00
24	Anthracene	1.144	1.028	10.1	128	0.00
25	Fluoranthene	1.525	1.314	13.8	120	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	123	0.00
27	Pyrene	1.141	1.045	8.4	121	0.00
28 S	Terphenyl-d14	0.638	0.558	12.5	118	0.00
29	Benzo(a)anthracene	1.235	1.100	10.9	115	0.00
30	Chrysene	1.352	1.255	7.2	115	0.00
31	Bis(2-ethylhexyl)phthalate	0.515	0.321	37.7#	88	0.00
32	Indeno(1,2,3-cd)pyrene	1.455	1.176	19.2	99	0.00
33 I	Perylene-d12	1.000	1.000	0.0	115	0.00
34	Benzo(b)fluoranthene	1.170	1.052	10.1	107	0.00
35	Benzo(k)fluoranthene	1.347	1.236	8.2	106	0.00
36 C	Benzo(a)pyrene	1.113	0.975	12.4	105	0.00
37	Dibenzo(a,h)anthracene	1.158	0.982	15.2	99	0.00
38	Benzo(g,h,i)perylene	1.240	1.065	14.1	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0