

Data Path : Z:\HPCHEM1\BNA_E\Data\BE092115\
 Data File : BE090917.D
 Acq On : 21 Sep 2015 14:21
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 22 02:16:22 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 18 18:17:34 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	128	0.00
2	1,4-Dioxane	1.000	0.961	3.9	113	0.00
3	n-Nitrosodimethylamine	1.000	0.811	18.9	99	0.00
4 S	2-Fluorophenol	1.000	0.738	26.2#	91	0.00
5 S	Phenol-d6	1.000	0.768	23.2	101	0.01
6	bis(2-Chloroethyl)ether	1.000	1.024	-2.4	133	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	130	0.00
8 S	Nitrobenzene-d5	1.000	0.856	14.4	112	0.00
9	Nitrobenzene	1.000	0.806	19.4	109	0.00
10	Naphthalene	1.000	0.966	3.4	121	0.00
11	Hexachlorobutadiene	1.000	1.046	-4.6	129	0.00
12	2-Methylnaphthalene	1.000	0.885	11.5	118	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	132	0.00
14 S	2,4,6-Tribromophenol	1.000	0.607	39.3#	83	0.00
15 S	2-Fluorobiphenyl	1.000	0.905	9.5	119	0.00
16	Acenaphthylene	1.000	0.879	12.1	117	0.00
17	Acenaphthene	1.000	0.959	4.1	128	0.00
18	Fluorene	1.000	0.935	6.5	123	0.00
19 I	Phenanthrene-d10	5.000	5.000	0.0	134	0.00
20	4-Bromophenyl-phenylether	1.000	0.876	12.4	119	0.00
21	Hexachlorobenzene	1.000	1.028	-2.8	130	0.00
22	Pentachlorophenol	1.000	0.509	49.1#	54	0.02
23	Phenanthrene	1.000	0.972	2.8	126	0.00
24	Anthracene	1.000	0.875	12.5	121	0.00
25	Fluoranthene	1.000	0.823	17.7	111	0.00
26 I	Chrysene-d12	5.000	5.000	0.0	117	0.00
27	Pyrene	1.000	0.882	11.8	111	0.00
28 S	Terphenyl-d14	1.000	0.880	12.0	107	0.00
29	Benzo(a)anthracene	1.000	0.854	14.6	105	0.00
30	Chrysene	1.000	0.973	2.7	109	0.00
31	Bis(2-ethylhexyl)phthalate	1.000	0.581	41.9#	63	0.00
32	Indeno(1,2,3-cd)pyrene	1.000	0.748	25.2#	87	0.00
33 I	Perylene-d12	5.000	5.000	0.0	104	0.00
34	Benzo(b)fluoranthene	1.000	0.865	13.5	93	0.00
35	Benzo(k)fluoranthene	1.000	0.914	8.6	96	0.00
36 C	Benzo(a)pyrene	1.000	0.829	17.1	90	0.00
37	Dibenzo(a,h)anthracene	1.000	0.810	19.0	85	0.00
38	Benzo(g,h,i)perylene	1.000	0.825	17.5	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0