

Data Path : Z:\HPCHEM1\BNA_E\Data\BE092115\
 Data File : BE090917.D
 Acq On : 21 Sep 2015 14:21
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 22 02:16:22 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 18 18:17:34 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	128	0.00
2	1,4-Dioxane	0.795	0.719	9.6	113	0.00
3	n-Nitrosodimethylamine	1.067	0.804	24.6	99	0.00
4 S	2-Fluorophenol	1.238	0.873	29.5#	91	0.00
5 S	Phenol-d6	1.559	1.197	23.2	101	0.01
6	bis(2-Chloroethyl)ether	1.546	1.471	4.9	133	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	130	0.00
8 S	Nitrobenzene-d5	0.345	0.295	14.5	112	0.00
9	Nitrobenzene	0.383	0.309	19.3	109	0.00
10	Naphthalene	1.126	1.030	8.5	121	0.00
11	Hexachlorobutadiene	0.166	0.166	0.0	129	0.00
12	2-Methylnaphthalene	0.672	0.595	11.5	118	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	132	0.00
14 S	2,4,6-Tribromophenol	0.196	0.107	45.4#	83	0.00
15 S	2-Fluorobiphenyl	1.418	1.283	9.5	119	0.00
16	Acenaphthylene	8.705	7.655	12.1	117	0.00
17	Acenaphthene	1.341	1.287	4.0	128	0.00
18	Fluorene	1.705	1.542	9.6	123	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	134	0.00
20	4-Bromophenyl-phenylether	0.193	0.169	12.4	119	0.00
21	Hexachlorobenzene	0.906	0.885	2.3	130	0.00
22	Pentachlorophenol	0.087	0.032	63.2#	54	0.02
23	Phenanthrene	1.292	1.185	8.3	126	0.00
24	Anthracene	1.144	1.001	12.5	121	0.00
25	Fluoranthene	1.525	1.254	17.8	111	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	117	0.00
27	Pyrene	1.141	1.006	11.8	111	0.00
28 S	Terphenyl-d14	0.638	0.530	16.9	107	0.00
29	Benzo(a)anthracene	1.235	1.055	14.6	105	0.00
30	Chrysene	1.352	1.254	7.2	109	0.00
31	Bis(2-ethylhexyl)phthalate	0.515	0.242	53.0#	63	0.00
32	Indeno(1,2,3-cd)pyrene	1.455	1.088	25.2#	87	0.00
33 I	Perylene-d12	1.000	1.000	0.0	104	0.00
34	Benzo(b)fluoranthene	1.170	1.012	13.5	93	0.00
35	Benzo(k)fluoranthene	1.347	1.231	8.6	96	0.00
36 C	Benzo(a)pyrene	1.113	0.923	17.1	90	0.00
37	Dibenzo(a,h)anthracene	1.158	0.937	19.1	85	0.00
38	Benzo(g,h,i)perylene	1.240	1.023	17.5	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0