

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE092115\
 Data File : BE090910.D
 Acq On : 21 Sep 2015 5:09
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001

Manual Integrations
 APPROVED

MMDadoda
 9/21/2015 2:24:59 PM

Quant Time: Sep 21 07:28:11 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 18 18:17:34 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.55	152	30348	5.00	ng	0.00
7) Naphthalene-d8	10.30	136	135714	5.00	ng	0.00
13) Acenaphthene-d10	14.17	164	73249	5.00	ng	0.00
19) Phenanthrene-d10	16.91	188	179721	5.00	ng	0.00
26) Chrysene-d12	21.07	240	241119	5.00	ng	0.00
33) Perylene-d12	23.35	264	227486	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.19	112	5644m	0.79	ng	0.00
5) Phenol-d6	6.77	99	7568	0.80	ng	0.01
8) Nitrobenzene-d5	8.71	82	8151	0.87	ng	0.00
14) 2,4,6-Tribromophenol	15.69	330	1992	0.77	ng	0.00
15) 2-Fluorobiphenyl	12.80	172	18873	0.91	ng	0.00
28) Terphenyl-d14	19.54	244	26892	0.93	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.15	88	4470	0.99	ng	99
3) n-Nitrosodimethylamine	3.45	42	4970	0.83	ng	83
6) bis(2-Chloroethyl)ether	6.99	93	8856m	1.02	ng	
9) Nitrobenzene	8.75	77	8625	0.83	ng	97
10) Naphthalene	10.35	128	27896	0.96	ng	98
11) Hexachlorobutadiene	10.64	225	4412	1.02	ng	99
12) 2-Methylnaphthalene	11.99	142	16729	0.92	ng	99
16) Acenaphthylene	13.88	152	114411	0.90	ng	100
17) Acenaphthene	14.22	154	18819	0.96	ng	89
18) Fluorene	15.22	166	22877	0.95	ng	97
20) 4-Bromophenyl-phenylether	16.13	248	6158	0.89	ng	88
21) Hexachlorobenzene	16.23	284	30358	0.98	ng	99
22) Pentachlorophenol	16.61	266	1345	0.57	ng	# 92
23) Phenanthrene	16.94	178	42252	0.96	ng	99
24) Anthracene	17.05	178	36941	0.90	ng	100
25) Fluoranthene	18.96	202	47227	0.86	ng	99
27) Pyrene	19.32	202	50411	0.92	ng	99
29) Benzo(a)anthracene	21.05	228	53070	0.89	ng	99
30) Chrysene	21.10	228	60516	0.97	ng	98
31) Bis(2-ethylhexyl)phthalate	21.00	149	15456	0.71	ng	99
32) Indeno(1,2,3-cd)pyrene	25.72	276	56690	0.81	ng	99
34) Benzo(b)fluoranthene	22.66	252	47872	0.90	ng	100
35) Benzo(k)fluoranthene	22.71	252	56240	0.92	ng	100
36) Benzo(a)pyrene	23.25	252	44371	0.88	ng	# 91
37) Dibenzo(a,h)anthracene	25.74	278	44676	0.85	ng	99
38) Benzo(g,h,i)perylene	26.44	276	48453	0.86	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_E\DATA\BE092115\
Data File : BE090910.D
Acq On : 21 Sep 2015 5:09
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDCCC001

Manual Integrations
APPROVED

MMDadoda
9/21/2015 2:24:59 PM

Quant Time: Sep 21 07:28:11 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 18 18:17:34 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.55 min Scan# 726

Delta R.T. -0.00 min

Lab File: BE090910.D


Acq: 21 Sep 2015 5:09

Instrument :

BNA_E

ClientSampleId :

SSTDCCC001



Tgt Ion: 152 Resp: 30348

Ion Ratio Lower Upper

152 100




150 155.9 122.2 183.4

115 67.6 51.0 76.4

Manual Integrations
APPROVED

MMDadoda

9/21/2015 2:24:59 PM



#2

1,4-Dioxane

Concen: 0.99 ng

RT: 3.15 min Scan# 82

Delta R.T. -0.00 min

Lab File: BE090910.D

Acq: 21 Sep 2015 5:09




Tgt Ion: 88 Resp: 4470


Ion Ratio Lower Upper

88 100

58 85.6 68.4 102.6

43 31.6 26.9 40.3





#3

n-Nitrosodimethylamine

Concen: 0.83 ng

RT: 3.45 min Scan# 126

Delta R.T. 0.00 min

Lab File: BE090910.D

Acq: 21 Sep 2015 5:09

Instrument :

BNA_E

ClientSampleId :





SSTDCCC001

Manual Integrations
APPROVED

MMDadoda

9/21/2015 2:24:59 PM

Tgt Ion:	42	Resp:	4970
Ion	Ratio	Lower	Upper
42	100		
74	123.9	83.7	125.5
44	12.1	10.0	15.0



#4

2-Fluorophenol

Concen: 0.79 ng m




RT: 5.19 min Scan# 382




Delta R.T. 0.01 min

Lab File: BE090910.D

Acq: 21 Sep 2015 5:09

Tgt Ion:	112	Resp:	5644
Ion	Ratio	Lower	Upper
112	100		
64	32.8	57.3	85.9#
63	17.1	29.2	43.8#





#5

Phenol-d6

Concen: 0.80 ng

RT: 6.77 min Scan# 611

Delta R.T. 0.01 min

Lab File: BE090910.D

Acq: 21 Sep 2015 5:09

Tgt Ion:	99	Resp:	7568
Ion	Ratio	Lower	Upper
99	100		
42	22.1	16.8	25.2
71	36.0	28.3	42.5

Instrument :





BNA_E

ClientSampleId :

SSTDCCC001

Manual Integrations
APPROVED

MMDadoda
9/21/2015 2:24:59 PM



#6

bis(2-Chloroethyl)ether

Concen: 1.02 ng m


RT: 6.99 min Scan# 644


Delta R.T. 0.01 min

Lab File: BE090910.D

Acq: 21 Sep 2015 5:09

Tgt Ion:	93	Resp:	8856
Ion	Ratio	Lower	Upper
93	100		
63	69.4	68.5	102.7
95	27.7	26.9	40.3





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.30 min Scan# 1211

Delta R.T. 0.00 min

Lab File: BE090910.D


Acq: 21 Sep 2015 5:09

Instrument :

BNA_E

ClientSampleId :




SSTDCCC001



Tgt Ion	Ratio	Resp	Lower	Upper
136	100	135714		
137	11.0	8.7	13.1	
54	11.8	9.0	13.6	
68	8.5	6.3	9.5	

Manual Integrations
APPROVED

MMDadoda
9/21/2015 2:24:59 PM



#8

Nitrobenzene-d5


Concen: 0.87 ng

RT: 8.71 min Scan# 934



Delta R.T. 0.01 min


Lab File: BE090910.D

Acq: 21 Sep 2015 5:09



Tgt Ion	Ratio	Resp	Lower	Upper
82	100	8151		
128	31.3	25.0	37.4	
54	57.5	46.3	69.5	





#9

Nitrobenzene

Concen: 0.83 ng

RT: 8.75 min Scan# 942

Delta R.T. 0.00 min

Lab File: BE090910.D

Acq: 21 Sep 2015 5:09

Instrument :

BNA_E





ClientSampleId :

SSTDCCC001

Manual Integrations
APPROVED

MMDadoda
9/21/2015 2:24:59 PM

Tgt Ion	Ratio	Lower	Upper
77	100		
123	33.3	25.1	37.7
65	11.6	9.9	14.9



#10

Naphthalene

Concen: 0.96 ng




RT: 10.35 min Scan# 1217


Delta R.T. -0.01 min

Lab File: BE090910.D

Acq: 21 Sep 2015 5:09

Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.3	8.4	12.6
127	13.4	9.9	14.9





#11

Hexachlorobutadiene

Concen: 1.02 ng

RT: 10.64 min Scan# 1256

Delta R.T. 0.00 min

Lab File: BE090910.D


Acq: 21 Sep 2015 5:09

Instrument :

BNA_E

ClientSampleId :




SSTDCCC001



Tgt Ion	Ratio	Lower	Upper
225	100		
223	62.7	50.6	76.0
227	64.5	51.0	76.6

Manual Integrations
APPROVED

MMDadoda
9/21/2015 2:24:59 PM



#12

2-Methylnaphthalene

Concen: 0.92 ng




RT: 11.99 min Scan# 1490


Delta R.T. 0.00 min

Lab File: BE090910.D

Acq: 21 Sep 2015 5:09

Tgt Ion	Ratio	Lower	Upper
142	100		
141	88.6	71.4	107.2
115	38.8	31.0	46.4









#13
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.17 min Scan# 1915
Delta R.T. -0.00 min
Lab File: BE090910.D
Acq: 21 Sep 2015 5:09

Instrument :
BNA_E
ClientSampleId :
SSTDCCC001

Tgt Ion	Ratio	Lower	Upper
164	100		
162	102.1	86.8	130.2
160	51.6	53.9	80.9#




Manual Integrations
APPROVED


MMDadoda
9/21/2015 2:24:59 PM



#14
2,4,6-Tribromophenol
Concen: 0.77 ng
RT: 15.69 min Scan# 2051
Delta R.T. 0.00 min
Lab File: BE090910.D
Acq: 21 Sep 2015 5:09

Tgt Ion	Ratio	Lower	Upper
330	100		
332	95.1	74.9	112.3
141	162.8	145.2	217.8









#15
2-Fluorobiphenyl
Concen: 0.91 ng
RT: 12.80 min Scan# 1669
Delta R.T. 0.00 min
Lab File: BE090910.D
Acq: 21 Sep 2015 5:09

Instrument :
BNA_E
ClientSampleId :
SSTDCCC001

Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.4	27.8	41.8
170	23.8	19.8	29.6




Manual Integrations
APPROVED


MMDadoda
9/21/2015 2:24:59 PM



#16
Acenaphthylene
Concen: 0.90 ng
RT: 13.88 min Scan# 1871
Delta R.T. -0.00 min
Lab File: BE090910.D
Acq: 21 Sep 2015 5:09

Tgt Ion	Ratio	Lower	Upper
152	100		
151	4.9	3.9	5.9
153	13.0	10.3	15.5









#17
Acenaphthene
Concen: 0.96 ng
RT: 14.22 min Scan# 1924
Delta R.T. -0.00 min
Lab File: BE090910.D
Acq: 21 Sep 2015 5:09

Instrument :
BNA_E
ClientSampleId :
SSTDCCC001

Tgt Ion	Ratio	Lower	Upper
154	100		
153	435.6	354.5	531.7
152	238.6	227.8	341.6




Manual Integrations
APPROVED


MMDadoda
9/21/2015 2:24:59 PM



#18
Fluorene
Concen: 0.95 ng
RT: 15.22 min Scan# 2024
Delta R.T. 0.00 min
Lab File: BE090910.D
Acq: 21 Sep 2015 5:09

Tgt Ion	Ratio	Lower	Upper
166	100		
165	100.3	82.7	124.1
167	13.4	10.7	16.1









#19
Phenanthrene-d10
Concen: 5.00 ng
RT: 16.91 min Scan# 2121
Delta R.T. 0.00 min
Lab File: BE090910.D
Acq: 21 Sep 2015 5:09

Instrument :
BNA_E
ClientSampleId :
SSTDCCC001

Tgt Ion	Ratio	Resp Lower	Resp Upper
188	100		
94	9.8	10.3	15.5#
80	10.5	11.0	16.6#




Manual Integrations
APPROVED


MMDadoda
9/21/2015 2:24:59 PM



#20
4-Bromophenyl-phenylether
Concen: 0.89 ng
RT: 16.13 min Scan# 2076
Delta R.T. 0.00 min
Lab File: BE090910.D
Acq: 21 Sep 2015 5:09

Tgt Ion	Ratio	Resp Lower	Resp Upper
248	100		
250	98.2	76.7	115.1
141	332.8	292.6	439.0









#21
Hexachlorobenzene
Concen: 0.98 ng
RT: 16.23 min Scan# 2082
Delta R.T. 0.00 min
Lab File: BE090910.D
Acq: 21 Sep 2015 5:09

Instrument :
BNA_E
ClientSampleId :
SSTDCCC001

Tgt Ion	Ratio	Lower	Upper
284	100		
142	42.8	35.0	52.4
249	31.1	25.5	38.3




Manual Integrations
APPROVED


MMDadoda
9/21/2015 2:24:59 PM



#22
Pentachlorophenol
Concen: 0.57 ng
RT: 16.61 min Scan# 2104
Delta R.T. 0.02 min
Lab File: BE090910.D
Acq: 21 Sep 2015 5:09

Tgt Ion	Ratio	Lower	Upper
266	100		
268	74.7	49.6	74.4#
264	67.9	53.9	80.9









#23
Phenanthrene
Concen: 0.96 ng
RT: 16.94 min Scan# 2123
Delta R.T. 0.00 min
Lab File: BE090910.D
Acq: 21 Sep 2015 5:09

Instrument :
BNA_E
ClientSampleId :
SSTDCCC001

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.6	23.4
179	15.6	12.8	19.2




Manual Integrations
APPROVED


MMDadoda
9/21/2015 2:24:59 PM



#24
Anthracene
Concen: 0.90 ng
RT: 17.05 min Scan# 2129
Delta R.T. 0.00 min
Lab File: BE090910.D
Acq: 21 Sep 2015 5:09

Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.0	22.6
179	15.2	12.3	18.5





#25

Fluoranthene

Concen: 0.86 ng

RT: 18.96 min Scan# 2347

Delta R.T. 0.00 min

Lab File: BE090910.D

Acq: 21 Sep 2015 5:09

Instrument :

BNA_E





ClientSampleId :

SSTDCCC001

Tgt	Ion	202	Resp:	47227
Ion	Ratio	Lower	Upper	
202	100			
101	14.4	11.0	16.6	
203	17.1	13.9	20.9	

Manual Integrations
APPROVED

MMDadoda
9/21/2015 2:24:59 PM



#26

Chrysene-d12

Concen: 5.00 ng




RT: 21.07 min Scan# 2705


Delta R.T. -0.00 min

Lab File: BE090910.D

Acq: 21 Sep 2015 5:09

Tgt	Ion	240	Resp:	241119
Ion	Ratio	Lower	Upper	
240	100			
120	15.3	10.1	15.1#	
236	27.9	21.9	32.9	









#27
 Pyrene
 Concen: 0.92 ng
 RT: 19.32 min Scan# 2415
 Delta R.T. 0.00 min
 Lab File: BE090910.D
 Acq: 21 Sep 2015 5:09

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001

Tgt Ion	Ratio	Resp	Lower	Upper
202	100	50411		
200	21.2	16.7	25.1	
203	17.3	13.8	20.6	




Manual Integrations
 APPROVED


MMDadoda
 9/21/2015 2:24:59 PM



#28
 Terphenyl-d14
 Concen: 0.93 ng
 RT: 19.54 min Scan# 2456
 Delta R.T. 0.01 min
 Lab File: BE090910.D
 Acq: 21 Sep 2015 5:09

Tgt Ion	Ratio	Resp	Lower	Upper
244	100	26892		
212	8.6	6.9	10.3	
122	14.4	12.9	19.3	









#29
Benzo(a)anthracene
Concen: 0.89 ng
RT: 21.05 min Scan# 2703
Delta R.T. -0.00 min
Lab File: BE090910.D
Acq: 21 Sep 2015 5:09

Instrument :
BNA_E
ClientSampleId :
SSTDCCC001

Tgt Ion	Ratio	Lower	Upper
228	100		
226	27.2	21.6	32.4
229	19.5	15.9	23.9




Manual Integrations
APPROVED


MMDadoda
9/21/2015 2:24:59 PM



#30
Chrysene
Concen: 0.97 ng
RT: 21.10 min Scan# 2710
Delta R.T. 0.00 min
Lab File: BE090910.D
Acq: 21 Sep 2015 5:09


Tgt Ion	Ratio	Lower	Upper
228	100		
226	31.5	24.2	36.4
229	19.6	15.3	22.9





#31
 Bis(2-ethylhexyl)phthalate
 Concen: 0.71 ng
 RT: 21.00 min Scan# 2695
 Delta R.T. -0.00 min
 Lab File: BE090910.D
 Acq: 21 Sep 2015 5:09




Instrument :
 BNA_E
 ClientSampleId :
 SSTCCCC001




Tgt Ion	Ratio	Resp Lower	Resp Upper
149	100		
167	25.5	20.1	30.1
279	3.1	2.3	3.5

Manual Integrations
 APPROVED



MMDadoda
 9/21/2015 2:24:59 PM




#32
 Indeno(1,2,3-cd)pyrene
 Concen: 0.81 ng
 RT: 25.72 min Scan# 3659
 Delta R.T. 0.00 min
 Lab File: BE090910.D
 Acq: 21 Sep 2015 5:09



Tgt Ion	Ratio	Resp Lower	Resp Upper
276	100		
138	29.9	23.3	34.9
277	25.0	19.8	29.6









#33
Perylene-d12
 Concen: 5.00 ng
 RT: 23.35 min Scan# 3143
 Delta R.T. 0.00 min
 Lab File: BE090910.D
 Acq: 21 Sep 2015 5:09

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001

Tgt Ion:	264	Resp:	227486
Ion Ratio	Lower	Upper	
264	100		
260	24.9	19.7	29.5
265	21.7	17.5	26.3




Manual Integrations
 APPROVED


MMDadoda
 9/21/2015 2:24:59 PM



#34
Benzo(b)fluoranthene
 Concen: 0.90 ng
 RT: 22.66 min Scan# 2991
 Delta R.T. 0.00 min
 Lab File: BE090910.D
 Acq: 21 Sep 2015 5:09

Tgt Ion:	252	Resp:	47872
Ion Ratio	Lower	Upper	
252	100		
253	22.2	17.9	26.9
125	13.1	10.6	16.0









#35
Benzo(k)fluoranthene
Concen: 0.92 ng
RT: 22.71 min Scan# 3001
Delta R.T. 0.00 min
Lab File: BE090910.D
Acq: 21 Sep 2015 5:09

Instrument :
BNA_E
ClientSampleId :
SSTDCCC001

Tgt Ion	Ratio	Resp	Lower	Upper
252	100	56240		
253	22.4	17.9	26.9	
125	13.0	10.1	15.1	




Manual Integrations
APPROVED


MMDadoda
9/21/2015 2:24:59 PM



#36
Benzo(a)pyrene
Concen: 0.88 ng
RT: 23.25 min Scan# 3121
Delta R.T. 0.00 min
Lab File: BE090910.D
Acq: 21 Sep 2015 5:09

Tgt Ion	Ratio	Resp	Lower	Upper
252	100	44371		
253	22.3	17.2	25.8	
125	15.2	18.0	27.0	









#37
Dibenzo(a,h)anthracene
Concen: 0.85 ng
RT: 25.74 min Scan# 3664
Delta R.T. 0.00 min
Lab File: BE090910.D
Acq: 21 Sep 2015 5:09

Instrument :
BNA_E
ClientSampleId :
SSTDCCC001

Tgt Ion	Ratio	Lower	Upper
278	100		
139	19.8	15.4	23.0
279	23.6	18.8	28.2

Manual Integrations
APPROVED

MMDadoda
9/21/2015 2:24:59 PM



#38
Benzo(g,h,i)perylene
Concen: 0.86 ng
RT: 26.44 min Scan# 3818
Delta R.T. 0.00 min
Lab File: BE090910.D
Acq: 21 Sep 2015 5:09

Tgt Ion	Ratio	Lower	Upper
276	100		
277	24.4	18.5	27.7
138	25.7	20.1	30.1

