

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE092118\  
 Data File : BE097609.D  
 Acq On : 21 Sep 2018 13:34  
 Operator : SJ/JU  
 Sample : SSTDCCC0.4  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 LabSampleId :  
 SSTDCCC0.4

Quant Time: Sep 22 04:19:01 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Sep 20 10:54:43 2018  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	90	0.00
2	1,4-Dioxane	0.400	0.376	6.0	80	0.00
3	n-Nitrosodimethylamine	0.400	0.365	8.8	93	0.00
4 S	2-Fluorophenol	0.400	0.361	9.8	88	0.00
5 S	Phenol-d6	0.400	0.413	-3.2	99	0.00
6	bis(2-Chloroethyl)ether	0.400	0.358	10.5	88	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	95	0.00
8 S	Nitrobenzene-d5	0.400	0.340	15.0	91	0.00
9	Naphthalene	0.400	0.400	0.0	102	0.00
10	Hexachlorobutadiene	0.400	0.391	2.3	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.431	-7.7	112	0.00
12	2-Methylnaphthalene	0.400	0.420	-5.0	110	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	121	0.00
14 S	2,4,6-Tribromophenol	0.400	0.448	-12.0	160	0.00
15 S	2-Fluorobiphenyl	0.400	0.360	10.0	117	0.00
16	Acenaphthylene	0.400	0.401	-0.3	136	0.00
17	Acenaphthene	0.400	0.393	1.8	127	0.00
18	Fluorene	0.400	0.418	-4.5	138	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	137	0.00
20	4-Bromophenyl-phenylether	0.400	0.381	4.8	142	0.00
21	Hexachlorobenzene	0.400	0.404	-1.0	147	0.00
22	Pentachlorophenol	0.400	0.348	13.0	120	0.00
23	Phenanthrene	0.400	0.393	1.8	147	0.00
24	Anthracene	0.400	0.364	9.0	140	0.00
25 SURR	Fluoranthene-d10	0.400	0.324	19.0	121	0.00
26	Fluoranthene	0.400	0.329	17.8	122	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	98	0.00
28	Pyrene	0.400	0.441	-10.2	118	0.00
29 S	Terphenyl-d14	0.400	0.407	-1.7	110	0.00
30	Benzo(a)anthracene	0.400	0.363	9.3	100	0.00
31	Chrysene	0.400	0.397	0.8	102	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.321	19.8	88	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.376	6.0	96	0.00
34 I	Perylene-d12	0.400	0.400	0.0	90	0.00
35	Benzo(b)fluoranthene	0.400	0.371	7.3	93	0.00
36	Benzo(k)fluoranthene	0.400	0.397	0.8	99	0.00
37 C	Benzo(a)pyrene	0.400	0.386	3.5	94	0.00
38	Dibenzo(a,h)anthracene	0.400	0.392	2.0	96	0.00
39	Benzo(a,h,i)perylene	0.400	0.395	1.3	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE092118\  
Data File : BE097609.D  
Acq On : 21 Sep 2018 13:34  
Operator : SJ/JU  
Sample : SSTDCCC0.4  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_E  
LabSampleId :  
SSTDCCC0.4

Quant Time: Sep 22 04:19:01 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Sep 20 10:54:43 2018  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
-----					
(#) = Out of Range	SPCC's out = 0 CCC's out = 0				