

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE092118\
 Data File : BE097609.D
 Acq On : 21 Sep 2018 13:34
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 22 04:19:01 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 20 10:54:43 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
2	1,4-Dioxane	0.707	0.645	8.8	80	0.00
3	n-Nitrosodimethylamine	0.624	0.529	15.2	93	0.00
4 S	2-Fluorophenol	1.199	1.081	9.8	88	0.00
5 S	Phenol-d6	1.532	1.583	-3.3	99	0.00
6	bis(2-Chloroethyl)ether	1.421	1.272	10.5	88	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
8 S	Nitrobenzene-d5	0.309	0.263	14.9	91	0.00
9	Naphthalene	3.602	3.598	0.1	102	0.00
10	Hexachlorobutadiene	0.166	0.162	2.4	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.551	0.594	-7.8	112	0.00
12	2-Methylnaphthalene	0.562	0.591	-5.2	110	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	121	0.00
14 S	2,4,6-Tribromophenol	0.166	0.173	-4.2	160#	0.00
15 S	2-Fluorobiphenyl	1.413	1.273	9.9	117	0.00
16	Acenaphthylene	1.581	1.586	-0.3	136	0.00
17	Acenaphthene	1.060	1.043	1.6	127	0.00
18	Fluorene	1.382	1.445	-4.6	138	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	137	0.00
20	4-Bromophenyl-phenylether	0.180	0.172	4.4	142	0.00
21	Hexachlorobenzene	0.198	0.200	-1.0	147	0.00
22	Pentachlorophenol	0.069	0.049	29.0#	120	0.00
23	Phenanthrene	0.950	0.932	1.9	147	0.00
24	Anthracene	0.839	0.764	8.9	140	0.00
25 SURR	Fluoranthene-d10	5.121	4.152	18.9	121	0.00
26	Fluoranthene	1.318	1.084	17.8	122	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
28	Pvrene	0.910	1.003	-10.2	118	0.00
29 S	Terphenyl-d14	0.732	0.744	-1.6	110	0.00
30	Benzo(a)anthracene	1.007	0.914	9.2	100	0.00
31	Chrysene	1.080	1.073	0.6	102	0.00
32	Bis(2-ethylhexyl)phthalate	2.006	1.315	34.4#	88	0.00
33	Indeno(1,2,3-cd)pyrene	1.301	1.223	6.0	96	0.00
34 I	Perylene-d12	1.000	1.000	0.0	90	0.00
35	Benzo(b)fluoranthene	1.026	0.952	7.2	93	0.00
36	Benzo(k)fluoranthene	1.171	1.161	0.9	99	0.00
37 C	Benzo(a)pyrene	1.053	0.958	9.0	94	0.00
38	Dibenzo(a,h)anthracene	1.075	1.053	2.0	96	0.00
39	Benzo(a,h,i)perylene	1.046	1.032	1.3	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE092118\
Data File : BE097609.D
Acq On : 21 Sep 2018 13:34
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 22 04:19:01 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 20 10:54:43 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			