

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE092118\
 Data File : BE097628.D
 Acq On : 22 Sep 2018 1:56
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 22 04:54:25 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 20 10:54:43 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
2	1,4-Dioxane	0.707	0.693	2.0	87	0.00
3	n-Nitrosodimethylamine	0.624	0.527	15.5	93	0.01
4 S	2-Fluorophenol	1.199	1.069	10.8	87	0.00
5 S	Phenol-d6	1.532	1.590	-3.8	100	0.01
6	bis(2-Chloroethyl)ether	1.421	1.285	9.6	89	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
8 S	Nitrobenzene-d5	0.309	0.249	19.4	88	0.01
9	Naphthalene	3.602	3.520	2.3	101	0.01
10	Hexachlorobutadiene	0.166	0.164	1.2	101	0.00
11 SURR	2-Methylnaphthalene-d10	0.551	0.561	-1.8	107	0.00
12	2-Methylnaphthalene	0.562	0.579	-3.0	109	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	120	0.00
14 S	2,4,6-Tribromophenol	0.166	0.159	4.2	147	0.01
15 S	2-Fluorobiphenyl	1.413	1.292	8.6	119	0.01
16	Acenaphthylene	1.581	1.556	1.6	133	0.00
17	Acenaphthene	1.060	1.062	-0.2	129	0.00
18	Fluorene	1.382	1.456	-5.4	139	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	139	0.00
20	4-Bromophenyl-phenylether	0.180	0.175	2.8	147	0.00
21	Hexachlorobenzene	0.198	0.201	-1.5	150	0.00
22	Pentachlorophenol	0.069	0.043	37.7#	107	0.01
23	Phenanthrene	0.950	0.915	3.7	147	0.00
24	Anthracene	0.839	0.755	10.0	140	0.00
25 SURR	Fluoranthene-d10	5.121	4.144	19.1	123	0.00
26	Fluoranthene	1.318	1.090	17.3	125	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
28	Pvrene	0.910	1.018	-11.9	118	0.00
29 S	Terphenyl-d14	0.732	0.755	-3.1	110	0.00
30	Benzo(a)anthracene	1.007	0.864	14.2	93	0.00
31	Chrysene	1.080	1.116	-3.3	105	0.00
32	Bis(2-ethylhexyl)phthalate	2.006	1.170	41.7#	77	-0.01
33	Indeno(1,2,3-cd)pyrene	1.301	1.269	2.5	98	0.00
34 I	Perylene-d12	1.000	1.000	0.0	91	0.00
35	Benzo(b)fluoranthene	1.026	0.926	9.7	92	0.00
36	Benzo(k)fluoranthene	1.171	1.206	-3.0	105	0.00
37 C	Benzo(a)pyrene	1.053	0.971	7.8	97	0.00
38	Dibenzo(a,h)anthracene	1.075	1.046	2.7	96	0.00
39	Benzo(a,h,i)perylene	1.046	1.020	2.5	96	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE092118\
Data File : BE097628.D
Acq On : 22 Sep 2018 1:56
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 22 04:54:25 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 20 10:54:43 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		