

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE092118\
 Data File : BE097642.D
 Acq On : 22 Sep 2018 10:35
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Sep 24 02:59:51 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 20 10:54:43 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	-0.01
2	1,4-Dioxane	0.707	0.667	5.7	107	0.00
3	n-Nitrosodimethylamine	0.624	0.503	19.4	114	0.01
4 S	2-Fluorophenol	1.199	1.039	13.3	108	0.00
5 S	Phenol-d6	1.532	1.501	2.0	121	0.00
6	bis(2-Chloroethyl)ether	1.421	1.269	10.7	112	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	120	0.00
8 S	Nitrobenzene-d5	0.309	0.259	16.2	114	0.00
9	Naphthalene	3.602	3.615	-0.4	130	0.00
10	Hexachlorobutadiene	0.166	0.161	3.0	123	0.00
11 SURR	2-Methylnaphthalene-d10	0.551	0.585	-6.2	139	0.00
12	2-Methylnaphthalene	0.562	0.587	-4.4	138	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	152#	0.00
14 S	2,4,6-Tribromophenol	0.166	0.167	-0.6	194#	0.00
15 S	2-Fluorobiphenyl	1.413	1.275	9.8	148	0.00
16	Acenaphthylene	1.581	1.584	-0.2	170#	0.00
17	Acenaphthene	1.060	1.059	0.1	162#	0.00
18	Fluorene	1.382	1.449	-4.8	175#	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	172#	0.00
20	4-Bromophenyl-phenylether	0.180	0.177	1.7	185#	0.00
21	Hexachlorobenzene	0.198	0.195	1.5	180#	0.00
22	Pentachlorophenol	0.069	0.047	31.9#	146	0.00
23	Phenanthrene	0.950	0.925	2.6	184#	0.00
24	Anthracene	0.839	0.764	8.9	176#	0.00
25 SURR	Fluoranthene-d10	5.121	4.082	20.3	150	0.00
26	Fluoranthene	1.318	1.081	18.0	154#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	120	0.00
28	Pvrene	0.910	1.022	-12.3	146	0.00
29 S	Terphenyl-d14	0.732	0.755	-3.1	135	0.00
30	Benzo(a)anthracene	1.007	0.900	10.6	119	0.00
31	Chrysene	1.080	1.089	-0.8	126	0.00
32	Bis(2-ethylhexyl)phthalate	2.006	1.161	42.1#	94	0.00
33	Indeno(1,2,3-cd)pyrene	1.301	1.256	3.5	120	0.00
34 I	Perylene-d12	1.000	1.000	0.0	111	0.00
35	Benzo(b)fluoranthene	1.026	0.945	7.9	114	0.00
36	Benzo(k)fluoranthene	1.171	1.176	-0.4	124	0.00
37 C	Benzo(a)pyrene	1.053	0.949	9.9	115	0.00
38	Dibenzo(a,h)anthracene	1.075	1.056	1.8	118	0.00
39	Benzo(a,h,i)perylene	1.046	1.024	2.1	117	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE092118\
Data File : BE097642.D
Acq On : 22 Sep 2018 10:35
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 24 02:59:51 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 20 10:54:43 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		