

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE092118\
 Data File : BE097642.D
 Acq On : 22 Sep 2018 10:35
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 24 02:59:51 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 20 10:54:43 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	115	-0.01
2	1,4-Dioxane	0.400	0.389	2.8	107	0.00
3	n-Nitrosodimethylamine	0.400	0.349	12.8	114	0.01
4 S	2-Fluorophenol	0.400	0.347	13.3	108	0.00
5 S	Phenol-d6	0.400	0.392	2.0	121	0.00
6	bis(2-Chloroethyl)ether	0.400	0.357	10.8	112	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	120	0.00
8 S	Nitrobenzene-d5	0.400	0.335	16.3	114	0.00
9	Naphthalene	0.400	0.401	-0.3	130	0.00
10	Hexachlorobutadiene	0.400	0.387	3.3	123	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.424	-6.0	139	0.00
12	2-Methylnaphthalene	0.400	0.418	-4.5	138	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	152	0.00
14 S	2,4,6-Tribromophenol	0.400	0.435	-8.7	194	0.00
15 S	2-Fluorobiphenyl	0.400	0.361	9.8	148	0.00
16	Acenaphthylene	0.400	0.401	-0.3	170	0.00
17	Acenaphthene	0.400	0.400	0.0	162	0.00
18	Fluorene	0.400	0.419	-4.7	175	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	172	0.00
20	4-Bromophenyl-phenylether	0.400	0.393	1.8	185	0.00
21	Hexachlorobenzene	0.400	0.393	1.8	180	0.00
22	Pentachlorophenol	0.400	0.341	14.8	146	0.00
23	Phenanthrene	0.400	0.390	2.5	184	0.00
24	Anthracene	0.400	0.364	9.0	176	0.00
25 SURR	Fluoranthene-d10	0.400	0.319	20.3	150	0.00
26	Fluoranthene	0.400	0.328	18.0	154	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	120	0.00
28	Pvrene	0.400	0.449	-12.2	146	0.00
29 S	Terphenyl-d14	0.400	0.413	-3.2	135	0.00
30	Benzo(a)anthracene	0.400	0.357	10.8	119	0.00
31	Chrysene	0.400	0.403	-0.8	126	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.285	28.8#	94	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.386	3.5	120	0.00
34 I	Perylene-d12	0.400	0.400	0.0	111	0.00
35	Benzo(b)fluoranthene	0.400	0.369	7.8	114	0.00
36	Benzo(k)fluoranthene	0.400	0.402	-0.5	124	0.00
37 C	Benzo(a)pyrene	0.400	0.383	4.3	115	0.00
38	Dibenzo(a,h)anthracene	0.400	0.393	1.8	118	0.00
39	Benzo(a,h,i)perylene	0.400	0.392	2.0	117	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE092118\
Data File : BE097642.D
Acq On : 22 Sep 2018 10:35
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 24 02:59:51 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE091918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 20 10:54:43 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				