

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE092216\
 Data File : BE092406.D
 Acq On : 22 Sep 2016 21:33
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 23 03:42:56 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 22 18:25:56 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	-0.02
2	1,4-Dioxane	0.635	0.666	-4.9	94	-0.01
3	n-Nitrosodimethylamine	0.588	0.514	12.6	88	0.00
4 S	2-Fluorophenol	0.999	0.898	10.1	88	0.00
5 S	Phenol-d6	1.388	1.065	23.3	80	0.00
6	bis(2-Chloroethyl)ether	1.222	0.845	30.9#	81	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	87	0.00
8 S	Nitrobenzene-d5	0.299	0.246	17.7	84	0.00
9	Naphthalene	1.082	1.078	0.4	87	0.00
10	Hexachlorobutadiene	0.170	0.171	-0.6	88	0.00
11	2-Methylnaphthalene	0.716	0.703	1.8	86	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	88	0.00
13 S	2,4,6-Tribromophenol	0.144	0.121	16.0	86	0.01
14 S	2-Fluorobiphenyl	1.184	1.160	2.0	88	0.01
15	Acenaphthylene	7.240	6.957	3.9	87	0.00
16	Acenaphthene	1.134	1.113	1.9	84	0.00
17	Fluorene	1.438	1.413	1.7	88	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.02
19	Hexachlorobenzene	0.214	0.200	6.5	89	0.00
20	Pentachlorophenol	0.050	0.037	26.0#	78	0.00
21	Phenanthrene	1.149	1.131	1.6	92	0.02
22	Anthracene	1.105	1.020	7.7	90	0.02
23	Fluoranthene	5.460	5.079	7.0	89	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
25	Pyrene	4.333	4.118	5.0	90	0.02
26 S	Terphenyl-d14	0.620	0.591	4.7	88	0.02
27	Benzo(a)anthracene	4.873	4.603	5.5	88	0.00
28	Chrysene	4.462	4.421	0.9	84	0.00
29	Bis(2-ethylhexyl)phthalate	0.420	0.359	14.5	77	0.00
30	Indeno(1,2,3-cd)pyrene	4.734	4.378	7.5	85	0.05
31 I	Perylene-d12	1.000	1.000	0.0	90	0.04
32	Benzo(b)fluoranthene	1.211	1.201	0.8	90	0.04
33	Benzo(k)fluoranthene	1.267	1.182	6.7	90	0.04
34 C	Benzo(a)pyrene	1.187	1.129	4.9	90	0.02
35	Dibenzo(a,h)anthracene	1.088	1.012	7.0	87	0.07
36	Benzo(g,h,i)perylene	4.673	4.524	3.2	90	0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0