

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE092216\
 Data File : BE092394.D
 Acq On : 22 Sep 2016 14:09
 Operator : UM/SJ
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2

Manual Integrations
 APPROVED

sohil
 9/23/2016 6:47:57 PM

Quant Time: Sep 22 17:03:45 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 22 16:57:02 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.16	152	9432	5.00	ng	0.00
7) Naphthalene-d8	10.95	136	48446	5.00	ng	0.00
12) Acenaphthene-d10	14.82	164	35759	5.00	ng	0.00
18) Phenanthrene-d10	17.58	188	71056	5.00	ng	0.00
24) Chrysene-d12	21.75	240	97124	5.00	ng	0.00
31) Perylene-d12	24.12	264	82229	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.67	112	6620	2.95	ng	-0.02
5) Phenol-d6	7.34	99	9967	3.33	ng	-0.02
8) Nitrobenzene-d5	9.34	82	10182	2.93	ng	-0.02
13) 2,4,6-Tribromophenol	16.31	330	3640	3.29	ng	-0.02
14) 2-Fluorobiphenyl	13.43	172	27465	2.50	ng	-0.01
26) Terphenyl-d14	20.17	244	38996	2.71	ng	-0.02
Target Compounds						
2) 1,4-Dioxane	3.41	88	3207	2.40	ng	Qvalue 94
3) n-Nitrosodimethylamine	3.76	42	4326	3.47	ng	98
6) bis(2-Chloroethyl)ether	7.59	93	7647	3.15	ng	# 93
9) Naphthalene	10.99	128	30506	2.81	ng	# 95
10) Hexachlorobutadiene	11.28	225	4724	2.79	ng	99
11) 2-Methylnaphthalene	12.62	142	21496	3.03	ng	93
15) Acenaphthylene	14.53	152	161274	2.62	ng	100
16) Acenaphthene	14.88	154	24403	2.45	ng	97
17) Fluorene	15.87	166	32205	2.58	ng	100
19) Hexachlorobenzene	16.89	284	9716	2.86	ng	# 66
20) Pentachlorophenol	17.23	266	3545	4.69	ng	# 86
21) Phenanthrene	17.60	178	49580	2.48	ng	# 95
22) Anthracene	17.69	178	50556	2.85	ng	99
23) Fluoranthene	19.61	202	253834	2.91	ng	100
25) Pyrene	19.97	202	261965	2.77	ng	100
27) Benzo(a)anthracene	21.72	228	286622m	2.73	ng	
28) Chrysene	21.77	228	250636m	2.32	ng	
29) Bis(2-ethylhexyl)phthalate	21.63	149	25584	1.76	ng	# 56
30) Indeno(1,2,3-cd)pyrene	26.61	276	280902	2.54	ng	99
32) Benzo(b)fluoranthene	23.38	252	63452	3.01	ng	# 96
33) Benzo(k)fluoranthene	23.43	252	64072	2.87	ng	94
34) Benzo(a)pyrene	24.01	252	61597	2.96	ng	# 94
35) Dibenzo(a,h)anthracene	26.63	278	57625	2.99	ng	95
36) Benzo(g,h,i)perylene	27.40	276	237352	2.93	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_E\DATA\BE092216\
Data File : BE092394.D
Acq On : 22 Sep 2016 14:09
Operator : UM/SJ
Sample : SSTDIC3.2
Misc :
ALS Vial : 4 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDIC3.2

Manual Integrations
APPROVED

sohil
9/23/2016 6:47:57 PM

Quant Time: Sep 22 17:03:45 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 22 16:57:02 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 8.16 min Scan# 576

Delta R.T. 0.00 min

Lab File: BE092394.D

Acq: 22 Sep 2016 14:09

Instrument :

BNA_E





ClientSampled :

SSTDICC3.2

Tgt Ion	Ratio	Lower	Upper
152	100		
150	169.2	106.0	159.0#
115	72.1	31.2	46.8#

Manual Integrations
APPROVED

sohil
9/23/2016 6:47:57 PM



#2

1,4-Dioxane

Concen: 2.40 ng




RT: 3.41 min Scan# 71


Delta R.T. 0.00 min

Lab File: BE092394.D

Acq: 22 Sep 2016 14:09

Tgt Ion	Ratio	Lower	Upper
88	100		
58	87.2	63.6	95.4
43	35.7	28.0	42.0





#3

n-Nitrosodimethylamine

Concen: 3.47 ng

RT: 3.76 min Scan# 101

Delta R.T. -0.03 min

Lab File: BE092394.D

Acq: 22 Sep 2016 14:09

Instrument :

BNA_E





ClientSampled :

SSTDICC3.2

Tgt Ion:	42	Resp:	4326
Ion	Ratio	Lower	Upper
42	100		
74	110.0	86.0	129.0
44	6.2	5.1	7.7

Manual Integrations
APPROVED

sohil
9/23/2016 6:47:57 PM



#4

2-Fluorophenol

Concen: 2.95 ng




RT: 5.67 min Scan# 363


Delta R.T. -0.02 min

Lab File: BE092394.D

Acq: 22 Sep 2016 14:09

Tgt Ion:	112	Resp:	6620
Ion	Ratio	Lower	Upper
112	100		
64	67.7	53.5	80.3
63	35.7	27.9	41.9





#5

Phenol-d6

Concen: 3.33 ng

RT: 7.34 min Scan# 540

Delta R.T. -0.02 min

Lab File: BE092394.D

Acq: 22 Sep 2016 14:09

Instrument :

BNA_E





ClientSampleId :

SSTDICC3.2

Tgt Ion	Ratio	Resp	Lower	Upper
99	100	9967		
42	0.0	16.0	24.0#	
71	33.5	30.6	45.8	

Manual Integrations
APPROVED

sohil
9/23/2016 6:47:57 PM



#6

bis(2-Chloroethyl)ether

Concen: 3.15 ng




RT: 7.59 min Scan# 551


Delta R.T. -0.02 min

Lab File: BE092394.D

Acq: 22 Sep 2016 14:09

Tgt Ion	Ratio	Resp	Lower	Upper
93	100	7647		
63	84.7	71.6	107.4	
95	36.0	24.0	36.0#	





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.95 min Scan# 708

Delta R.T. -0.00 min

Lab File: BE092394.D

Acq: 22 Sep 2016 14:09

Instrument :

BNA_E





ClientSampleId :

SSTDICC3.2

Tgt Ion	Ratio	Resp	Lower	Upper
136	100	48446		
137	11.0	8.2	12.4	
54	6.4	9.6	14.4	
68	5.0	6.7	10.1	

Manual Integrations
APPROVED

sohil
9/23/2016 6:47:57 PM



#8

Nitrobenzene-d5

Concen: 2.93 ng




RT: 9.34 min Scan# 627


Delta R.T. -0.02 min

Lab File: BE092394.D

Acq: 22 Sep 2016 14:09

Tgt Ion	Ratio	Resp	Lower	Upper
82	100	10182		
128	41.4	39.0	58.4	
54	41.9	44.8	67.2	





#9

Naphthalene

Concen: 2.81 ng

RT: 10.99 min Scan# 710

Delta R.T. -0.02 min

Lab File: BE092394.D

Acq: 22 Sep 2016 14:09

Instrument :

BNA_E





ClientSampleId :

SSTDICC3.2

Tgt Ion:	128	Resp:	30506
Ion Ratio	Lower	Upper	
128	100		
129	10.4	11.2	16.8#
127	14.3	11.6	17.4

Manual Integrations
APPROVED

sohil
9/23/2016 6:47:57 PM



#10

Hexachlorobutadiene

Concen: 2.79 ng




RT: 11.28 min Scan# 725


Delta R.T. -0.02 min

Lab File: BE092394.D

Acq: 22 Sep 2016 14:09

Tgt Ion:	225	Resp:	4724
Ion Ratio	Lower	Upper	
225	100		
223	63.4	50.6	76.0
227	63.3	51.4	77.0









#11
2-Methylnaphthalene
Concen: 3.03 ng
RT: 12.62 min Scan# 814
Delta R.T. -0.02 min
Lab File: BE092394.D
Acq: 22 Sep 2016 14:09

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Tgt Ion:142 Resp: 21496
Ion Ratio Lower Upper
142 100
141 87.5 73.7 110.5
115 34.9 34.0 51.0




Manual Integrations
APPROVED


sohil
9/23/2016 6:47:57 PM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.82 min Scan# 980
Delta R.T. 0.00 min
Lab File: BE092394.D
Acq: 22 Sep 2016 14:09

Tgt Ion:164 Resp: 35759
Ion Ratio Lower Upper
164 100
162 108.2 71.4 107.0#
160 49.6 27.4 41.2#









#13
 2,4,6-Tribromophenol
 Concen: 3.29 ng
 RT: 16.31 min Scan# 1088
 Delta R.T. -0.02 min
 Lab File: BE092394.D
 Acq: 22 Sep 2016 14:09

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:330 Resp: 3640
 Ion Ratio Lower Upper
 330 100
 332 97.1 75.4 113.0
 141 41.4 31.0 46.6




Manual Integrations
 APPROVED


sohil
 9/23/2016 6:47:57 PM



#14
 2-Fluorobiphenyl
 Concen: 2.50 ng
 RT: 13.43 min Scan# 884
 Delta R.T. -0.01 min
 Lab File: BE092394.D
 Acq: 22 Sep 2016 14:09

Tgt Ion:172 Resp: 27465
 Ion Ratio Lower Upper
 172 100
 171 36.3 30.1 45.1
 170 25.2 21.8 32.6









#15
Acenaphthylene
Concen: 2.62 ng
RT: 14.53 min Scan# 963
Delta R.T. -0.02 min
Lab File: BE092394.D
Acq: 22 Sep 2016 14:09

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Tgt Ion	Ratio	Lower	Upper
152	100		
151	4.9	3.9	5.9
153	13.1	10.4	15.6




Manual Integrations
APPROVED


sohil
9/23/2016 6:47:57 PM



#16
Acenaphthene
Concen: 2.45 ng
RT: 14.88 min Scan# 984
Delta R.T. 0.00 min
Lab File: BE092394.D
Acq: 22 Sep 2016 14:09

Tgt Ion	Ratio	Lower	Upper
154	100		
153	445.2	350.8	526.2
152	219.4	171.1	256.7









#17
Fluorene
 Concen: 2.58 ng
 RT: 15.87 min Scan# 1050
 Delta R.T. -0.00 min
 Lab File: BE092394.D
 Acq: 22 Sep 2016 14:09

Instrument :
 BNA_E
 ClientSampled :
 SSTDICC3.2

Tgt Ion	Ratio	Resp	Lower	Upper
166	100	32205		
165	98.1		78.2	117.4
167	13.3		11.0	16.4




Manual Integrations
 APPROVED


sohil
 9/23/2016 6:47:57 PM



#18
Phenanthrene-d10
 Concen: 5.00 ng
 RT: 17.58 min Scan# 1154
 Delta R.T. -0.00 min
 Lab File: BE092394.D
 Acq: 22 Sep 2016 14:09

Tgt Ion	Ratio	Resp	Lower	Upper
188	100	71056		
94	3.4		3.2	4.8
80	2.8		2.6	3.8









#19
Hexachlorobenzene
Concen: 2.86 ng
RT: 16.89 min Scan# 1124
Delta R.T. -0.00 min
Lab File: BE092394.D
Acq: 22 Sep 2016 14:09

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Tgt Ion	Ratio	Lower	Upper
284	100		
142	0.0	29.1	43.7#
249	31.8	27.9	41.9




Manual Integrations
APPROVED


sohil
9/23/2016 6:47:57 PM



#20
Pentachlorophenol
Concen: 4.69 ng
RT: 17.23 min Scan# 1139
Delta R.T. -0.02 min
Lab File: BE092394.D
Acq: 22 Sep 2016 14:09

Tgt Ion	Ratio	Lower	Upper
266	100		
268	54.4	62.5	93.7#
264	71.2	57.2	85.8









#21
Phenanthrene
Concen: 2.48 ng
RT: 17.60 min Scan# 1155
Delta R.T. -0.00 min
Lab File: BE092394.D
Acq: 22 Sep 2016 14:09

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Tgt Ion	Ratio	Lower	Upper
178	100		
176	20.1	15.8	23.8
179	15.2	16.0	24.0




Manual Integrations
APPROVED


sohil
9/23/2016 6:47:57 PM



#22
Anthracene
Concen: 2.85 ng
RT: 17.69 min Scan# 1159
Delta R.T. -0.02 min
Lab File: BE092394.D
Acq: 22 Sep 2016 14:09

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.0	22.4
179	15.1	12.2	18.2









#23
 Fluoranthene
 Concen: 2.91 ng
 RT: 19.61 min Scan# 1266
 Delta R.T. 0.00 min
 Lab File: BE092394.D
 Acq: 22 Sep 2016 14:09

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2

Tgt Ion	Ratio	Resp	Lower	Upper
202	100	253834		
101	2.7		2.2	3.4
203	17.3		13.7	20.5




Manual Integrations
 APPROVED


sohil
 9/23/2016 6:47:57 PM



#24
 Chrysene-d12
 Concen: 5.00 ng
 RT: 21.75 min Scan# 1375
 Delta R.T. -0.00 min
 Lab File: BE092394.D
 Acq: 22 Sep 2016 14:09

Tgt Ion	Ratio	Resp	Lower	Upper
240	100	97124		
120	2.8		2.6	4.0
236	29.1		24.6	37.0









#25
 Pyrene
 Concen: 2.77 ng
 RT: 19.97 min Scan# 1287
 Delta R.T. -0.02 min
 Lab File: BE092394.D
 Acq: 22 Sep 2016 14:09

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2

Tgt Ion	Ratio	Resp Lower	Resp Upper
202	100		
200	5.3	4.2	6.4
203	17.4	13.9	20.9




Manual Integrations
 APPROVED


sohil
 9/23/2016 6:47:57 PM



#26
 Terphenyl-d14
 Concen: 2.71 ng
 RT: 20.17 min Scan# 1299
 Delta R.T. -0.02 min
 Lab File: BE092394.D
 Acq: 22 Sep 2016 14:09

Tgt Ion	Ratio	Resp Lower	Resp Upper
244	100		
212	9.9	6.7	10.1
122	16.0	3.6	5.4









#27
Benzo(a)anthracene
Concen: 2.73 ng m
RT: 21.72 min Scan# 1374
Delta R.T. -0.00 min
Lab File: BE092394.D
Acq: 22 Sep 2016 14:09

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.6	23.6	35.4
229	15.8	13.3	19.9




Manual Integrations
APPROVED


sohil
9/23/2016 6:47:57 PM



#28
Chrysene
Concen: 2.32 ng m
RT: 21.77 min Scan# 1376
Delta R.T. -0.00 min
Lab File: BE092394.D
Acq: 22 Sep 2016 14:09

Tgt Ion	Ratio	Lower	Upper
228	100		
226	42.9	36.3	54.5
229	11.7	10.6	16.0









#29
 Bis(2-ethylhexyl)phthalate
 Concen: 1.76 ng
 RT: 21.63 min Scan# 1370
 Delta R.T. -0.00 min
 Lab File: BE092394.D
 Acq: 22 Sep 2016 14:09

Instrument :
 BNA_E
 ClientSampled :
 SSTDICC3.2

Tgt Ion	Ratio	Lower	Upper
149	100		
167	0.0	16.0	24.0#
279	0.0	16.0	24.0#




Manual Integrations
 APPROVED


sohil
 9/23/2016 6:47:57 PM



#30
 Indeno(1,2,3-cd)pyrene
 Concen: 2.54 ng
 RT: 26.61 min Scan# 1717
 Delta R.T. -0.02 min
 Lab File: BE092394.D
 Acq: 22 Sep 2016 14:09

Tgt Ion	Ratio	Lower	Upper
276	100		
138	26.2	20.3	30.5
277	24.8	19.7	29.5









#31
 Perylene-d12
 Concen: 5.00 ng
 RT: 24.12 min Scan# 1547
 Delta R.T. 0.00 min
 Lab File: BE092394.D
 Acq: 22 Sep 2016 14:09

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC3.2

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.8	20.1	30.1
265	22.6	17.9	26.9




Manual Integrations
 APPROVED


sohil
 9/23/2016 6:47:57 PM



#32
 Benzo(b)fluoranthene
 Concen: 3.01 ng
 RT: 23.38 min Scan# 1483
 Delta R.T. -0.01 min
 Lab File: BE092394.D
 Acq: 22 Sep 2016 14:09

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.2	18.7	28.1
125	10.0	10.5	15.7





#33

Benzo(k)fluoranthene

Concen: 2.87 ng

RT: 23.43 min Scan# 1487

Delta R.T. -0.01 min

Lab File: BE092394.D

Acq: 22 Sep 2016 14:09

Instrument :

BNA_E





ClientSampleId :

SSTDICC3.2

Tgt Ion:	252	Resp:	64072
Ion Ratio	Lower	Upper	
252	100		
253	21.8	20.3	30.5
125	10.2	9.6	14.4

Manual Integrations
APPROVED

sohil
9/23/2016 6:47:57 PM



#34

Benzo(a)pyrene

Concen: 2.96 ng




RT: 24.01 min Scan# 1537


Delta R.T. 0.00 min

Lab File: BE092394.D

Acq: 22 Sep 2016 14:09

Tgt Ion:	252	Resp:	61597
Ion Ratio	Lower	Upper	
252	100		
253	21.8	19.0	28.6
125	11.0	11.8	17.8#









#35
Dibenzo(a,h)anthracene
Concen: 2.99 ng
RT: 26.63 min Scan# 1718
Delta R.T. -0.02 min
Lab File: BE092394.D
Acq: 22 Sep 2016 14:09

Instrument :
BNA_E
ClientSampleId :
SSTDICC3.2

Tgt Ion	Ratio	Lower	Upper
278	100		
139	18.3	13.8	20.8
279	23.3	21.4	32.2

Manual Integrations
APPROVED

sohil
9/23/2016 6:47:57 PM



#36
Benzo(g,h,i)perylene
Concen: 2.93 ng
RT: 27.40 min Scan# 1763
Delta R.T. -0.00 min
Lab File: BE092394.D
Acq: 22 Sep 2016 14:09

Tgt Ion	Ratio	Lower	Upper
276	100		
277	24.9	19.8	29.8
138	19.7	19.8	29.6#

