

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE092315\
 Data File : BE090962.D
 Acq On : 23 Sep 2015 22:59
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 24 02:57:21 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 18 18:17:34 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	129	0.00
2	1,4-Dioxane	0.795	0.840	-5.7	133	0.00
3	n-Nitrosodimethylamine	1.067	0.841	21.2	105	0.00
4 S	2-Fluorophenol	1.238	0.954	22.9	101	0.00
5 S	Phenol-d6	1.559	1.325	15.0	113	0.03
6	bis(2-Chloroethyl)ether	1.546	1.574	-1.8	144	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	137	0.00
8 S	Nitrobenzene-d5	0.345	0.283	18.0	114	0.00
9	Nitrobenzene	0.383	0.279	27.2#	104	0.00
10	Naphthalene	1.126	1.106	1.8	137	0.00
11	Hexachlorobutadiene	0.166	0.188	-13.3	154#	0.00
12	2-Methylnaphthalene	0.672	0.663	1.3	139	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	150	0.00
14 S	2,4,6-Tribromophenol	0.196	0.135	31.1#	118	0.00
15 S	2-Fluorobiphenyl	1.418	1.376	3.0	144	0.00
16	Acenaphthylene	8.705	8.406	3.4	146	0.00
17	Acenaphthene	1.341	1.379	-2.8	156#	0.00
18	Fluorene	1.705	1.606	5.8	145	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	137	0.00
20	4-Bromophenyl-phenylether	0.193	0.194	-0.5	139	0.00
21	Hexachlorobenzene	0.906	1.007	-11.1	151#	0.00
22	Pentachlorophenol	0.087	0.052	40.2#	90	0.04
23	Phenanthrene	1.292	1.253	3.0	136	0.00
24	Anthracene	1.144	1.152	-0.7	143	0.00
25	Fluoranthene	1.525	1.344	11.9	122	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
27	Pyrene	1.141	1.234	-8.2	121	0.00
28 S	Terphenyl-d14	0.638	0.633	0.8	113	0.01
29	Benzo(a)anthracene	1.235	1.045	15.4	92	0.00
30	Chrysene	1.352	1.538	-13.8	119	0.00
31	Bis(2-ethylhexyl)phthalate	0.515	0.392	23.9	90	0.00
32	Indeno(1,2,3-cd)pyrene	1.455	1.180	18.9	84	0.00
33 I	Perylene-d12	1.000	1.000	0.0	92	0.00
34	Benzo(b)fluoranthene	1.170	1.031	11.9	84	0.00
35	Benzo(k)fluoranthene	1.347	1.630	-21.0	112	0.00
36 C	Benzo(a)pyrene	1.113	1.032	7.3	89	0.00
37	Dibenzo(a,h)anthracene	1.158	0.999	13.7	81	0.00
38	Benzo(g,h,i)perylene	1.240	1.138	8.2	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0