

Data Path : Z:\HPCHEM1\BNA E\DATA\BE092316\
 Data File : BE092408.D
 Acq On : 23 Sep 2016 9:17
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 23 10:00:49 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 22 18:25:56 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	-0.02
2	1,4-Dioxane	0.635	0.666	-4.9	103	-0.01
3	n-Nitrosodimethylamine	0.588	0.461	21.6	86	0.00
4 S	2-Fluorophenol	0.999	0.829	17.0	88	0.00
5 S	Phenol-d6	1.388	1.236	11.0	101	0.00
6	bis(2-Chloroethyl)ether	1.222	0.829	32.2#	86	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
8 S	Nitrobenzene-d5	0.299	0.251	16.1	93	0.00
9	Naphthalene	1.082	1.079	0.3	93	0.00
10	Hexachlorobutadiene	0.170	0.168	1.2	93	0.00
11	2-Methylnaphthalene	0.716	0.711	0.7	94	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
13 S	2,4,6-Tribromophenol	0.144	0.115	20.1	89	0.01
14 S	2-Fluorobiphenyl	1.184	1.154	2.5	95	0.00
15	Acenaphthylene	7.240	6.966	3.8	95	0.00
16	Acenaphthene	1.134	1.147	-1.1	94	0.00
17	Fluorene	1.438	1.411	1.9	95	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.02
19	Hexachlorobenzene	0.214	0.210	1.9	98	0.00
20	Pentachlorophenol	0.050	0.037	26.0#	83	0.00
21	Phenanthrene	1.149	1.115	3.0	96	0.00
22	Anthracene	1.105	1.033	6.5	96	0.02
23	Fluoranthene	5.460	5.184	5.1	97	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
25	Pyrene	4.333	4.158	4.0	95	0.02
26 S	Terphenyl-d14	0.620	0.597	3.7	93	0.02
27	Benzo(a)anthracene	4.873	4.579	6.0	91	0.00
28	Chrysene	4.462	5.040	-13.0	100	0.00
29	Bis(2-ethylhexyl)phthalate	0.420	0.341	18.8	76	0.00
30	Indeno(1,2,3-cd)pyrene	4.734	4.306	9.0	88	0.05
31 I	Perylene-d12	1.000	1.000	0.0	93	0.02
32	Benzo(b)fluoranthene	1.211	1.093	9.7	85	0.02
33	Benzo(k)fluoranthene	1.267	1.261	0.5	100	0.02
34 C	Benzo(a)pyrene	1.187	1.112	6.3	92	0.02
35	Dibenzo(a,h)anthracene	1.088	0.992	8.8	89	0.05
36	Benzo(g,h,i)perylene	4.673	4.440	5.0	91	0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0