

Data Path : Z:\HPCHEM1\BNA E\DATA\BE092316\
 Data File : BE092423.D
 Acq On : 23 Sep 2016 19:50
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 23 23:17:43 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 22 18:25:56 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	98	-0.02
2	1,4-Dioxane	0.400	0.363	9.3	96	-0.01
3	n-Nitrosodimethylamine	0.400	0.407	-1.7	83	0.00
4 S	2-Fluorophenol	0.400	0.363	9.3	79	0.00
5 S	Phenol-d6	0.400	0.318	20.5	81	0.00
6	bis(2-Chloroethyl)ether	0.400	0.325	18.8	82	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	95	0.00
8 S	Nitrobenzene-d5	0.400	0.425	-6.2	97	0.00
9	Naphthalene	0.400	0.385	3.8	92	0.00
10	Hexachlorobutadiene	0.400	0.384	4.0	92	0.00
11	2-Methylnaphthalene	0.400	0.379	5.3	91	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	96	0.00
13 S	2,4,6-Tribromophenol	0.400	0.382	4.5	77	0.01
14 S	2-Fluorobiphenyl	0.400	0.384	4.0	93	0.01
15	Acenaphthylene	0.400	0.367	8.3	90	0.02
16	Acenaphthene	0.400	0.390	2.5	90	0.00
17	Fluorene	0.400	0.369	7.8	89	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	90	0.02
19	Hexachlorobenzene	0.400	0.385	3.8	92	0.00
20	Pentachlorophenol	0.400	0.322	19.5	62	0.00
21	Phenanthrene	0.400	0.383	4.3	90	0.02
22	Anthracene	0.400	0.356	11.0	88	0.02
23	Fluoranthene	0.400	0.359	10.3	87	0.02
24 I	Chrysene-d12	5.000	5.000	0.0	82	0.00
25	Pyrene	0.400	0.399	0.3	86	0.02
26 S	Terphenyl-d14	0.400	0.410	-2.5	85	0.02
27	Benzo(a)anthracene	0.400	0.347	13.3	73	0.00
28	Chrysene	0.400	0.443	-10.7	85	0.02
29	Bis(2-ethylhexyl)phthalate	0.400	0.309	22.8	63	0.02
30	Indeno(1,2,3-cd)pyrene	0.400	0.359	10.3	75	0.07
31 I	Perylene-d12	5.000	5.000	0.0	88	0.05
32	Benzo(b)fluoranthene	0.400	0.347	13.3	76	0.03
33	Benzo(k)fluoranthene	0.400	0.375	6.3	88	0.03
34 C	Benzo(a)pyrene	0.400	0.355	11.3	82	0.03
35	Dibenzo(a,h)anthracene	0.400	0.341	14.8	78	0.07
36	Benzo(g,h,i)perylene	0.400	0.355	11.3	80	0.07

(#) = Out of Range

SPCC's out = 0 CCC's out = 0