

Data Path : Z:\HPCHEM1\BNA E\DATA\BE092317\
 Data File : BE094096.D
 Acq On : 24 Sep 2017 2:18
 Operator : SJ/JU
 Sample : I5340-10
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :

Quant Time: Sep 25 08:50:37 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 25 07:59:29 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.91	152	2142	0.40	ng	0.00
7) Naphthalene-d8	10.69	136	11975	0.40	ng	0.00
13) Acenaphthene-d10	14.51	164	7748	0.40	ng	0.00
19) Phenanthrene-d10	17.24	188	24126	0.40	ng	0.00
27) Chrysene-d12	21.40	240	14528	0.40	ng	0.00
34) Perylene-d12	23.91	264	13492	0.40	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.50	112	1980	0.05	ng	0.00
5) Phenol-d6	7.09	99	3281	0.26	ng	0.00
8) Nitrobenzene-d5	9.06	82	2981	0.28	ng	0.00
11) 2-Methylnaphthalene-d10	12.27	152	4399	0.22	ng	0.00
14) 2,4,6-Tribromophenol	16.00	330	1685	0.29	ng	0.00
15) 2-Fluorobiphenyl	13.13	172	6018	0.24	ng	0.00
25) Fluoranthene-d10	19.27	212	34992	0.11	ng	0.00
29) Terphenyl-d14	19.85	244	4837	0.21	ng	0.00


Target Compounds				Qvalue		
3) n-Nitrosodimethylamine	3.79	42	54	Below Cal	#	1
6) bis(2-Chloroethyl)ether	7.34	93	859	0.091 ng	#	1
9) Naphthalene	10.74	128	5167	Below Cal	#	95
12) 2-Methylnaphthalene	12.35	142	779	0.032 ng	#	92
16) Acenaphthylene	14.23	152	5944	0.152 ng		99
17) Acenaphthene	14.57	154	1214	0.047 ng		93
18) Fluorene	15.56	166	2428	0.063 ng		90
23) Phenanthrene	17.27	178	62208	0.720 ng		98
24) Anthracene	17.37	178	9063	0.136 ng	#	93
26) Fluoranthene	19.30	202	114770	1.238 ng		99
28) Pvrene	19.66	202	110160	2.741 ng	#	96
30) Benzo(a)anthracene	21.38	228	45010	0.925 ng		96
31) Chrysene	21.44	228	55615	1.166 ng		97
32) Bis(2-ethylhexyl)phthalate	21.29	149	29018	0.222 ng		99
33) Indeno(1,2,3-cd)pvrene	26.55	276	31247	0.848 ng		97
35) Benzo(b)fluoranthene	23.14	252	67526	1.383 ng		95
36) Benzo(k)fluoranthene	23.14	252	67526	1.378 ng		97
37) Benzo(a)pvrene	23.79	252	46014	1.049 ng		98
38) Dibenzo(a,h)anthracene	26.56	278	7874	0.224 ng	#	70
39) Benzo(g,h,i)perylene	27.38	276	32036	0.913 ng		99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE092317\
Data File : BE094096.D
Acq On : 24 Sep 2017 2:18
Operator : SJ/JU
Sample : I5340-10
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :

Quant Time: Sep 25 08:50:37 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 25 07:59:29 2017
Response via : Initial Calibration









#1

1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.91 min Scan# 401
Delta R.T. -0.00 min
Lab File: BE094096.D
Acq: 24 Sep 2017 2:18

Instrument :
BNA_E
ClientSampled :




Tot Ion:152 Resp: 2142
Ion Ratio Lower Upper
152 100
150 151.2 117.2 175.8
115 68.3 57.0 85.6




#3

n-Nitrosodimethylamine
Concen: Below Cal
RT: 3.79 min Scan# 74
Delta R.T. 0.06 min
Lab File: BE094096.D
Acq: 24 Sep 2017 2:18

Tgt Ion: 42 Resp: 54
Ion Ratio Lower Upper
42 100
74 0.0 100.3 150.5#
44 181.5 17.0 25.4#





#4

2-Fluorophenol

Concen: 0.051 ng

RT: 5.50 min Scan# 210

Delta R.T. -0.00 min

Lab File: BE094096.D

Acq: 24 Sep 2017 2:18

Instrument :

BNA_E

ClientSampled :


Tot Ion:112 Resp: 1980

Ion Ratio Lower Upper

112 100

64 77.8 60.1 90.1




63 148.2 116.8 175.2



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BE0

Ion 63.00 (62.70 to 63.70): BE0



#5

Phenol-d6

Concen: 0.259 ng

RT: 7.09 min Scan# 336

Delta R.T. -0.00 min

Lab File: BE094096.D

Acq: 24 Sep 2017 2:18


Tgt Ion: 99 Resp: 3281

Ion Ratio Lower Upper

99 100

42 29.8 20.2 30.2



71 34.3 28.4 42.6




Abundance Ion 99.00 (98.70 to 99.70): BE0

Ion 42.00 (41.70 to 42.70): BE0

Ion 71.00 (70.70 to 71.70): BE0





#6

bis(2-Chloroethyl)ether

Concen: 0.091 ng

RT: 7.34 min Scan# 356

Delta R.T. 0.01 min

Lab File: BE094096.D

Acq: 24 Sep 2017 2:18

Instrument :

BNA_E

ClientSampled :


Tot Ion: 93 Resp: 859

Ion Ratio Lower Upper

93 100

63 15.3 275.3 412.9#




95 3.1 25.2 37.8#



Abundance Ion 93.00 (92.70 to 93.70): BE0

Ion 63.00 (62.70 to 63.70): BE0

Ion 95.00 (94.70 to 95.70): BE0



#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.69 min Scan# 576

Delta R.T. -0.00 min

Lab File: BE094096.D

Acq: 24 Sep 2017 2:18

Tgt Ion: 136 Resp: 11975


Ion Ratio Lower Upper

136 100

137 11.4 9.5 14.3

54 44.1 29.1 43.7#

68 9.0 6.2 9.2






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0





#8

Nitrobenzene-d5

Concen: 0.280 ng

RT: 9.06 min Scan# 479

Delta R.T. -0.00 min

Lab File: BE094096.D

Acq: 24 Sep 2017 2:18

Instrument :

BNA_E

ClientSampled :


Tot Ion: 82 Resp: 2981

Ion Ratio Lower Upper

82 100

128 170.2 150.0 225.0




54 193.3 154.2 231.4



Abundance Ion 82.00 (81.70 to 82.70): BE094096.D (-467) (-)

Ion 128.00 (127.70 to 128.70): BE094096.D (-467) (-)

Ion 54.00 (53.70 to 54.70): BE094096.D (-467) (-)



#9

Naphthalene

Concen: Below Cal

RT: 10.74 min Scan# 579

Delta R.T. -0.00 min

Lab File: BE094096.D

Acq: 24 Sep 2017 2:18


Tgt Ion: 128 Resp: 5167

Ion Ratio Lower Upper

128 100

129 4.9 2.6 3.8#



127 4.8 2.8 4.2#




Abundance Ion 128.00 (127.70 to 128.70): BE094096.D (-567) (-)

Ion 129.00 (128.70 to 129.70): BE094096.D (-567) (-)

Ion 127.00 (126.70 to 127.70): BE094096.D (-567) (-)





#11

2-Methylnaphthalene-d10

Concen: 0.217 ng

RT: 12.27 min Scan# 820

Delta R.T. 0.00 min

Lab File: BE094096.D

Acq: 24 Sep 2017 2:18

Instrument :

BNA_E





ClientSampleId :

Tot Ion:152 Resp: 4399

Ion Ratio Lower Upper

152 100

151 19.2 15.4 23.0



#12

2-Methylnaphthalene

Concen: 0.032 ng

RT: 12.35 min Scan# 840

Delta R.T. 0.00 min

Lab File: BE094096.D

Acq: 24 Sep 2017 2:18




Tgt Ion:142 Resp: 779


Ion Ratio Lower Upper

142 100

141 90.4 70.4 105.6

115 51.8 31.7 47.5#









#13
 Acenaphthene-d10
 Concen: 0.400 ng
 RT: 14.51 min Scan# 1111
 Delta R.T. 0.00 min
 Lab File: BE094096.D
 Acq: 24 Sep 2017 2:18




Instrument :
 BNA_E
 ClientSampleId :


Tot Ion:164 Resp: 7748
 Ion Ratio Lower Upper
 164 100
 162 104.1 83.1 124.7
 160 46.6 37.1 55.7



#14
 2,4,6-Tribromophenol
 Concen: 0.287 ng
 RT: 16.00 min Scan# 1205
 Delta R.T. 0.00 min
 Lab File: BE094096.D
 Acq: 24 Sep 2017 2:18

Tgt Ion:330 Resp: 1685
 Ion Ratio Lower Upper
 330 100
 332 93.2 152.7 229.1#
 141 0.0 33.7 50.5#









#15
2-Fluorobiphenyl
Concen: 0.236 ng
RT: 13.13 min Scan# 1018
Delta R.T. 0.00 min
Lab File: BE094096.D
Acq: 24 Sep 2017 2:18




Instrument :
BNA_E
ClientSampleId :


Tot Ion:172 Resp: 6018
Ion Ratio Lower Upper
172 100
171 38.0 29.9 44.9
170 25.8 21.8 32.8



#16
Acenaphthylene
Concen: 0.152 ng
RT: 14.23 min Scan# 1092
Delta R.T. 0.00 min
Lab File: BE094096.D
Acq: 24 Sep 2017 2:18

Tgt Ion:152 Resp: 5944
Ion Ratio Lower Upper
152 100
151 20.3 15.7 23.5
153 12.9 10.5 15.7





#17

Acenaphthene

Concen: 0.047 ng

RT: 14.57 min Scan# 1115

Delta R.T. 0.00 min

Lab File: BE094096.D

Acq: 24 Sep 2017 2:18

Instrument :

BNA_E

ClientSampleId :


Tot Ion:154 Resp: 1214

Ion Ratio Lower Upper

154 100

153 107.2 89.2 133.8




152 61.8 42.0 63.0



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#18

Fluorene

Concen: 0.063 ng

RT: 15.56 min Scan# 1182

Delta R.T. 0.02 min

Lab File: BE094096.D

Acq: 24 Sep 2017 2:18


Tgt Ion:166 Resp: 2428

Ion Ratio Lower Upper

166 100

165 105.8 77.8 116.6



167 61.8 42.4 63.6




Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.24 min Scan# 1243

Delta R.T. 0.00 min

Lab File: BE094096.D

Acq: 24 Sep 2017 2:18

Instrument :

BNA_E

ClientSampleId :


Tot Ion:188 Resp: 24126

Ion Ratio Lower Upper

188 100

94 5.7 3.9 5.9




80 5.1 3.6 5.4



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BE0

Ion 80.00 (79.70 to 80.70): BE0



#23

Phenanthrene

Concen: 0.720 ng

RT: 17.27 min Scan# 1244

Delta R.T. 0.00 min

Lab File: BE094096.D

Acq: 24 Sep 2017 2:18


Tgt Ion:178 Resp: 62208

Ion Ratio Lower Upper

178 100

176 20.0 15.1 22.7



179 15.5 11.8 17.6




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#24

Anthracene

Concen: 0.136 ng

RT: 17.37 min Scan# 1247

Delta R.T. 0.00 min

Lab File: BE094096.D

Acq: 24 Sep 2017 2:18

Instrument :

BNA_E

ClientSampleId :





Tot Ion:178 Resp: 9063

Ion Ratio Lower Upper

178 100

176 17.7 12.8 19.2

179 11.7 13.0 19.6#



#25

Fluoranthene-d10

Concen: 0.113 ng

RT: 19.27 min Scan# 1431

Delta R.T. 0.01 min

Lab File: BE094096.D

Acq: 24 Sep 2017 2:18




Tgt Ion:212 Resp: 34992


Ion Ratio Lower Upper

212 100

106 3.8 2.8 4.2

104 2.3 1.6 2.4









#26
 Fluoranthene
 Concen: 1.238 ng
 RT: 19.30 min Scan# 1435
 Delta R.T. 0.01 min
 Lab File: BE094096.D
 Acq: 24 Sep 2017 2:18




Instrument :
 BNA_E
 ClientSampled :


Tot Ion:202 Resp: 114770
 Ion Ratio Lower Upper
 202 100
 101 12.4 9.8 14.8
 203 17.5 13.7 20.5



#27
 Chrysene-d12
 Concen: 0.400 ng
 RT: 21.40 min Scan# 1654
 Delta R.T. -0.00 min
 Lab File: BE094096.D
 Acq: 24 Sep 2017 2:18

Tgt Ion:240 Resp: 14528
 Ion Ratio Lower Upper
 240 100
 120 12.2 5.5 8.3#
 236 27.8 20.7 31.1









#28
Pvrene
Concen: 2.741 ng
RT: 19.66 min Scan# 1484
Delta R.T. 0.01 min
Lab File: BE094096.D
Acq: 24 Sep 2017 2:18

Instrument :
BNA_E
ClientSampleId :

Tot Ion:202 Resp: 110160
Ion Ratio Lower Upper
202 100
200 21.0 16.6 25.0
203 21.2 13.8 20.8#




Abundance Ion 202.00 (201.70 to 202.70): E
Ion 200.00 (199.70 to 200.70): E
Ion 203.00 (202.70 to 203.70): E






#29
Terphenyl-d14
Concen: 0.213 ng
RT: 19.85 min Scan# 1510
Delta R.T. -0.00 min
Lab File: BE094096.D
Acq: 24 Sep 2017 2:18

Tgt Ion:244 Resp: 4837
Ion Ratio Lower Upper
244 100
212 46.0 25.4 38.0#
122 18.0 8.1 12.1#



Abundance Ion 244.00 (243.70 to 244.70): E
Ion 212.00 (211.70 to 212.70): E
Ion 122.00 (121.70 to 122.70): E





#30

Benzo(a)anthracene

Concen: 0.925 ng

RT: 21.38 min Scan# 1652

Delta R.T. -0.00 min

Lab File: BE094096.D

Acq: 24 Sep 2017 2:18

Instrument :

BNA_E

ClientSampleId :


Tot Ion:228 Resp: 45010

Ion Ratio Lower Upper

228 100

226 29.4 21.4 32.0




229 22.0 16.4 24.6



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#31

Chrysene

Concen: 1.166 ng

RT: 21.44 min Scan# 1657

Delta R.T. -0.00 min

Lab File: BE094096.D

Acq: 24 Sep 2017 2:18


Tgt Ion:228 Resp: 55615

Ion Ratio Lower Upper

228 100

226 31.4 23.6 35.4



229 22.1 16.4 24.6




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#32

Bis(2-ethylhexyl)phthalate

Concen: 0.222 ng

RT: 21.29 min Scan# 1644

Delta R.T. -0.00 min

Lab File: BE094096.D

Acq: 24 Sep 2017 2:18

Instrument :

BNA_E

ClientSampleId :


Tot Ion:149 Resp: 29018

Ion Ratio Lower Upper

149 100

167 7.2 5.4 8.0




279 1.0 0.7 1.1



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#33

Indeno(1,2,3-cd)pyrene

Concen: 0.848 ng

RT: 26.55 min Scan# 2697

Delta R.T. 0.01 min

Lab File: BE094096.D

Acq: 24 Sep 2017 2:18


Tgt Ion:276 Resp: 31247

Ion Ratio Lower Upper

276 100

138 25.0 22.5 33.7



277 25.0 19.9 29.9




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#34
Pervlene-d12
Concen: 0.400 ng
RT: 23.91 min Scan# 2122
Delta R.T. 0.01 min
Lab File: BE094096.D
Acq: 24 Sep 2017 2:18




Instrument :
BNA_E
ClientSampleId :


Tot Ion:264 Resp: 13492
Ion Ratio Lower Upper
264 100
260 27.5 20.5 30.7
265 34.5 22.8 34.2#



#35
Benzo(b)fluoranthene
Concen: 1.383 ng
RT: 23.14 min Scan# 1953
Delta R.T. 0.01 min
Lab File: BE094096.D
Acq: 24 Sep 2017 2:18

Tgt Ion:252 Resp: 67526
Ion Ratio Lower Upper
252 100
253 24.8 18.2 27.2
125 14.6 10.1 15.1





#36

Benzo(k)fluoranthene

Concen: 1.378 ng

RT: 23.14 min Scan# 1953

Delta R.T. -0.04 min

Lab File: BE094096.D

Acq: 24 Sep 2017 2:18

Instrument :

BNA_E

ClientSampleId :


Tot Ion:252 Resp: 67526

Ion Ratio Lower Upper

252 100

253 24.8 18.6 27.8




125 14.6 10.7 16.1



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#37

Benzo(a)pyrene

Concen: 1.049 ng

RT: 23.79 min Scan# 2096

Delta R.T. 0.00 min

Lab File: BE094096.D

Acq: 24 Sep 2017 2:18


Tgt Ion:252 Resp: 46014

Ion Ratio Lower Upper

252 100

253 24.6 19.0 28.6



125 16.0 11.8 17.8




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#38
Dibenzo(a,h)anthracene
Concen: 0.224 ng
RT: 26.56 min Scan# 2700
Delta R.T. -0.00 min
Lab File: BE094096.D
Acq: 24 Sep 2017 2:18

Instrument :
BNA_E
ClientSampleId :

Tot Ion:278 Resp: 7874
Ion Ratio Lower Upper
278 100
139 37.9 16.3 24.5#
279 36.3 19.9 29.9#



#39
Benzo(g,h,i)perylene
Concen: 0.913 ng
RT: 27.38 min Scan# 2879
Delta R.T. 0.01 min
Lab File: BE094096.D
Acq: 24 Sep 2017 2:18

Tgt Ion:276 Resp: 32036
Ion Ratio Lower Upper
276 100
277 25.9 20.8 31.2
138 28.2 21.5 32.3

