


n-Nitrosodimethylamine

Response Ratio




Resp Ratio = 8.48e-001 * Amt + 1.28e-001
Coef of Det (r^2) = 0.999 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092317.M
Calibration Table Last Updated: Mon Sep 25 07:59:29 2017


2-Fluorophenol

Response Ratio



Resp Ratio = 1.42e+000 * Amt + 7.43e-001
Coef of Det (r^2) = 0.991 Curve Fit: Linear


Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092317.M
Calibration Table Last Updated: Mon Sep 25 07:59:29 2017



Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092317.M
Calibration Table Last Updated: Mon Sep 25 07:59:29 2017

1,4-Dioxane

Response Ratio



Resp Ratio = $5.70 \times 10^{-1} \times \text{Amt} + 2.63 \times 10^{-1}$
Coef of Det (r^2) = 0.997 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092317.M
Calibration Table Last Updated: Mon Sep 25 07:59:29 2017