

Data Path : Z:\HPCHEM1\BNA E\Data\BE092317\
 Data File : BE094081.D
 Acq On : 23 Sep 2017 17:00
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 25 13:09:31 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 25 13:07:29 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.835	0.784	6.1	95	0.00
3	n-Nitrosodimethylamine	1.034	0.841	18.7	97	0.00
4 S	2-Fluorophenol	2.100	1.985	5.5	94	0.00
5 S	Phenol-d6	2.365	2.188	7.5	93	0.00
6	bis(2-Chloroethyl)ether	1.767	1.631	7.7	92	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
8 S	Nitrobenzene-d5	0.356	0.333	6.5	89	0.00
9	Naphthalene	5.096	4.705	7.7	89	0.00
10	Hexachlorobutadiene	0.162	0.166	-2.5	97	0.00
11 SURR	2-Methylnaphthalene-d10	0.678	0.656	3.2	93	0.00
12	2-Methylnaphthalene	0.808	0.774	4.2	92	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.00
14 S	2,4,6-Tribromophenol	0.303	0.109	64.0#	41#	0.00
15 S	2-Fluorobiphenyl	1.314	1.338	-1.8	93	0.00
16	Acenaphthylene	2.014	1.918	4.8	88	0.00
17	Acenaphthene	1.336	1.279	4.3	88	0.00
18	Fluorene	1.989	1.855	6.7	88	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	87	0.00
20	4-Bromophenyl-phenylether	0.214	0.214	0.0	92	0.00
21	Hexachlorobenzene	0.157	0.174	-10.8	106	0.00
22	Pentachlorophenol	0.011	0.257	-2236.4#	3032#	0.00
23	Phenanthrene	1.433	1.430	0.2	91	0.00
24	Anthracene	1.105	1.076	2.6	81	0.00
25 SURR	Fluoranthene-d10	5.132	5.267	-2.6	96	0.00
26	Fluoranthene	1.537	1.579	-2.7	96	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
28	Pvrene	1.107	1.118	-1.0	96	0.00
29 S	Terphenyl-d14	0.627	0.639	-1.9	97	0.00
30	Benzo(a)anthracene	1.340	1.341	-0.1	101	0.00
31	Chrysene	1.313	1.296	1.3	97	0.00
32	Bis(2-ethylhexyl)phthalate	3.600	3.169	12.0	88	0.00
33	Indeno(1,2,3-cd)pyrene	1.015	0.957	5.7	98	0.00
34 I	Perylene-d12	1.000	1.000	0.0	96	0.00
35	Benzo(b)fluoranthene	1.448	1.468	-1.4	96	0.00
36	Benzo(k)fluoranthene	1.453	1.447	0.4	95	0.00
37 C	Benzo(a)pyrene	1.300	1.318	-1.4	98	0.00
38	Dibenzo(a,h)anthracene	1.041	1.069	-2.7	97	0.00
39	Benzo(a,h,i)perylene	1.040	1.078	-3.7	97	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE092317\
Data File : BE094081.D
Acq On : 23 Sep 2017 17:00
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 25 13:09:31 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 25 13:07:29 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			