

Data Path : Z:\HPCHEM1\BNA E\DATA\BE092317\
 Data File : BE094094.D
 Acq On : 24 Sep 2017 1:03
 Operator : SJ/JU
 Sample : I5340-12MSD
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 LT-TS-012MSD

Manual Integrations
 APPROVED

Sohil
 9/25/2017 6:24:50 PM

Quant Time: Sep 25 13:34:16 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092317.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 25 13:07:29 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.91	152	2262	0.40	ng	0.00
7) Naphthalene-d8	10.69	136	12554	0.40	ng	0.00
13) Acenaphthene-d10	14.51	164	7964	0.40	ng	0.00
19) Phenanthrene-d10	17.23	188	25446	0.40	ng	0.00
27) Chrysene-d12	21.40	240	15253	0.40	ng	0.00
34) Perylene-d12	23.91	264	14663	0.40	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.50	112	3687	0.25	ng	0.00
5) Phenol-d6	7.09	99	5021	0.38	ng	0.00
8) Nitrobenzene-d5	9.05	82	4344	0.39	ng	-0.02
11) 2-Methylnaphthalene-d10	12.27	152	7443	0.35	ng	0.00
14) 2,4,6-Tribromophenol	16.00	330	1432	0.24	ng	0.00
15) 2-Fluorobiphenyl	13.13	172	10139	0.39	ng	0.00
25) Fluoranthene-d10	19.27	212	70977	0.22	ng	0.00
29) Terphenyl-d14	19.85	244	10832	0.45	ng	0.00

Target Compounds				Qvalue		
3) n-Nitrosodimethylamine	3.65	42	897	0.127	ng	# 1
9) Naphthalene	10.74	128	20603	0.129	ng	100
12) 2-Methylnaphthalene	12.34	142	3378	0.133	ng	98
16) Acenaphthylene	14.23	152	13943	0.348	ng	99
17) Acenaphthene	14.57	154	23336	0.877	ng	96
18) Fluorene	15.56	166	27632	0.698	ng	95
22) Pentachlorophenol	16.91	266	33	0.046	ng	99
23) Phenanthrene	17.27	178	623626	6.842	ng	97
24) Anthracene	17.37	178	139566	1.986	ng	# 93
26) Fluoranthene	19.30	202	851320	8.705	ng	99
28) Pyrene	19.66	202	737684	17.480	ng	# 93
30) Benzo(a)anthracene	21.38	228	312021	6.108	ng	98
31) Chrysene	21.44	228	281877	5.631	ng	98
32) Bis(2-ethylhexyl)phthalate	21.29	149	49721	0.362	ng	100
33) Indeno(1,2,3-cd)pyrene	26.55	276	173279	4.478	ng	96
35) Benzo(b)fluoranthene	23.14	252	355337	6.696	ng	99
36) Benzo(k)fluoranthene	23.19	252	132324m	2.485	ng	
37) Benzo(a)pyrene	23.80	252	260881	5.475	ng	98
38) Dibenzo(a,h)anthracene	26.56	278	41637	1.091	ng	95
39) Benzo(g,h,i)perylene	27.38	276	173897	4.560	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA E\DATA\BE092317\
Data File : BE094094.D
Acq On : 24 Sep 2017 1:03
Operator : SJ/JU
Sample : I5340-12MSD
Misc :
ALS Vial : 21 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
LT-TS-012MSD

Manual Integrations
APPROVED

Sohil
9/25/2017 6:24:50 PM

Quant Time: Sep 25 13:34:16 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092317.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 25 13:07:29 2017
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 0.400 ng

RT: 7.91 min Scan# 401

Delta R.T. -0.00 min

Lab File: BE094094.D

Acq: 24 Sep 2017 1:03

Instrument :

BNA_E

ClientSampleId :

LT-TS-012MSD

Tot Ion:152 Resp: 2262

Ion Ratio Lower Upper

152 100


150 154.6 117.2 175.8

115 69.7 57.0 85.6

Manual Integrations
APPROVED

Sohil




9/25/2017 6:24:50 PM



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#3

n-Nitrosodimethylamine

Concen: 0.127 ng

RT: 3.65 min Scan# 63

Delta R.T. -0.08 min

Lab File: BE094094.D

Acq: 24 Sep 2017 1:03


Tgt Ion: 42 Resp: 897

Ion Ratio Lower Upper

42 100

74 4.3 100.3 150.5#



44 150.8 17.0 25.4#




Abundance Ion 42.00 (41.70 to 42.70): BE0

Ion 74.00 (73.70 to 74.70): BE0

Ion 44.00 (43.70 to 44.70): BE0





#4

2-Fluorophenol

Concen: 0.249 ng

RT: 5.50 min Scan# 210

Delta R.T. -0.00 min

Lab File: BE094094.D

Acq: 24 Sep 2017 1:03

Instrument :

BNA_E





ClientSampleId :

LT-TS-012MSD

Tot Ion:	112	Resp:	3687
Ion Ratio	Lower	Upper	
112	100		
64	74.7	60.1	90.1
63	153.9	116.8	175.2

Manual Integrations
APPROVED

Sohil
9/25/2017 6:24:50 PM



#5

Phenol-d6

Concen: 0.375 ng




RT: 7.09 min Scan# 336


Delta R.T. -0.00 min

Lab File: BE094094.D

Acq: 24 Sep 2017 1:03

Tgt Ion:	99	Resp:	5021
Ion Ratio	Lower	Upper	
99	100		
42	28.7	20.2	30.2
71	35.1	28.4	42.6





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.69 min Scan# 576

Delta R.T. 0.00 min

Lab File: BE094094.D

Acq: 24 Sep 2017 1:03

Instrument :

BNA_E


ClientSampled :

LT-TS-012MSD

Tot Ion	Ratio	Resp	Lower	Upper
136	100	12554		
137	11.5	9.5	14.3	
54	37.5	29.1	43.7	
68	8.2	6.2	9.2	

Manual Integrations
APPROVED

Sohil
9/25/2017 6:24:50 PM






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0



#8

Nitrobenzene-d5

Concen: 0.389 ng


RT: 9.05 min Scan# 478

Delta R.T. -0.02 min

Lab File: BE094094.D

Acq: 24 Sep 2017 1:03



Tgt Ion	Ratio	Resp	Lower	Upper
82	100	4344		
128	102.7	150.0	225.0#	
54	291.5	154.2	231.4#	




Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0





#9

Naphthalene

Concen: 0.129 ng

RT: 10.74 min Scan# 579

Delta R.T. 0.00 min

Lab File: BE094094.D

Acq: 24 Sep 2017 1:03

Instrument :

BNA_E





ClientSampleId :

LT-TS-012MSD

Tot Ion:	128	Resp:	20603
Ion Ratio	Lower	Upper	
128	100		
129	3.3	2.6	3.8
127	3.5	2.8	4.2

Manual Integrations
APPROVED

Sohil
9/25/2017 6:24:50 PM



#11

2-Methylnaphthalene-d10

Concen: 0.350 ng




RT: 12.27 min Scan# 820


Delta R.T. -0.00 min

Lab File: BE094094.D

Acq: 24 Sep 2017 1:03

Tgt Ion:	152	Resp:	7443
Ion Ratio	Lower	Upper	
152	100		
151	18.9	15.4	23.0









#12
2-Methylnaphthalene
Concen: 0.133 ng
RT: 12.34 min Scan# 839
Delta R.T. -0.00 min
Lab File: BE094094.D
Acq: 24 Sep 2017 1:03

Instrument :
BNA_E
ClientSampleId :
LT-TS-012MSD

Tot Ion	Ratio	Lower	Upper
142	100		
141	88.9	70.4	105.6
115	42.9	31.7	47.5




Manual Integrations
APPROVED




Sohil
9/25/2017 6:24:50 PM



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.51 min Scan# 1111
Delta R.T. 0.00 min
Lab File: BE094094.D
Acq: 24 Sep 2017 1:03

Tgt Ion	Ratio	Lower	Upper
164	100		
162	104.0	83.1	124.7
160	46.8	37.1	55.7









#14
2,4,6-Tribromophenol
Concen: 0.237 ng
RT: 16.00 min Scan# 1205
Delta R.T. 0.00 min
Lab File: BE094094.D
Acq: 24 Sep 2017 1:03

Instrument :
BNA_E
ClientSampleId :
LT-TS-012MSD

Tot Ion	Ratio	Lower	Upper
330	100		
332	87.3	152.7	229.1#
141	0.0	33.7	50.5#


Manual Integrations
APPROVED


Sohil
9/25/2017 6:24:50 PM



#15
2-Fluorobiphenyl
Concen: 0.388 ng
RT: 13.13 min Scan# 1018
Delta R.T. 0.00 min
Lab File: BE094094.D
Acq: 24 Sep 2017 1:03

Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.8	29.9	44.9
170	25.9	21.8	32.8






#16
Acenaphthylene
Concen: 0.348 ng
RT: 14.23 min Scan# 1092
Delta R.T. 0.00 min
Lab File: BE094094.D
Acq: 24 Sep 2017 1:03

Instrument :
BNA_E
ClientSampleId :
LT-TS-012MSD

Tot Ion	Ratio	Lower	Upper
152	100		
151	19.8	15.7	23.5
153	14.0	10.5	15.7

Manual Integrations
APPROVED

Sohil
9/25/2017 6:24:50 PM






Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#17
Acenaphthene
Concen: 0.877 ng
RT: 14.57 min Scan# 1115
Delta R.T. 0.00 min
Lab File: BE094094.D
Acq: 24 Sep 2017 1:03




Tgt Ion	Ratio	Lower	Upper
154	100		
153	115.0	89.2	133.8
152	56.5	42.0	63.0


Abundance

Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E









#18
Fluorene
 Concen: 0.698 ng
 RT: 15.56 min Scan# 1182
 Delta R.T. 0.02 min
 Lab File: BE094094.D
 Acq: 24 Sep 2017 1:03

Instrument :
 BNA_E
 ClientSampleId :
 LT-TS-012MSD

Tot Ion	Ratio	Resp	Lower	Upper
166	100	27632		
165	99.0	77.8	116.6	
167	59.7	42.4	63.6	




Manual Integrations
 APPROVED




Sohil
 9/25/2017 6:24:50 PM



#19
Phenanthrene-d10
 Concen: 0.400 ng
 RT: 17.23 min Scan# 1243
 Delta R.T. 0.00 min
 Lab File: BE094094.D
 Acq: 24 Sep 2017 1:03

Tgt Ion	Ratio	Resp	Lower	Upper
188	100	25446		
94	6.2	3.9	5.9#	
80	5.9	3.6	5.4#	









#22
 Pentachlorophenol
 Concen: 0.046 ng
 RT: 16.91 min Scan# 1233
 Delta R.T. 0.00 min
 Lab File: BE094094.D
 Acq: 24 Sep 2017 1:03

Instrument :
 BNA_E
 ClientSampleId :
 LT-TS-012MSD

Tot Ion	Ratio	Lower	Upper
266	100		
264	89.9	72.5	108.7
268	92.8	73.8	110.6


Manual Integrations
 APPROVED


Sohil
 9/25/2017 6:24:50 PM



#23
 Phenanthrene
 Concen: 6.842 ng
 RT: 17.27 min Scan# 1244
 Delta R.T. 0.00 min
 Lab File: BE094094.D
 Acq: 24 Sep 2017 1:03

Tgt Ion	Ratio	Lower	Upper
178	100		
176	21.2	15.1	22.7
179	15.1	11.8	17.6





#24

Anthracene

Concen: 1.986 ng

RT: 17.37 min Scan# 1247

Delta R.T. 0.00 min

Lab File: BE094094.D

Acq: 24 Sep 2017 1:03

Instrument :

BNA_E





ClientSampleId :

LT-TS-012MSD

Tot Ion:	178	Resp:	139566
Ion Ratio	Lower	Upper	
178	100		
176	16.7	12.8	19.2
179	21.5	13.0	19.6#

Manual Integrations
APPROVED

Sohil
9/25/2017 6:24:50 PM



#25

Fluoranthene-d10

Concen: 0.217 ng




RT: 19.27 min Scan# 1431


Delta R.T. 0.01 min

Lab File: BE094094.D

Acq: 24 Sep 2017 1:03

Tgt Ion:	212	Resp:	70977
Ion Ratio	Lower	Upper	
212	100		
106	3.6	2.8	4.2
104	2.2	1.6	2.4









#26
Fluoranthene
Concen: 8.705 ng
RT: 19.30 min Scan# 1435
Delta R.T. 0.01 min
Lab File: BE094094.D
Acq: 24 Sep 2017 1:03

Instrument :
BNA_E
ClientSampleId :
LT-TS-012MSD

Tot Ion	Ratio	Resp	Lower	Upper
202	100	851320		
101	12.5		9.8	14.8
203	17.6		13.7	20.5




Manual Integrations
APPROVED


Sohil
9/25/2017 6:24:50 PM



#27
Chrysene-d12
Concen: 0.400 ng
RT: 21.40 min Scan# 1654
Delta R.T. -0.00 min
Lab File: BE094094.D
Acq: 24 Sep 2017 1:03

Tgt Ion	Ratio	Resp	Lower	Upper
240	100	15253		
120	12.8		5.5	8.3#
236	28.7		20.7	31.1









#28
Pvrene
Concen: 17.480 ng
RT: 19.66 min Scan# 1484
Delta R.T. 0.01 min
Lab File: BE094094.D
Acq: 24 Sep 2017 1:03

Instrument :
BNA_E
ClientSampleId :
LT-TS-012MSD

Tot Ion	Ratio	Resp Lower	Upper
202	100		
200	21.4	16.6	25.0
203	22.9	13.8	20.8




Manual Integrations
APPROVED


Sohil
9/25/2017 6:24:50 PM



#29
Terphenyl-d14
Concen: 0.453 ng
RT: 19.85 min Scan# 1510
Delta R.T. 0.00 min
Lab File: BE094094.D
Acq: 24 Sep 2017 1:03

Tgt Ion	Ratio	Resp Lower	Upper
244	100		
212	38.8	25.4	38.0
122	12.9	8.1	12.1









#30
Benzo(a)anthracene
Concen: 6.108 ng
RT: 21.38 min Scan# 1652
Delta R.T. -0.00 min
Lab File: BE094094.D
Acq: 24 Sep 2017 1:03

Instrument :
BNA_E
ClientSampleId :
LT-TS-012MSD

Tot Ion	Ratio	Lower	Upper
228	100		
226	28.5	21.4	32.0
229	20.4	16.4	24.6




Manual Integrations
APPROVED


Sohil
9/25/2017 6:24:50 PM



#31
Chrysene
Concen: 5.631 ng
RT: 21.44 min Scan# 1657
Delta R.T. -0.00 min
Lab File: BE094094.D
Acq: 24 Sep 2017 1:03

Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.9	23.6	35.4
229	21.5	16.4	24.6





#32

Bis(2-ethylhexyl)phthalate

Concen: 0.362 ng

RT: 21.29 min Scan# 1644

Delta R.T. -0.00 min

Lab File: BE094094.D

Acq: 24 Sep 2017 1:03

Instrument :

BNA_E





ClientSampleId :

LT-TS-012MSD

Tot Ion:	149	Resp:	49721
Ion Ratio	Lower	Upper	
149	100		
167	6.9	5.4	8.0
279	0.9	0.7	1.1

Manual Integrations
APPROVED

Sohil
9/25/2017 6:24:50 PM



#33

Indeno(1,2,3-cd)pyrene

Concen: 4.478 ng




RT: 26.55 min Scan# 2698


Delta R.T. 0.01 min

Lab File: BE094094.D

Acq: 24 Sep 2017 1:03

Tgt Ion:	276	Resp:	173279
Ion Ratio	Lower	Upper	
276	100		
138	24.0	22.5	33.7
277	24.6	19.9	29.9









#34
Pervlene-d12
Concen: 0.400 ng
RT: 23.91 min Scan# 2122
Delta R.T. 0.00 min
Lab File: BE094094.D
Acq: 24 Sep 2017 1:03

Instrument :
BNA_E
ClientSampleId :
LT-TS-012MSD

Tot Ion	Ratio	Lower	Upper
264	100		
260	27.0	20.5	30.7
265	39.2	22.8	34.2




Manual Integrations
APPROVED


Sohil
9/25/2017 6:24:50 PM



#35
Benzo(b)fluoranthene
Concen: 6.696 ng
RT: 23.14 min Scan# 1953
Delta R.T. 0.01 min
Lab File: BE094094.D
Acq: 24 Sep 2017 1:03

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.0	18.2	27.2
125	12.4	10.1	15.1









#36
Benzo(k)fluoranthene
Concen: 2.485 ng m
RT: 23.19 min Scan# 1963
Delta R.T. 0.00 min
Lab File: BE094094.D
Acq: 24 Sep 2017 1:03

Instrument :
BNA_E
ClientSampleId :
LT-TS-012MSD

Tot Ion	Ratio	Lower	Upper
252	100		
253	23.6	18.6	27.8
125	15.1	10.7	16.1




Manual Integrations
APPROVED


Sohil
9/25/2017 6:24:50 PM



#37
Benzo(a)pyrene
Concen: 5.475 ng
RT: 23.80 min Scan# 2097
Delta R.T. 0.00 min
Lab File: BE094094.D
Acq: 24 Sep 2017 1:03

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.1	19.0	28.6
125	13.2	11.8	17.8









#38
Dibenzo(a,h)anthracene
Concen: 1.091 ng
RT: 26.56 min Scan# 2701
Delta R.T. 0.00 min
Lab File: BE094094.D
Acq: 24 Sep 2017 1:03

Instrument :
BNA_E
ClientSampleId :
LT-TS-012MSD

Tot Ion	Ratio	Lower	Upper
278	100		
139	22.8	16.3	24.5
279	27.2	19.9	29.9

Manual Integrations
APPROVED

Sohil
9/25/2017 6:24:50 PM



#39
Benzo(g,h,i)perylene
Concen: 4.560 ng
RT: 27.38 min Scan# 2880
Delta R.T. 0.02 min
Lab File: BE094094.D
Acq: 24 Sep 2017 1:03

Tgt Ion	Ratio	Lower	Upper
276	100		
277	24.6	20.8	31.2
138	25.3	21.5	32.3

