

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE092415\
 Data File : BE090967.D
 Acq On : 24 Sep 2015 5:54
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Sep 24 08:36:38 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE091815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 18 18:17:34 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	133	0.00
2	1,4-Dioxane	1.000	1.087	-8.7	131	0.00
3	n-Nitrosodimethylamine	1.000	0.833	16.7	105	0.00
4 S	2-Fluorophenol	1.000	0.746	25.4#	96	0.00
5 S	Phenol-d6	1.000	0.662	33.8#	91	0.04
6	bis(2-Chloroethyl)ether	1.000	1.054	-5.4	142	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	141	0.00
8 S	Nitrobenzene-d5	1.000	0.812	18.8	115	0.00
9	Nitrobenzene	1.000	0.742	25.8#	109	0.00
10	Naphthalene	1.000	1.040	-4.0	139	0.00
11	Hexachlorobutadiene	1.000	1.196	-19.6	157	0.00
12	2-Methylnaphthalene	1.000	0.967	3.3	140	0.01
13 I	Acenaphthene-d10	5.000	5.000	0.0	151	0.00
14 S	2,4,6-Tribromophenol	1.000	0.671	32.9#	104	0.00
15 S	2-Fluorobiphenyl	1.000	0.957	4.3	143	0.00
16	Acenaphthylene	1.000	0.955	4.5	145	0.00
17	Acenaphthene	1.000	1.018	-1.8	155	0.00
18	Fluorene	1.000	0.971	2.9	145	0.00
19 I	Phenanthrene-d10	5.000	5.000	0.0	137	0.00
20	4-Bromophenyl-phenylether	1.000	0.999	0.1	139	0.00
21	Hexachlorobenzene	1.000	1.206	-20.6	154	0.00
22	Pentachlorophenol	1.000	0.541	45.9#	60	0.04
23	Phenanthrene	1.000	1.004	-0.4	133	0.00
24	Anthracene	1.000	0.902	9.8	128	0.00
25	Fluoranthene	1.000	0.879	12.1	121	0.00
26 I	Chrysene-d12	5.000	5.000	0.0	107	0.00
27	Pyrene	1.000	1.048	-4.8	120	0.00
28 S	Terphenyl-d14	1.000	1.025	-2.5	112	0.01
29	Benzo(a)anthracene	1.000	0.848	15.2	95	0.00
30	Chrysene	1.000	1.144	-14.4	115	0.00
31	Bis(2-ethylhexyl)phthalate	1.000	0.787	21.3	86	0.00
32	Indeno(1,2,3-cd)pyrene	1.000	0.831	16.9	88	0.00
33 I	Perylene-d12	5.000	5.000	0.0	97	0.00
34	Benzo(b)fluoranthene	1.000	0.882	11.8	88	0.00
35	Benzo(k)fluoranthene	1.000	1.090	-9.0	106	0.00
36 C	Benzo(a)pyrene	1.000	1.044	-4.4	105	0.00
37	Dibenzo(a,h)anthracene	1.000	0.860	14.0	84	0.00
38	Benzo(g,h,i)perylene	1.000	0.909	9.1	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0