

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE092518\
 Data File : BE097669.D
 Acq On : 25 Sep 2018 22:13
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 26 08:31:52 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 24 18:40:34 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.716	0.659	8.0	110	0.00
3	n-Nitrosodimethylamine	0.421	0.451	-7.1	121	0.00
4 S	2-Fluorophenol	1.053	1.033	1.9	119	0.00
5 S	Phenol-d6	1.487	1.515	-1.9	124	0.00
6	bis(2-Chloroethyl)ether	1.204	1.383	-14.9	118	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
8 S	Nitrobenzene-d5	0.232	0.253	-9.1	123	0.00
9	Naphthalene	3.636	3.654	-0.5	115	-0.01
10	Hexachlorobutadiene	0.165	0.161	2.4	113	0.00
11 SURR	2-Methylnaphthalene-d10	0.592	0.571	3.5	109	0.00
12	2-Methylnaphthalene	0.576	0.588	-2.1	116	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
14 S	2,4,6-Tribromophenol	0.155	0.163	-5.2	130	0.00
15 S	2-Fluorobiphenyl	1.305	1.294	0.8	114	0.00
16	Acenaphthylene	1.608	1.525	5.2	114	0.00
17	Acenaphthene	1.050	1.037	1.2	115	0.00
18	Fluorene	1.477	1.429	3.2	114	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
20	4-Bromophenyl-phenylether	0.177	0.178	-0.6	114	0.00
21	Hexachlorobenzene	0.204	0.204	0.0	113	0.00
22	Pentachlorophenol	0.049	0.044	10.2	135	0.00
23	Phenanthrene	0.946	0.926	2.1	111	0.00
24	Anthracene	0.786	0.765	2.7	115	0.00
25 SURR	Fluoranthene-d10	4.361	4.155	4.7	112	0.00
26	Fluoranthene	1.135	1.095	3.5	113	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
28	Pvrene	1.044	1.050	-0.6	113	0.00
29 S	Terphenyl-d14	0.787	0.760	3.4	110	0.00
30	Benzo(a)anthracene	0.945	0.902	4.6	115	0.00
31	Chrysene	1.106	1.119	-1.2	109	0.00
32	Bis(2-ethylhexyl)phthalate	1.080	1.190	-10.2	131	0.00
33	Indeno(1,2,3-cd)pyrene	1.165	1.272	-9.2	121	0.00
34 I	Perylene-d12	1.000	1.000	0.0	112	0.00
35	Benzo(b)fluoranthene	1.023	0.947	7.4	112	0.00
36	Benzo(k)fluoranthene	1.211	1.203	0.7	120	0.00
37 C	Benzo(a)pyrene	1.010	0.990	2.0	122	0.00
38	Dibenzo(a,h)anthracene	1.062	1.037	2.4	123	0.00
39	Benzo(a,h,i)perylene	1.056	1.003	5.0	115	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE092518\
Data File : BE097669.D
Acq On : 25 Sep 2018 22:13
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 26 08:31:52 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 24 18:40:34 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			