

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE092518\
 Data File : BE097679.D
 Acq On : 26 Sep 2018 4:23
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 26 06:57:54 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 24 18:40:34 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	-0.01
2	1,4-Dioxane	0.716	0.678	5.3	121	0.00
3	n-Nitrosodimethylamine	0.421	0.481	-14.3	138	-0.01
4 S	2-Fluorophenol	1.053	1.176	-11.7	144	-0.01
5 S	Phenol-d6	1.487	1.557	-4.7	136	-0.01
6	bis(2-Chloroethyl)ether	1.204	1.408	-16.9	128	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	119	-0.01
8 S	Nitrobenzene-d5	0.232	0.278	-19.8	152#	-0.01
9	Naphthalene	3.636	3.541	2.6	125	-0.01
10	Hexachlorobutadiene	0.165	0.157	4.8	123	0.00
11 SURR	2-Methylnaphthalene-d10	0.592	0.577	2.5	124	0.00
12	2-Methylnaphthalene	0.576	0.587	-1.9	130	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	119	-0.01
14 S	2,4,6-Tribromophenol	0.155	0.178	-14.8	157#	-0.01
15 S	2-Fluorobiphenyl	1.305	1.302	0.2	126	-0.01
16	Acenaphthylene	1.608	1.594	0.9	131	0.00
17	Acenaphthene	1.050	1.057	-0.7	129	0.00
18	Fluorene	1.477	1.424	3.6	125	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	113	0.00
20	4-Bromophenyl-phenylether	0.177	0.180	-1.7	124	0.00
21	Hexachlorobenzene	0.204	0.195	4.4	116	0.00
22	Pentachlorophenol	0.049	0.058	-18.4	194#	-0.01
23	Phenanthrene	0.946	0.937	1.0	121	0.00
24	Anthracene	0.786	0.790	-0.5	128	0.00
25 SURR	Fluoranthene-d10	4.361	4.212	3.4	122	0.00
26	Fluoranthene	1.135	1.095	3.5	122	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	115	0.00
28	Pvrene	1.044	1.049	-0.5	123	0.00
29 S	Terphenyl-d14	0.787	0.769	2.3	122	0.00
30	Benzo(a)anthracene	0.945	0.980	-3.7	137	0.00
31	Chrysene	1.106	1.086	1.8	116	0.00
32	Bis(2-ethylhexyl)phthalate	1.080	1.465	-35.6#	177#	0.00
33	Indeno(1,2,3-cd)pyrene	1.165	1.295	-11.2	135	0.00
34 I	Perylene-d12	1.000	1.000	0.0	124	0.00
35	Benzo(b)fluoranthene	1.023	0.969	5.3	126	0.00
36	Benzo(k)fluoranthene	1.211	1.162	4.0	128	0.00
37 C	Benzo(a)pyrene	1.010	0.970	4.0	132	0.00
38	Dibenzo(a,h)anthracene	1.062	1.070	-0.8	140	-0.01
39	Benzo(a,h,i)perylene	1.056	1.022	3.2	130	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE092518\
Data File : BE097679.D
Acq On : 26 Sep 2018 4:23
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 26 06:57:54 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 24 18:40:34 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		