

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE092518\
 Data File : BE097679.D
 Acq On : 26 Sep 2018 4:23
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Sep 26 06:57:54 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 24 18:40:34 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	114	-0.01
2	1,4-Dioxane	0.400	0.394	1.5	121	0.00
3	n-Nitrosodimethylamine	0.400	0.425	-6.2	138	-0.01
4 S	2-Fluorophenol	0.400	0.447	-11.7	144	-0.01
5 S	Phenol-d6	0.400	0.419	-4.7	136	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.449	-12.2	128	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	119	-0.01
8 S	Nitrobenzene-d5	0.400	0.446	-11.5	152	-0.01
9	Naphthalene	0.400	0.390	2.5	125	-0.01
10	Hexachlorobutadiene	0.400	0.380	5.0	123	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.390	2.5	124	0.00
12	2-Methylnaphthalene	0.400	0.408	-2.0	130	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	119	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.436	-9.0	157	-0.01
15 S	2-Fluorobiphenyl	0.400	0.399	0.3	126	-0.01
16	Acenaphthylene	0.400	0.396	1.0	131	0.00
17	Acenaphthene	0.400	0.403	-0.8	129	0.00
18	Fluorene	0.400	0.386	3.5	125	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	113	0.00
20	4-Bromophenyl-phenylether	0.400	0.405	-1.3	124	0.00
21	Hexachlorobenzene	0.400	0.381	4.8	116	0.00
22	Pentachlorophenol	0.400	0.564	-41.0#	194	-0.01
23	Phenanthrene	0.400	0.396	1.0	121	0.00
24	Anthracene	0.400	0.402	-0.5	128	0.00
25 SURR	Fluoranthene-d10	0.400	0.386	3.5	122	0.00
26	Fluoranthene	0.400	0.386	3.5	122	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	115	0.00
28	Pvrene	0.400	0.402	-0.5	123	0.00
29 S	Terphenyl-d14	0.400	0.391	2.3	122	0.00
30	Benzo(a)anthracene	0.400	0.415	-3.7	137	0.00
31	Chrysene	0.400	0.393	1.8	116	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.543	-35.8#	177	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.444	-11.0	135	0.00
34 I	Perylene-d12	0.400	0.400	0.0	124	0.00
35	Benzo(b)fluoranthene	0.400	0.379	5.3	126	0.00
36	Benzo(k)fluoranthene	0.400	0.384	4.0	128	0.00
37 C	Benzo(a)pyrene	0.400	0.384	4.0	132	0.00
38	Dibenzo(a,h)anthracene	0.400	0.403	-0.8	140	-0.01
39	Benzo(a,h,i)perylene	0.400	0.387	3.3	130	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE092518\
Data File : BE097679.D
Acq On : 26 Sep 2018 4:23
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 26 06:57:54 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 24 18:40:34 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			