

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE092616\
 Data File : BE092425.D
 Acq On : 26 Sep 2016 12:37
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 26 13:56:07 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 22 18:25:56 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	144	-0.02
2	1,4-Dioxane	0.635	0.577	9.1	130	-0.01
3	n-Nitrosodimethylamine	0.588	0.525	10.7	143	0.00
4 S	2-Fluorophenol	0.999	0.854	14.5	132	0.00
5 S	Phenol-d6	1.388	1.088	21.6	130	0.00
6	bis(2-Chloroethyl)ether	1.222	1.025	16.1	156#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	140	0.00
8 S	Nitrobenzene-d5	0.299	0.260	13.0	144	0.00
9	Naphthalene	1.082	1.039	4.0	135	-0.02
10	Hexachlorobutadiene	0.170	0.162	4.7	135	-0.02
11	2-Methylnaphthalene	0.716	0.681	4.9	135	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	137	0.00
13 S	2,4,6-Tribromophenol	0.144	0.108	25.0	120	0.00
14 S	2-Fluorobiphenyl	1.184	1.174	0.8	138	0.00
15	Acenaphthylene	7.240	6.696	7.5	130	0.00
16	Acenaphthene	1.134	1.088	4.1	127	0.00
17	Fluorene	1.438	1.342	6.7	130	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	131	0.02
19	Hexachlorobenzene	0.214	0.207	3.3	136	0.00
20	Pentachlorophenol	0.050	0.060	-20.0	188#	0.00
21	Phenanthrene	1.149	1.173	-2.1	141	0.00
22	Anthracene	1.105	1.034	6.4	135	0.02
23	Fluoranthene	5.460	5.434	0.5	142	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	139	0.00
25	Pyrene	4.333	4.507	-4.0	150#	0.00
26 S	Terphenyl-d14	0.620	0.622	-0.3	141	0.00
27	Benzo(a)anthracene	4.873	4.616	5.3	134	0.00
28	Chrysene	4.462	4.443	0.4	129	0.00
29	Bis(2-ethylhexyl)phthalate	0.420	0.458	-9.0	150	0.00
30	Indeno(1,2,3-cd)pyrene	4.734	4.157	12.2	124	0.04
31 I	Perylene-d12	1.000	1.000	0.0	131	0.02
32	Benzo(b)fluoranthene	1.211	1.161	4.1	127	0.02
33	Benzo(k)fluoranthene	1.267	1.219	3.8	136	0.02
34 C	Benzo(a)pyrene	1.187	1.140	4.0	134	0.02
35	Dibenzo(a,h)anthracene	1.088	1.004	7.7	127	0.05
36	Benzo(g,h,i)perylene	4.673	4.196	10.2	122	0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0