

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE092616\
 Data File : BE092425.D
 Acq On : 26 Sep 2016 12:37
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 26 13:56:07 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 22 18:25:56 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	144	-0.02
2	1,4-Dioxane	0.400	0.325	18.8	130	-0.01
3	n-Nitrosodimethylamine	0.400	0.446	-11.5	143	0.00
4 S	2-Fluorophenol	0.400	0.398	0.5	132	0.00
5 S	Phenol-d6	0.400	0.341	14.8	130	0.00
6	bis(2-Chloroethyl)ether	0.400	0.395	1.3	156	-0.02
7 I	Naphthalene-d8	5.000	5.000	0.0	140	0.00
8 S	Nitrobenzene-d5	0.400	0.426	-6.5	144	0.00
9	Naphthalene	0.400	0.384	4.0	135	-0.02
10	Hexachlorobutadiene	0.400	0.380	5.0	135	-0.02
11	2-Methylnaphthalene	0.400	0.380	5.0	135	-0.01
12 I	Acenaphthene-d10	5.000	5.000	0.0	137	0.00
13 S	2,4,6-Tribromophenol	0.400	0.401	-0.3	120	0.00
14 S	2-Fluorobiphenyl	0.400	0.397	0.8	138	0.00
15	Acenaphthylene	0.400	0.370	7.5	130	0.00
16	Acenaphthene	0.400	0.384	4.0	127	0.00
17	Fluorene	0.400	0.373	6.8	130	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	131	0.02
19	Hexachlorobenzene	0.400	0.387	3.3	136	0.00
20	Pentachlorophenol	0.400	0.478	-19.5	188	0.00
21	Phenanthrene	0.400	0.408	-2.0	141	0.00
22	Anthracene	0.400	0.374	6.5	135	0.02
23	Fluoranthene	0.400	0.398	0.5	142	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	139	0.00
25	Pyrene	0.400	0.416	-4.0	150	0.00
26 S	Terphenyl-d14	0.400	0.401	-0.3	141	0.00
27	Benzo(a)anthracene	0.400	0.379	5.3	134	0.00
28	Chrysene	0.400	0.398	0.5	129	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.436	-9.0	150	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.351	12.3	124	0.04
31 I	Perylene-d12	5.000	5.000	0.0	131	0.02
32	Benzo(b)fluoranthene	0.400	0.383	4.3	127	0.02
33	Benzo(k)fluoranthene	0.400	0.385	3.8	136	0.02
34 C	Benzo(a)pyrene	0.400	0.384	4.0	134	0.02
35	Dibenzo(a,h)anthracene	0.400	0.369	7.8	127	0.05
36	Benzo(g,h,i)perylene	0.400	0.359	10.3	122	0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0