

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE092616\
 Data File : BE092430.D
 Acq On : 26 Sep 2016 17:48
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 26 19:24:45 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 22 18:25:56 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	154#	-0.02
2	1,4-Dioxane	0.635	0.583	8.2	140	-0.02
3	n-Nitrosodimethylamine	0.588	0.592	-0.7	172#	-0.02
4 S	2-Fluorophenol	0.999	0.869	13.0	144	-0.01
5 S	Phenol-d6	1.388	1.109	20.1	142	0.00
6	bis(2-Chloroethyl)ether	1.222	1.082	11.5	175#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	150	0.00
8 S	Nitrobenzene-d5	0.299	0.274	8.4	162#	0.00
9	Naphthalene	1.082	1.041	3.8	145	-0.02
10	Hexachlorobutadiene	0.170	0.160	5.9	142	-0.02
11	2-Methylnaphthalene	0.716	0.681	4.9	145	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	142	0.00
13 S	2,4,6-Tribromophenol	0.144	0.113	21.5	130	0.00
14 S	2-Fluorobiphenyl	1.184	1.200	-1.4	146	0.00
15	Acenaphthylene	7.240	6.685	7.7	135	0.00
16	Acenaphthene	1.134	1.103	2.7	134	0.00
17	Fluorene	1.438	1.341	6.7	134	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	131	0.02
19	Hexachlorobenzene	0.214	0.219	-2.3	143	0.00
20	Pentachlorophenol	0.050	0.043	14.0	135	0.00
21	Phenanthrene	1.149	1.162	-1.1	139	0.00
22	Anthracene	1.105	1.079	2.4	140	0.02
23	Fluoranthene	5.460	5.692	-4.2	148	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	133	0.00
25	Pyrene	4.333	4.810	-11.0	153#	0.00
26 S	Terphenyl-d14	0.620	0.658	-6.1	142	0.00
27	Benzo(a)anthracene	4.873	5.238	-7.5	146	0.00
28	Chrysene	4.462	4.484	-0.5	124	0.00
29	Bis(2-ethylhexyl)phthalate	0.420	0.428	-1.9	134	0.00
30	Indeno(1,2,3-cd)pyrene	4.734	4.639	2.0	132	0.03
31 I	Perylene-d12	1.000	1.000	0.0	136	0.04
32	Benzo(b)fluoranthene	1.211	1.150	5.0	130	0.02
33	Benzo(k)fluoranthene	1.267	1.260	0.6	145	0.04
34 C	Benzo(a)pyrene	1.187	1.158	2.4	140	0.02
35	Dibenzo(a,h)anthracene	1.088	1.040	4.4	136	0.05
36	Benzo(g,h,i)perylene	4.673	4.348	7.0	130	0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0